

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120624\
 Data File : BN035440.D
 Acq On : 06 Dec 2024 00:15
 Operator : RC/JU
 Sample : PB165398BL
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK398

Quant Time: Dec 06 02:04:00 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN120224.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 02 18:25:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.299	152	90991	20.000	ng/u1	0.00
20) Naphthalene-d8	10.046	136	391668	20.000	ng/u1	0.00
38) Acenaphthene-d10	13.957	164	307322	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.728	188	690346	20.000	ng/u1	0.00
79) Chrysene-d12	20.974	240	724033	20.000	ng/u1	0.00
88) Perylene-d12	23.651	264	807120	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.958	96	10365	5.383	ng/uL	0.00
4) Pyridine-d5	3.340	84	142040	27.983	ng/u1	0.00
7) Phenol-d5	6.499	99	180302	27.815	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.657	67	105315	29.133	ng/u1	0.00
11) 2-Chlorophenol-d4	6.840	132	153988	28.034	ng/u1	0.00
15) 4-Methylphenol-d8	8.010	113	156864	27.768	ng/u1	0.00
21) Nitrobenzene-d5	8.434	128	75904	29.829	ng/u1	0.00
24) 2-Nitrophenol-d4	9.146	143	89947	30.539	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.681	165	173109	26.891	ng/u1	0.00
31) 4-Chloroaniline-d4	10.193	131	234020	29.441	ng/u1	0.00
46) Dimethylphthalate-d6	13.387	166	646063	29.277	ng/u1	0.00
49) Acenaphthylene-d8	13.645	160	688627	28.336	ng/u1	0.00
54) 4-Nitrophenol-d4	14.192	143	97925	26.949	ng/u1	0.00
60) Fluorene-d10	14.969	176	535112	28.080	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.098	200	89540	24.594	ng/u1	0.00
73) Anthracene-d10	16.822	188	900147	29.563	ng/u1	0.00
81) Pyrene-d10	19.145	212	1093750	28.602	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.468	264	1156599	29.593	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120624\
Data File : BN035440.D
Acq On : 06 Dec 2024 00:15
Operator : RC/JU
Sample : PB165398BL
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK398

Quant Time: Dec 06 02:04:00 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN120224.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Dec 02 18:25:21 2024
Response via : Initial Calibration

