

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120624\
 Data File : BN035463.D
 Acq On : 06 Dec 2024 15:53
 Operator : RC/JU
 Sample : P5107-05
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BHJ82

Quant Time: Dec 06 16:28:13 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN120224.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 02 18:25:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.298	152	97849	20.000	ng/u1	0.00
20) Naphthalene-d8	10.045	136	417874	20.000	ng/u1	0.00
38) Acenaphthene-d10	13.957	164	328950	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.721	188	745842	20.000	ng/u1	0.00
79) Chrysene-d12	20.974	240	778132	20.000	ng/u1	0.00
88) Perylene-d12	23.645	264	872683	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.963	96	10116	4.885	ng/uL	0.00
4) Pyridine-d5	3.340	84	59153	10.837	ng/u1	0.00
7) Phenol-d5	6.498	99	54704	7.848	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.657	67	108652	27.950	ng/u1	0.00
11) 2-Chlorophenol-d4	6.840	132	143371	24.271	ng/u1	0.00
15) 4-Methylphenol-d8	8.010	113	109938	18.097	ng/u1	0.00
21) Nitrobenzene-d5	8.434	128	78580	28.944	ng/u1	0.00
24) 2-Nitrophenol-d4	9.145	143	94078	29.938	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.681	165	185405	26.995	ng/u1	0.00
31) 4-Chloroaniline-d4	10.192	131	219007	25.825	ng/u1	0.00
46) Dimethylphthalate-d6	13.386	166	748618	31.694	ng/u1	0.00
49) Acenaphthylene-d8	13.645	160	768319	29.537	ng/u1	0.00
54) 4-Nitrophenol-d4	14.192	143	28875	7.424	ng/u1	0.00
60) Fluorene-d10	14.969	176	645449	31.643	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.098	200	118321	30.081	ng/u1	0.00
73) Anthracene-d10	16.821	188	1056125	32.105	ng/u1	0.00
81) Pyrene-d10	19.145	212	1296497	31.547	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.468	264	1423833	33.694	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120624\
Data File : BN035463.D
Acq On : 06 Dec 2024 15:53
Operator : RC/JU
Sample : P5107-05
Misc :
ALS Vial : 43 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BHJ82

Quant Time: Dec 06 16:28:13 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN120224.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Dec 02 18:25:21 2024
Response via : Initial Calibration

