

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120624\
 Data File : BN035475.D
 Acq On : 07 Dec 2024 00:19
 Operator : RC/JU
 Sample : P5107-12
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BHJ89

Quant Time: Dec 07 01:23:46 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN120224.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 02 18:25:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.293	152	94223	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.045	136	398636	20.000	ng/u1	0.00
38) Acenaphthene-d10	13.957	164	314807	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.721	188	707992	20.000	ng/u1	0.00
79) Chrysene-d12	20.974	240	712652	20.000	ng/u1	0.00
88) Perylene-d12	23.645	264	773835	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.957	96	8733	4.379	ng/uL	0.00
4) Pyridine-d5	3.340	84	39624	7.539	ng/u1	0.00
7) Phenol-d5	6.498	99	49704	7.405	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.657	67	102727	27.443	ng/u1	0.00
11) 2-Chlorophenol-d4	6.834	132	133633	23.494	ng/u1	-0.01
15) 4-Methylphenol-d8	8.010	113	99938	17.084	ng/u1	0.00
21) Nitrobenzene-d5	8.434	128	75054	28.979	ng/u1	0.00
24) 2-Nitrophenol-d4	9.145	143	87887	29.318	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.681	165	173437	26.471	ng/u1	0.00
31) 4-Chloroaniline-d4	10.192	131	209646	25.914	ng/u1	0.00
46) Dimethylphthalate-d6	13.386	166	700277	30.980	ng/u1	0.00
49) Acenaphthylene-d8	13.639	160	713193	28.650	ng/u1	0.00
54) 4-Nitrophenol-d4	14.192	143	22087	5.934	ng/u1	0.00
60) Fluorene-d10	14.963	176	600301	30.752	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.098	200	103785	27.796	ng/u1	0.00
73) Anthracene-d10	16.821	188	981009	31.415	ng/u1	0.00
81) Pyrene-d10	19.145	212	1200684	31.900	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.468	264	1262402	33.689	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120624\
Data File : BN035475.D
Acq On : 07 Dec 2024 00:19
Operator : RC/JU
Sample : P5107-12
Misc :
ALS Vial : 56 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BHJ89

Quant Time: Dec 07 01:23:46 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN120224.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Dec 02 18:25:21 2024
Response via : Initial Calibration

