


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120624\
Data File : BN035423.D
Acq On : 05 Dec 2024 12:33
Operator : RC/JU
Sample : DFTPP367
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP367

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN120224.MA.M
Title : SVOA CALIBRATION
Last Update : Mon Dec 02 18:25:21 2024



AutoFind: Scans 2436, 2437, 2438; Background Corrected with Scan 2430

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	22.9	73152	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	30.5	97184	PASS
70	69	0.00	2	0.5	482	PASS
127	198	10	80	40.3	128669	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	318976	PASS
199	198	5	9	6.8	21544	PASS
275	198	10	60	28.9	92312	PASS
365	198	1	100	4.5	14376	PASS
441	198	0.01	100	12.7	40501	PASS
442	198	50	100	71.2	227029	PASS
443	442	15	24	19.7	44763	PASS

m/z	Abundance
40.00	366.0
44.00	1500.0
73.00	525.0
132.90	166.0
146.90	183.0
190.90	175.0
207.00	1521.0
208.00	254.0
209.00	255.0
281.00	302.0

Instrument :

BNA_N

ClientSampleId :

DFTPP367

m/z	Abundance
37.00	382.0
38.00	1121.0
39.00	5932.0
40.00	599.0
41.10	191.0
44.00	1264.0
50.00	24136.0
51.00	85808.0
52.00	4428.0
53.00	229.0
55.00	819.0
56.00	2665.0
57.00	6117.0
58.00	277.0
60.90	1363.0
62.00	1670.0
63.00	3903.0
64.00	520.0
65.00	1840.0
66.90	270.0
69.00	112024.0
70.00	532.0
73.00	1999.0
74.00	12293.0
75.00	18896.0
76.10	6712.0
77.00	126456.0
78.00	8844.0
79.00	9598.0
80.00	7047.0
81.00	9432.0
82.00	2347.0
83.00	1979.0
84.10	272.0
85.00	1667.0
86.00	3102.0
87.00	1453.0
87.80	459.0
88.90	310.0
90.20	156.0
90.90	2251.0
92.00	2695.0
93.00	17336.0
94.00	1220.0
95.00	215.0
96.10	745.0
98.00	13614.0
99.00	9263.0
100.00	858.0
101.00	5152.0
101.90	329.0
103.00	1889.0
104.00	3396.0
105.00	3370.0
107.00	42912.0
108.00	6222.0
110.00	75552.0
111.00	11653.0
112.00	1494.0
112.90	498.0
114.80	248.0
116.10	2442.0
117.00	41064.0
118.00	3055.0
118.90	416.0
119.90	667.0
120.90	261.0
121.90	2886.0
123.00	4317.0
124.00	1921.0
124.90	1819.0
127.00	140352.0
128.00	11344.0
129.00	56704.0
130.00	5212.0
131.00	939.0
132.00	490.0
133.00	269.0
134.00	1878.0
135.00	4895.0
136.00	1994.0
137.00	2701.0
137.80	503.0
138.90	398.0
140.00	786.0
141.00	7656.0
142.00	2614.0
142.90	1635.0
143.90	561.0
144.90	414.0
145.90	1773.0
147.00	4827.0

Instrument :
BNA_N
ClientSampleId :
DFTPP367

148.00	11248.0
148.90	1987.0
150.00	569.0
151.00	1188.0
151.70	704.0
153.00	2507.0
153.90	1689.0
155.00	4610.0
156.00	6766.0
157.00	1242.0
157.90	1478.0
158.90	1234.0
159.90	2670.0
161.00	3953.0
161.90	1072.0
162.90	407.0
164.00	379.0
164.90	2977.0
166.00	2607.0
167.00	17408.0
168.00	9087.0
168.90	1428.0
169.90	582.0
170.90	768.0
172.00	1509.0
173.00	1900.0
174.00	3590.0
175.00	6576.0
176.00	1682.0
176.90	3096.0
177.90	1046.0
178.90	12454.0
180.00	8354.0
180.90	4183.0
182.00	667.0
183.00	340.0
183.90	1125.0
185.00	6209.0
186.00	46920.0
187.00	14357.0
188.00	1510.0
188.90	3214.0
190.00	502.0
191.00	1593.0
192.00	4080.0
193.00	4588.0
194.00	1099.0
195.10	444.0
195.90	9234.0
197.90	323648.0
198.90	22936.0
200.00	1785.0
201.40	1544.0
203.00	2361.0
204.00	12828.0
205.00	22408.0
206.00	90232.0
207.00	12609.0
208.00	3583.0
209.00	1028.0
210.20	1312.0
211.00	3468.0
211.90	649.0
212.90	157.0
215.00	1128.0
216.00	2343.0
216.90	23360.0
217.90	3214.0
219.00	357.0
221.00	11396.0
221.80	3295.0
223.00	5419.0
224.00	48240.0
225.00	12685.0
226.00	1537.0
227.00	22456.0
227.90	3352.0
229.00	4795.0
230.00	736.0
231.00	1890.0
232.00	280.0
232.90	458.0
233.90	1414.0
234.90	1626.0
236.00	1331.0
237.00	1577.0
237.90	293.0
238.90	842.0
239.90	772.0
241.00	1193.0
242.00	2533.0
243.10	2910.0
244.00	35952.0
245.00	5130.0
246.00	8572.0
246.90	1772.0

Instrument :
BNA_N
ClientSampleId :
DFTPP367

248.00	378.0
248.90	1471.0
249.90	223.0
250.90	331.0
251.90	337.0
253.00	943.0
255.00	197312.0
256.00	27704.0
256.90	2441.0
258.00	12193.0
258.90	2101.0
259.90	247.0
260.90	430.0
264.10	374.0
265.00	5612.0
266.00	902.0
266.90	327.0
269.80	270.0
271.10	918.0
273.00	5712.0
274.00	16224.0
275.00	88544.0
276.00	11999.0
276.90	7996.0
278.00	1221.0
279.10	259.0
281.10	781.0
281.90	421.0
282.90	915.0
283.90	673.0
285.00	1385.0
286.00	337.0
288.90	371.0
289.80	257.0
290.90	265.0
292.10	417.0
293.00	1804.0
293.90	482.0
295.90	27000.0
297.00	4213.0
297.90	257.0
300.90	357.0
302.10	495.0
303.00	3010.0
304.10	782.0
304.90	190.0
307.90	340.0
308.90	251.0
310.00	313.0
312.90	208.0
314.00	1421.0
315.00	3044.0
316.00	1835.0
317.00	231.0
321.00	752.0
321.90	441.0
323.00	7787.0
324.00	1734.0
325.00	194.0
325.80	256.0
326.90	1666.0
327.90	857.0
328.90	223.0
332.00	560.0
333.00	822.0
334.00	5897.0
335.00	1458.0
336.00	189.0
341.00	1092.0
341.90	241.0
346.00	1742.0
346.90	281.0
351.10	154.0
352.00	2263.0
352.90	1838.0
354.00	2923.0
355.00	791.0
356.00	226.0
359.00	199.0
365.00	12545.0
365.90	1867.0
369.80	290.0
371.00	603.0
372.00	3967.0
373.00	971.0
382.90	1155.0
384.00	415.0
390.10	498.0
390.80	400.0
392.10	292.0
401.00	342.0
402.00	1490.0
403.00	2400.0
404.00	866.0
421.00	1934.0
422.00	1797.0

Instrument :
BNA_N
ClientSampleId :
DFTPP367

423.10	14225.0
424.10	2748.0
425.00	389.0
437.90	159.0
441.10	34536.0
442.00	173888.0
443.10	37536.0
444.00	3315.0
444.90	265.0

Instrument :
BNA_N
ClientSampleId :
DFTPP367

m/z	Abundance
37.00	493.0
38.00	1089.0
39.00	5982.0
40.00	675.0
41.00	232.0
44.00	1423.0
45.00	177.0
49.00	857.0
50.00	22792.0
51.00	84064.0
52.00	3939.0
53.00	285.0
55.00	583.0
56.00	2767.0
57.00	6229.0
58.00	318.0
61.00	1255.0
62.00	1657.0
63.00	3953.0
64.10	644.0
65.00	1880.0
66.10	198.0
66.70	157.0
69.00	111792.0
70.00	636.0
73.00	2056.0
74.00	12997.0
75.00	18992.0
76.10	6605.0
77.00	126760.0
78.00	8457.0
79.00	9699.0
80.00	7010.0
81.00	10104.0
82.00	2448.0
83.00	2132.0
83.90	319.0
85.00	1625.0
86.00	3069.0
86.90	1350.0
88.00	433.0
88.90	309.0
90.90	2464.0
92.00	2365.0
93.00	17504.0
94.00	1180.0
95.00	295.0
96.00	949.0
98.00	14142.0
99.00	10731.0
99.90	816.0
101.00	5539.0
102.00	374.0
102.90	1910.0
104.00	3465.0
105.00	3442.0
106.00	1169.0
107.00	46448.0
108.00	6876.0
110.00	78880.0
111.00	11478.0
112.00	1566.0
113.00	564.0
114.90	224.0
116.00	2943.0
117.00	43808.0
118.00	2893.0
119.00	533.0
120.00	673.0
120.80	224.0
122.00	2975.0
123.00	4740.0
123.90	2297.0
125.00	2170.0
127.00	151104.0
128.00	11605.0
129.00	61312.0
130.00	5254.0
131.00	1033.0
132.00	484.0
133.00	389.0
134.00	2029.0
135.00	5028.0
136.00	2415.0
137.00	2779.0
138.80	305.0
140.00	871.0
141.00	8873.0
141.90	2690.0
143.00	2004.0
143.90	552.0
145.00	462.0

Instrument :
BNA_N
ClientSampleId :
DFTPP367

146.00	1533.0
147.00	5527.0
148.00	11766.0
149.00	2440.0
149.90	702.0
151.10	1247.0
152.90	2614.0
154.00	1913.0
155.00	4765.0
156.00	7853.0
157.00	1337.0
158.00	1692.0
159.00	1355.0
159.90	3101.0
161.00	4241.0
161.90	1180.0
162.90	295.0
163.90	488.0
164.90	3540.0
166.00	2942.0
167.00	19656.0
168.00	9879.0
169.00	1508.0
170.10	630.0
170.90	1107.0
171.90	1871.0
172.90	2145.0
174.00	3838.0
175.00	7561.0
176.00	1903.0
177.00	3393.0
178.00	1266.0
178.90	14891.0
180.00	9621.0
181.00	4416.0
182.00	835.0
182.70	377.0
184.00	1198.0
185.00	6533.0
186.00	53760.0
187.00	16051.0
188.00	1308.0
188.90	3660.0
190.00	671.0
191.00	1940.0
192.00	4927.0
193.00	5143.0
194.00	998.0
194.90	526.0
196.00	10000.0
197.90	369024.0
198.90	25104.0
200.00	2027.0
201.50	1374.0
202.90	3446.0
204.00	14702.0
205.00	24224.0
206.00	101616.0
207.00	14723.0
208.00	3858.0
208.90	1495.0
210.00	1776.0
211.00	3945.0
211.80	664.0
213.00	300.0
213.90	195.0
214.90	1428.0
216.00	2669.0
216.90	27752.0
217.90	3716.0
218.90	323.0
221.00	13709.0
221.80	3976.0
223.00	6344.0
224.00	56968.0
225.00	15021.0
226.10	1839.0
226.90	27816.0
228.00	3868.0
228.90	5557.0
230.00	626.0
231.00	2238.0
231.90	287.0
233.00	542.0
233.90	1780.0
234.90	1996.0
235.90	1262.0
237.00	1925.0
237.90	324.0
239.00	1070.0
239.90	743.0
240.90	1596.0
242.00	3294.0
243.10	3095.0
244.00	43056.0
245.00	6524.0

Instrument :
BNA_N
ClientSampleId :
DFTPP367

246.00	10222.0
247.00	1804.0
248.00	498.0
248.90	1718.0
249.80	321.0
251.00	421.0
252.00	441.0
253.00	933.0
255.00	249856.0
256.00	36616.0
257.00	2753.0
258.00	15755.0
259.00	2364.0
260.00	419.0
260.90	349.0
262.90	196.0
265.00	6815.0
265.90	1005.0
267.00	226.0
269.90	319.0
271.00	925.0
272.00	835.0
272.90	7318.0
274.00	20848.0
275.00	109672.0
276.00	13901.0
277.00	9935.0
277.90	1748.0
279.00	490.0
281.00	568.0
281.90	297.0
283.00	1209.0
283.90	881.0
285.00	1975.0
286.00	403.0
288.90	412.0
290.00	397.0
290.90	372.0
292.00	529.0
292.90	2310.0
294.00	648.0
296.00	35704.0
297.00	4804.0
297.90	196.0
301.10	427.0
301.90	681.0
303.00	3823.0
304.00	1220.0
307.90	522.0
309.00	288.0
309.90	445.0
311.90	193.0
313.00	294.0
314.00	1711.0
315.00	4003.0
316.00	2020.0
317.00	394.0
321.00	1327.0
322.00	639.0
323.00	11035.0
324.00	1933.0
325.00	217.0
326.00	287.0
326.90	2106.0
327.90	1086.0
328.90	208.0
331.90	881.0
333.00	1009.0
334.00	7710.0
335.00	2126.0
336.00	290.0
339.00	175.0
341.00	1459.0
342.00	484.0
345.90	2418.0
346.90	571.0
350.90	200.0
352.00	2951.0
353.00	2411.0
354.00	3178.0
355.00	952.0
356.10	174.0
359.00	354.0
364.90	17808.0
365.90	2465.0
367.00	190.0
370.00	387.0
371.00	918.0
372.00	5708.0
373.00	1567.0
374.00	154.0
383.00	1574.0
383.90	625.0
390.00	689.0
391.00	515.0
392.00	268.0

Instrument :
BNA_N
ClientSampleId :
DFTPP367

401.00	412.0
402.00	2519.0
403.00	3415.0
404.10	1380.0
405.00	205.0
415.00	165.0
420.90	2747.0
422.00	2866.0
423.00	18104.0
424.10	4143.0
425.00	322.0
438.00	211.0
439.00	205.0
441.10	49544.0
442.00	282560.0
443.00	53744.0
444.00	4753.0
445.00	259.0

Instrument :
BNA_N
ClientSampleId :
DFTPP367

m/z	Abundance
37.10	260.0
38.00	611.0
39.00	3510.0
39.90	524.0
41.00	184.0
44.00	1649.0
44.90	166.0
50.00	13257.0
51.00	49584.0
52.00	2433.0
52.90	156.0
55.00	369.0
56.00	1959.0
57.00	3698.0
60.90	807.0
62.00	889.0
63.00	2585.0
64.00	463.0
65.00	1342.0
69.00	67736.0
70.00	279.0
73.00	1352.0
74.00	7777.0
75.00	11314.0
76.10	4049.0
77.00	78456.0
78.00	5423.0
79.00	5436.0
80.00	4387.0
81.00	5666.0
81.90	1659.0
83.00	1328.0
84.00	153.0
85.00	1024.0
86.00	1862.0
86.90	765.0
87.80	235.0
91.00	1370.0
92.10	1376.0
93.00	10799.0
94.00	750.0
95.00	167.0
95.90	435.0
96.80	235.0
98.00	8564.0
99.00	6517.0
100.00	590.0
101.00	3309.0
101.90	207.0
103.00	1087.0
104.00	2369.0
104.90	2375.0
106.10	885.0
107.00	28144.0
108.00	4315.0
110.00	47848.0
111.00	7505.0
111.90	853.0
112.90	315.0
114.70	218.0
116.10	1787.0
117.00	26968.0
117.90	1734.0
119.00	267.0
119.90	393.0
122.00	1952.0
123.00	2993.0
124.00	1370.0
125.00	1188.0
127.00	94552.0
128.00	7193.0
129.00	38640.0
129.90	3355.0
131.00	620.0
131.90	327.0
132.90	293.0
133.90	1330.0
134.90	3045.0
135.90	1286.0
137.00	1958.0
138.10	483.0
138.90	219.0
140.10	581.0
141.00	5198.0
142.00	1783.0
142.90	1313.0
144.00	319.0
144.80	191.0
146.00	1179.0
147.00	3138.0
148.00	7460.0
149.00	1616.0

Instrument :
BNA_N
ClientSampleId :
DFTPP367

150.00	454.0
151.00	778.0
151.70	392.0
153.00	1979.0
154.00	1548.0
155.00	3285.0
156.00	4613.0
157.10	1084.0
157.90	1058.0
159.00	844.0
160.00	1785.0
161.00	2884.0
161.90	807.0
162.80	258.0
164.20	350.0
164.90	2265.0
166.00	1957.0
167.00	13176.0
167.90	6672.0
168.90	1121.0
170.00	369.0
171.00	530.0
172.00	1159.0
172.90	1635.0
174.00	2590.0
175.00	5170.0
176.00	1117.0
176.90	2410.0
177.90	895.0
179.00	9615.0
180.00	6203.0
181.00	3005.0
182.00	442.0
182.70	176.0
183.30	275.0
183.90	817.0
185.00	4483.0
186.00	36328.0
187.00	10370.0
188.00	1199.0
188.90	2624.0
189.90	515.0
191.00	1146.0
192.00	2911.0
193.00	3722.0
193.90	815.0
196.00	7250.0
198.00	264256.0
199.00	16592.0
200.00	1413.0
201.40	1027.0
202.90	1871.0
204.00	10631.0
205.00	17240.0
206.00	67632.0
207.00	9818.0
208.00	2704.0
208.90	1010.0
210.90	3010.0
213.10	293.0
213.80	171.0
214.90	798.0
215.90	1569.0
216.90	19672.0
218.00	2513.0
219.10	270.0
221.00	9972.0
221.90	2597.0
222.90	4507.0
224.00	40152.0
225.00	10206.0
226.00	1207.0
227.00	18328.0
228.00	2502.0
228.90	3824.0
230.00	426.0
231.00	1473.0
232.10	240.0
233.00	283.0
233.90	1215.0
235.00	1174.0
235.90	939.0
236.90	1397.0
238.10	237.0
238.90	625.0
239.90	588.0
240.90	1113.0
242.00	2113.0
243.00	2526.0
244.00	31824.0
245.00	3783.0
246.00	7481.0
247.00	1386.0
248.00	325.0
248.90	1250.0
249.90	283.0

Instrument :
BNA_N
ClientSampleId :
DFTPP367

250.90	382.0
251.90	338.0
253.00	806.0
255.00	177920.0
256.00	24616.0
257.00	1679.0
257.90	10290.0
259.00	1588.0
260.00	203.0
260.90	253.0
264.10	339.0
265.00	4461.0
266.00	562.0
266.80	180.0
271.00	826.0
272.00	590.0
273.00	5480.0
274.00	14702.0
275.00	78720.0
276.00	9756.0
276.90	7323.0
277.90	1239.0
278.90	279.0
281.10	682.0
281.90	309.0
283.00	848.0
283.90	575.0
284.90	1213.0
285.90	280.0
288.90	373.0
289.80	242.0
291.00	324.0
292.00	320.0
292.90	1655.0
293.90	381.0
296.00	25056.0
297.00	3582.0
297.80	276.0
300.90	284.0
301.90	469.0
303.00	2623.0
304.00	782.0
307.90	379.0
308.90	280.0
310.00	419.0
312.90	269.0
314.00	1147.0
314.90	3293.0
316.00	1442.0
317.00	230.0
321.00	865.0
321.90	500.0
323.00	7938.0
324.10	1458.0
324.90	167.0
325.90	234.0
326.90	1703.0
327.90	929.0
332.00	572.0
333.10	860.0
334.00	5588.0
335.00	1237.0
341.00	1214.0
342.00	357.0
346.00	2067.0
346.90	284.0
352.00	2833.0
353.00	1755.0
354.00	2831.0
355.00	637.0
359.00	250.0
365.00	12775.0
365.90	1721.0
366.90	189.0
370.00	226.0
370.90	690.0
372.00	4567.0
373.00	1199.0
376.90	175.0
382.90	1299.0
383.90	375.0
384.90	164.0
390.00	539.0
391.00	474.0
391.90	368.0
401.00	464.0
402.00	2221.0
403.00	2749.0
404.00	966.0
405.00	214.0
421.00	2406.0
422.10	2457.0
423.00	13765.0
424.00	2817.0
425.00	348.0
438.00	255.0

Instrument :
BNA_N
ClientSampleId :
DFTPP367

439.10	155.0
441.00	37424.0
442.00	224640.0
443.00	43008.0
444.00	3875.0
445.00	201.0

Instrument :
BNA_N
ClientSampleId :
DFTPP367