

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120721\
 Data File : BN017779.D
 Acq On : 08 Dec 2021 01:24
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Dec 08 07:34:48 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 08 01:18:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.715	152	21612	0.400	ng	0.00
7) Naphthalene-d8	10.494	136	88070	0.400	ng	0.00
13) Acenaphthene-d10	14.347	164	58385	0.400	ng	0.00
19) Phenanthrene-d10	17.092	188	118557	0.400	ng	0.00
29) Chrysene-d12	21.282	240	121707	0.400	ng	0.00
35) Perylene-d12	23.586	264	103873	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.327	112	20259	0.388	ng	0.00
5) Phenol-d6	6.903	99	18704	0.379	ng	0.00
8) Nitrobenzene-d5	8.859	82	20654	0.353	ng	0.00
11) 2-Methylnaphthalene-d10	12.090	152	66984	0.381	ng	0.00
14) 2,4,6-Tribromophenol	15.844	330	10298	0.337	ng	0.00
15) 2-Fluorobiphenyl	12.964	172	87502	0.409	ng	0.00
27) Fluoranthene-d10	19.129	212	161330	0.407	ng	0.00
31) Terphenyl-d14	19.735	244	106630	0.408	ng	0.00
Target Compounds						
						Qvalue
3) n-Nitrosodimethylamine	3.602	42	8579	0.385	ng	100
6) bis(2-Chloroethyl)ether	7.143	93	23503	0.420	ng	100
9) Naphthalene	10.545	128	88345	0.383	ng	100
10) Hexachlorobutadiene	10.836	225	26787	0.400	ng	# 100
12) 2-Methylnaphthalene	12.161	142	64312	0.387	ng	99
16) Acenaphthylene	14.055	152	86674	0.378	ng	100
17) Acenaphthene	14.398	154	61659	0.403	ng	100
18) Fluorene	15.400	166	75981	0.389	ng	100
20) 4,6-Dinitro-2-methylph...	15.489	198	421	0.368	ng	86
21) 4-Bromophenyl-phenylether	16.289	248	27457	0.383	ng	96
22) Hexachlorobenzene	16.399	284	35512	0.416	ng	99
23) Atrazine	16.557	200	18852	0.363	ng	99
24) Pentachlorophenol	16.764	266	5961	0.466	ng	99
25) Phenanthrene	17.129	178	123174	0.395	ng	100
26) Anthracene	17.226	178	104990	0.383	ng	100
28) Fluoranthene	19.159	202	158222	0.417	ng	100
30) Pyrene	19.522	202	162262	0.421	ng	100
32) Benzo(a)anthracene	21.272	228	151041	0.403	ng	100
33) Chrysene	21.325	228	152293	0.389	ng	100
34) Bis(2-ethylhexyl)phtha...	21.219	149	70658	0.416	ng	99
36) Indeno(1,2,3-cd)pyrene	25.944	276	101293	0.352	ng	99
37) Benzo(b)fluoranthene	22.891	252	139700	0.396	ng	99
38) Benzo(k)fluoranthene	22.935	252	136044	0.408	ng	100
39) Benzo(a)pyrene	23.483	252	126366	0.398	ng	99
40) Dibenzo(a,h)anthracene	25.965	278	74370	0.333	ng	98
41) Benzo(g,h,i)perylene	26.663	276	86319	0.360	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120721\
Data File : BN017779.D
Acq On : 08 Dec 2021 01:24
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Dec 08 07:34:48 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 08 01:18:09 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.715 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: BN017779.D
 Acq: 08 Dec 2021 01:24




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:152 Resp: 21612
 Ion Ratio Lower Upper
 152 100
 150 154.8 123.0 184.4
 115 57.0 45.3 67.9



#3
 n-Nitrosodimethylamine
 Concen: 0.385 ng
 RT: 3.602 min Scan# 37
 Delta R.T. -0.000 min
 Lab File: BN017779.D
 Acq: 08 Dec 2021 01:24

Tgt Ion: 42 Resp: 8579
 Ion Ratio Lower Upper
 42 100
 74 126.6 101.3 151.9
 44 5.8 5.0 7.4









#4
2-Fluorophenol
Concen: 0.388 ng
RT: 5.327 min Scan# 211
Delta R.T. -0.000 min
Lab File: BN017779.D
Acq: 08 Dec 2021 01:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:112 Resp: 20259
Ion Ratio Lower Upper
112 100
64 61.3 49.3 73.9
63 32.8 25.8 38.8



#5
Phenol-d6
Concen: 0.379 ng
RT: 6.903 min Scan# 395
Delta R.T. -0.000 min
Lab File: BN017779.D
Acq: 08 Dec 2021 01:24

Tgt Ion: 99 Resp: 18704
Ion Ratio Lower Upper
99 100
42 21.0 16.6 25.0
71 36.6 29.0 43.4









#6
bis(2-Chloroethyl)ether
Concen: 0.420 ng
RT: 7.143 min Scan# 41
Delta R.T. -0.000 min
Lab File: BN017779.D
Acq: 08 Dec 2021 01:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 93 Resp: 23503
Ion Ratio Lower Upper
93 100
63 77.7 62.0 93.0
95 32.8 26.4 39.6



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.494 min Scan# 720
Delta R.T. -0.000 min
Lab File: BN017779.D
Acq: 08 Dec 2021 01:24

Tgt Ion: 136 Resp: 88070
Ion Ratio Lower Upper
136 100
137 11.1 8.8 13.2
54 7.3 5.9 8.9
68 6.1 5.0 7.6









#8
Nitrobenzene-d5
Concen: 0.353 ng
RT: 8.859 min Scan# 591
Delta R.T. -0.000 min
Lab File: BN017779.D
Acq: 08 Dec 2021 01:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	82	Resp:	20654
Ion Ratio	Lower	Upper	
82	100		
128	38.2	30.2	45.2
54	48.7	39.0	58.6



#9
Naphthalene
Concen: 0.383 ng
RT: 10.545 min Scan# 724
Delta R.T. -0.000 min
Lab File: BN017779.D
Acq: 08 Dec 2021 01:24

Tgt Ion:	128	Resp:	88345
Ion Ratio	Lower	Upper	
128	100		
129	10.9	8.8	13.2
127	13.1	10.6	15.8









#10
Hexachlorobutadiene
Concen: 0.400 ng
RT: 10.836 min Scan# 747
Delta R.T. -0.000 min
Lab File: BN017779.D
Acq: 08 Dec 2021 01:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:225 Resp: 26787
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.9 50.9 76.3



#11
2-Methylnaphthalene-d10
Concen: 0.381 ng
RT: 12.090 min Scan# 929
Delta R.T. -0.000 min
Lab File: BN017779.D
Acq: 08 Dec 2021 01:24

Tgt Ion:152 Resp: 66984
Ion Ratio Lower Upper
152 100
151 20.3 16.4 24.6





#12

2-Methylnaphthalene

Concen: 0.387 ng

RT: 12.161 min Scan# 945

Delta R.T. -0.000 min

Lab File: BN017779.D


Acq: 08 Dec 2021 01:24

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4






Tgt Ion:142 Resp: 64312

Ion Ratio Lower Upper

142 100

141 88.3 70.5 105.7

115 35.8 28.0 42.0



#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.347 min Scan# 1168

Delta R.T. -0.000 min

Lab File: BN017779.D

Acq: 08 Dec 2021 01:24




Tgt Ion:164 Resp: 58385


Ion Ratio Lower Upper

164 100

162 98.3 79.0 118.6

160 43.1 34.6 52.0









#14
2,4,6-Tribromophenol
Concen: 0.337 ng
RT: 15.844 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN017779.D
Acq: 08 Dec 2021 01:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:330 Resp: 10298
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 26.4 21.5 32.3



#15
2-Fluorobiphenyl
Concen: 0.409 ng
RT: 12.964 min Scan# 1059
Delta R.T. -0.000 min
Lab File: BN017779.D
Acq: 08 Dec 2021 01:24

Tgt Ion:172 Resp: 87502
Ion Ratio Lower Upper
172 100
171 35.3 28.1 42.1
170 23.6 18.9 28.3









#16
Acenaphthylene
Concen: 0.378 ng
RT: 14.055 min Scan# 1145
Delta R.T. -0.000 min
Lab File: BN017779.D
Acq: 08 Dec 2021 01:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 86674
Ion Ratio Lower Upper
152 100
151 19.5 15.7 23.5
153 13.0 10.4 15.6



#17
Acenaphthene
Concen: 0.403 ng
RT: 14.398 min Scan# 1172
Delta R.T. -0.000 min
Lab File: BN017779.D
Acq: 08 Dec 2021 01:24

Tgt Ion:154 Resp: 61659
Ion Ratio Lower Upper
154 100
153 112.6 90.1 135.1
152 54.6 43.7 65.5









#18
Fluorene
Concen: 0.389 ng
RT: 15.400 min Scan# 1251
Delta R.T. 0.012 min
Lab File: BN017779.D
Acq: 08 Dec 2021 01:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:166 Resp: 75981
Ion Ratio Lower Upper
166 100
165 97.2 77.8 116.8
167 13.6 11.0 16.4



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.092 min Scan# 1388
Delta R.T. -0.000 min
Lab File: BN017779.D
Acq: 08 Dec 2021 01:24

Tgt Ion:188 Resp: 118557
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.1 5.1 7.7









#20
4,6-Dinitro-2-methylphenol
Concen: 0.368 ng
RT: 15.489 min Scan# 1258
Delta R.T. 0.000 min
Lab File: BN017779.D
Acq: 08 Dec 2021 01:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:198 Resp: 421
Ion Ratio Lower Upper
198 100
182 45.3 44.2 66.2
77 0.0 0.0 0.0



#21
4-Bromophenyl-phenylether
Concen: 0.383 ng
RT: 16.289 min Scan# 1322
Delta R.T. -0.000 min
Lab File: BN017779.D
Acq: 08 Dec 2021 01:24

Tgt Ion:248 Resp: 27457
Ion Ratio Lower Upper
248 100
250 97.7 80.3 120.5
141 41.1 29.8 44.8









#22
Hexachlorobenzene
Concen: 0.416 ng
RT: 16.399 min Scan# 1331
Delta R.T. -0.000 min
Lab File: BN017779.D
Acq: 08 Dec 2021 01:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:284 Resp: 35512
Ion Ratio Lower Upper
284 100
142 25.9 21.0 31.4
249 29.5 23.9 35.9



#23
Atrazine
Concen: 0.363 ng
RT: 16.557 min Scan# 1344
Delta R.T. -0.000 min
Lab File: BN017779.D
Acq: 08 Dec 2021 01:24

Tgt Ion:200 Resp: 18852
Ion Ratio Lower Upper
200 100
173 28.6 21.4 32.2
215 49.3 39.3 58.9









#24
 Pentachlorophenol
 Concen: 0.466 ng
 RT: 16.764 min Scan# 1361
 Delta R.T. -0.000 min
 Lab File: BN017779.D
 Acq: 08 Dec 2021 01:24




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:	266	Resp:	5961
Ion Ratio	Lower	Upper	
266	100		
264	62.8	50.8	76.2
268	63.8	51.8	77.6



#25
 Phenanthrene
 Concen: 0.395 ng
 RT: 17.129 min Scan# 1391
 Delta R.T. -0.000 min
 Lab File: BN017779.D
 Acq: 08 Dec 2021 01:24

Tgt Ion:	178	Resp:	123174
Ion Ratio	Lower	Upper	
178	100		
176	18.8	15.0	22.6
179	15.6	12.5	18.7









#26
 Anthracene
 Concen: 0.383 ng
 RT: 17.226 min Scan# 1399
 Delta R.T. -0.000 min
 Lab File: BN017779.D
 Acq: 08 Dec 2021 01:24




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:178 Resp: 104990
 Ion Ratio Lower Upper
 178 100
 176 18.2 14.5 21.7
 179 15.4 12.2 18.4



#27
 Fluoranthene-d10
 Concen: 0.407 ng
 RT: 19.129 min Scan# 1688
 Delta R.T. -0.000 min
 Lab File: BN017779.D
 Acq: 08 Dec 2021 01:24

Tgt Ion:212 Resp: 161330
 Ion Ratio Lower Upper
 212 100
 106 10.2 8.2 12.2
 104 5.7 4.6 6.8









#28
Fluoranthene
Concen: 0.417 ng
RT: 19.159 min Scan# 1694
Delta R.T. -0.000 min
Lab File: BN017779.D
Acq: 08 Dec 2021 01:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:202 Resp: 158222
Ion Ratio Lower Upper
202 100
101 9.0 7.2 10.8
203 17.1 13.7 20.5



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.282 min Scan# 1985
Delta R.T. -0.000 min
Lab File: BN017779.D
Acq: 08 Dec 2021 01:24

Tgt Ion:240 Resp: 121707
Ion Ratio Lower Upper
240 100
120 11.0 7.9 11.9
236 27.2 21.4 32.2









#30
 Pyrene
 Concen: 0.421 ng
 RT: 19.522 min Scan# 1767
 Delta R.T. -0.000 min
 Lab File: BN017779.D
 Acq: 08 Dec 2021 01:24




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:202 Resp: 162262
 Ion Ratio Lower Upper
 202 100
 200 20.6 16.5 24.7
 203 17.5 14.0 21.0



#31
 Terphenyl-d14
 Concen: 0.408 ng
 RT: 19.735 min Scan# 1810
 Delta R.T. -0.000 min
 Lab File: BN017779.D
 Acq: 08 Dec 2021 01:24

Tgt Ion:244 Resp: 106630
 Ion Ratio Lower Upper
 244 100
 212 7.4 5.8 8.8
 122 9.5 7.1 10.7









#32
Benzo(a)anthracene
Concen: 0.403 ng
RT: 21.272 min Scan# 1984
Delta R.T. -0.000 min
Lab File: BN017779.D
Acq: 08 Dec 2021 01:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:228 Resp: 151041
Ion Ratio Lower Upper
228 100
226 26.7 21.4 32.0
229 19.9 15.7 23.5



#33
Chrysene
Concen: 0.389 ng
RT: 21.325 min Scan# 1989
Delta R.T. -0.000 min
Lab File: BN017779.D
Acq: 08 Dec 2021 01:24

Tgt Ion:228 Resp: 152293
Ion Ratio Lower Upper
228 100
226 29.3 23.6 35.4
229 19.7 15.6 23.4









#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.416 ng
 RT: 21.219 min Scan# 1979
 Delta R.T. -0.000 min
 Lab File: BN017779.D
 Acq: 08 Dec 2021 01:24




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:	149	Resp:	70658
Ion Ratio	Lower	Upper	
149	100		
167	29.7	23.5	35.3
279	4.8	3.8	5.6



#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.586 min Scan# 2435
 Delta R.T. 0.003 min
 Lab File: BN017779.D
 Acq: 08 Dec 2021 01:24

Tgt Ion:	264	Resp:	103873
Ion Ratio	Lower	Upper	
264	100		
260	24.8	20.0	30.0
265	25.2	20.5	30.7









#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.352 ng
 RT: 25.944 min Scan# 3123
 Delta R.T. 0.003 min
 Lab File: BN017779.D
 Acq: 08 Dec 2021 01:24




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:276 Resp: 101293
 Ion Ratio Lower Upper
 276 100
 138 19.3 15.0 22.4
 277 25.0 19.8 29.8



#37
 Benzo(b)fluoranthene
 Concen: 0.396 ng
 RT: 22.891 min Scan# 2232
 Delta R.T. 0.003 min
 Lab File: BN017779.D
 Acq: 08 Dec 2021 01:24

Tgt Ion:252 Resp: 139700
 Ion Ratio Lower Upper
 252 100
 253 22.5 17.7 26.5
 125 7.9 6.2 9.2









#38
Benzo(k)fluoranthene
Concen: 0.408 ng
RT: 22.935 min Scan# 21
Delta R.T. 0.003 min
Lab File: BN017779.D
Acq: 08 Dec 2021 01:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 136044
Ion Ratio Lower Upper
252 100
253 22.3 17.8 26.6
125 8.0 6.2 9.2



#39
Benzo(a)pyrene
Concen: 0.398 ng
RT: 23.483 min Scan# 2405
Delta R.T. -0.000 min
Lab File: BN017779.D
Acq: 08 Dec 2021 01:24

Tgt Ion:252 Resp: 126366
Ion Ratio Lower Upper
252 100
253 22.8 18.1 27.1
125 9.2 7.0 10.4









#40
 Dibenzo(a,h)anthracene
 Concen: 0.333 ng
 RT: 25.965 min Scan# 3128
 Delta R.T. 0.007 min
 Lab File: BN017779.D
 Acq: 08 Dec 2021 01:24

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Tgt Ion: 278 Resp: 74370
 Ion Ratio Lower Upper
 278 100
 139 14.5 10.8 16.2
 279 24.6 19.1 28.7



#41
 Benzo(g,h,i)perylene
 Concen: 0.360 ng
 RT: 26.663 min Scan# 3334
 Delta R.T. 0.007 min
 Lab File: BN017779.D
 Acq: 08 Dec 2021 01:24

Tgt Ion: 276 Resp: 86319
 Ion Ratio Lower Upper
 276 100
 277 23.9 19.1 28.7
 138 16.2 13.2 19.8

