

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120721\
 Data File : BN017782.D
 Acq On : 08 Dec 2021 03:47
 Operator : CG/JU
 Sample : PB141179BS
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB141179BS

Quant Time: Dec 08 07:35:12 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 08 01:18:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.715	152	23497	0.400	ng	0.00
7) Naphthalene-d8	10.494	136	96370	0.400	ng	0.00
13) Acenaphthene-d10	14.348	164	64452	0.400	ng	0.00
19) Phenanthrene-d10	17.093	188	134213	0.400	ng	0.00
29) Chrysene-d12	21.293	240	132891	0.400	ng	# 0.01
35) Perylene-d12	23.586	264	115098	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.327	112	19050	0.336	ng	0.00
5) Phenol-d6	6.904	99	18118	0.337	ng	0.00
8) Nitrobenzene-d5	8.859	82	19148	0.299	ng	0.00
11) 2-Methylnaphthalene-d10	12.090	152	62168	0.323	ng	0.00
14) 2,4,6-Tribromophenol	15.845	330	9081	0.270	ng	0.00
15) 2-Fluorobiphenyl	12.965	172	80100	0.339	ng	0.00
27) Fluoranthene-d10	19.130	212	148684	0.331	ng	0.00
31) Terphenyl-d14	19.735	244	97445	0.341	ng	0.00
Target Compounds						
						Qvalue
3) n-Nitrosodimethylamine	3.603	42	8221	0.339	ng	99
6) bis(2-Chloroethyl)ether	7.143	93	21791	0.359	ng	99
9) Naphthalene	10.545	128	82883	0.328	ng	100
10) Hexachlorobutadiene	10.836	225	24279	0.331	ng	# 99
12) 2-Methylnaphthalene	12.166	142	59864	0.329	ng	99
16) Acenaphthylene	14.056	152	81024	0.320	ng	100
17) Acenaphthene	14.398	154	57920	0.343	ng	100
18) Fluorene	15.388	166	71957	0.334	ng	100
20) 4,6-Dinitro-2-methylph...	15.502	198	238	0.241	ng	98
21) 4-Bromophenyl-phenylether	16.290	248	25315	0.312	ng	96
22) Hexachlorobenzene	16.399	284	33170	0.344	ng	99
23) Atrazine	16.569	200	16209	0.276	ng	95
24) Pentachlorophenol	16.764	266	3892	0.387	ng	99
25) Phenanthrene	17.129	178	118974	0.337	ng	100
26) Anthracene	17.227	178	99020	0.319	ng	100
28) Fluoranthene	19.159	202	147729	0.344	ng	100
30) Pyrene	19.522	202	150230	0.357	ng	100
32) Benzo(a)anthracene	21.272	228	135671	0.332	ng	100
33) Chrysene	21.325	228	143272	0.335	ng	100
34) Bis(2-ethylhexyl)phtha...	21.219	149	51344	0.277	ng	100
36) Indeno(1,2,3-cd)pyrene	25.944	276	91265	0.286	ng	100
37) Benzo(b)fluoranthene	22.888	252	129980	0.332	ng	100
38) Benzo(k)fluoranthene	22.936	252	127462	0.345	ng	100
39) Benzo(a)pyrene	23.483	252	116075	0.330	ng	100
40) Dibenzo(a,h)anthracene	25.965	278	64626	0.262	ng	98
41) Benzo(g,h,i)perylene	26.660	276	79984	0.301	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120721\
Data File : BN017782.D
Acq On : 08 Dec 2021 03:47
Operator : CG/JU
Sample : PB141179BS
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB141179BS

Quant Time: Dec 08 07:35:12 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 08 01:18:09 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.715 min Scan# 483
Delta R.T. 0.000 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47




Instrument :
BNA_N
ClientSampleId :
PB141179BS

Tgt Ion:152 Resp: 23497
Ion Ratio Lower Upper
152 100
150 151.0 123.0 184.4
115 55.6 45.3 67.9



#3
n-Nitrosodimethylamine
Concen: 0.339 ng
RT: 3.603 min Scan# 37
Delta R.T. -0.000 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47

Tgt Ion: 42 Resp: 8221
Ion Ratio Lower Upper
42 100
74 125.5 101.3 151.9
44 6.3 5.0 7.4









#4
2-Fluorophenol
Concen: 0.336 ng
RT: 5.327 min Scan# 211
Delta R.T. -0.000 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47




Instrument :
BNA_N
ClientSampleId :
PB141179BS


Tgt Ion:112 Resp: 19050
Ion Ratio Lower Upper
112 100
64 61.6 49.3 73.9
63 32.8 25.8 38.8



#5
Phenol-d6
Concen: 0.337 ng
RT: 6.904 min Scan# 395
Delta R.T. -0.000 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47

Tgt Ion: 99 Resp: 18118
Ion Ratio Lower Upper
99 100
42 20.9 16.6 25.0
71 35.9 29.0 43.4









#6
bis(2-Chloroethyl)ether
Concen: 0.359 ng
RT: 7.143 min Scan# 41
Delta R.T. -0.000 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47




Instrument :
BNA_N
ClientSampleId :
PB141179BS


Tgt Ion: 93 Resp: 21791
Ion Ratio Lower Upper
93 100
63 77.2 62.0 93.0
95 32.4 26.4 39.6



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.494 min Scan# 720
Delta R.T. 0.000 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47

Tgt Ion: 136 Resp: 96370
Ion Ratio Lower Upper
136 100
137 10.9 8.8 13.2
54 7.1 5.9 8.9
68 5.7 5.0 7.6









#8
 Nitrobenzene-d5
 Concen: 0.299 ng
 RT: 8.859 min Scan# 591
 Delta R.T. -0.000 min
 Lab File: BN017782.D
 Acq: 08 Dec 2021 03:47




Instrument :
 BNA_N
 ClientSampleId :
 PB141179BS


Tgt Ion	Ratio	Lower	Upper
82	100		
128	39.2	30.2	45.2
54	48.0	39.0	58.6



#9
 Naphthalene
 Concen: 0.328 ng
 RT: 10.545 min Scan# 724
 Delta R.T. -0.000 min
 Lab File: BN017782.D
 Acq: 08 Dec 2021 03:47

Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.8	13.2
127	13.1	10.6	15.8









#10
Hexachlorobutadiene
Concen: 0.331 ng
RT: 10.836 min Scan# 747
Delta R.T. -0.000 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47




Instrument :
BNA_N
ClientSampleId :
PB141179BS


Tgt Ion:225 Resp: 24279
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.3 50.9 76.3



#11
2-Methylnaphthalene-d10
Concen: 0.323 ng
RT: 12.090 min Scan# 929
Delta R.T. -0.000 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47

Tgt Ion:152 Resp: 62168
Ion Ratio Lower Upper
152 100
151 20.5 16.4 24.6





#12

2-Methylnaphthalene

Concen: 0.329 ng

RT: 12.166 min Scan# 945

Delta R.T. 0.004 min

Lab File: BN017782.D

Acq: 08 Dec 2021 03:47

Instrument :

BNA_N

ClientSampleId :

PB141179BS





Tgt Ion:142 Resp: 59864

Ion Ratio Lower Upper

142 100

141 86.6 70.5 105.7

115 34.5 28.0 42.0



#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.348 min Scan# 1168

Delta R.T. -0.000 min

Lab File: BN017782.D

Acq: 08 Dec 2021 03:47




Tgt Ion:164 Resp: 64452


Ion Ratio Lower Upper

164 100

162 98.1 79.0 118.6


160 42.7 34.6 52.0








#14
2,4,6-Tribromophenol
Concen: 0.270 ng
RT: 15.845 min Scan# 1286
Delta R.T. -0.000 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47

Instrument :
BNA_N
ClientSampleId :
PB141179BS







Tgt Ion: 330 Resp: 9081
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 25.9 21.5 32.3



#15
2-Fluorobiphenyl
Concen: 0.339 ng
RT: 12.965 min Scan# 1059
Delta R.T. -0.000 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47

Tgt Ion: 172 Resp: 80100
Ion Ratio Lower Upper
172 100
171 35.4 28.1 42.1
170 23.8 18.9 28.3









#16
Acenaphthylene
Concen: 0.320 ng
RT: 14.056 min Scan# 1145
Delta R.T. -0.000 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47




Instrument :
BNA_N
ClientSampleId :
PB141179BS


Tgt Ion:152 Resp: 81024
Ion Ratio Lower Upper
152 100
151 19.6 15.7 23.5
153 13.1 10.4 15.6



#17
Acenaphthene
Concen: 0.343 ng
RT: 14.398 min Scan# 1172
Delta R.T. -0.000 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47

Tgt Ion:154 Resp: 57920
Ion Ratio Lower Upper
154 100
153 112.7 90.1 135.1
152 55.1 43.7 65.5









#18
 Fluorene
 Concen: 0.334 ng
 RT: 15.388 min Scan# 1250
 Delta R.T. -0.000 min
 Lab File: BN017782.D
 Acq: 08 Dec 2021 03:47




Instrument :
 BNA_N
 ClientSampleId :
 PB141179BS


Tgt Ion:166 Resp: 71957
 Ion Ratio Lower Upper
 166 100
 165 97.6 77.8 116.8
 167 13.5 11.0 16.4



#19
 Phenanthrene-d10
 Concen: 0.400 ng
 RT: 17.093 min Scan# 1388
 Delta R.T. -0.000 min
 Lab File: BN017782.D
 Acq: 08 Dec 2021 03:47

Tgt Ion:188 Resp: 134213
 Ion Ratio Lower Upper
 188 100
 94 0.0 0.0 0.0
 80 7.1 5.1 7.7









#20
4,6-Dinitro-2-methylphenol
Concen: 0.241 ng
RT: 15.502 min Scan# 1259
Delta R.T. 0.013 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47




Instrument :
BNA_N
ClientSampleId :
PB141179BS


Tgt Ion:198 Resp: 238
Ion Ratio Lower Upper
198 100
182 53.7 44.2 66.2
77 0.0 0.0 0.0



#21
4-Bromophenyl-phenylether
Concen: 0.312 ng
RT: 16.290 min Scan# 1322
Delta R.T. 0.000 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47

Tgt Ion:248 Resp: 25315
Ion Ratio Lower Upper
248 100
250 96.9 80.3 120.5
141 41.5 29.8 44.8









#22
Hexachlorobenzene
Concen: 0.344 ng
RT: 16.399 min Scan# 1331
Delta R.T. -0.000 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47




Instrument :
BNA_N
ClientSampleId :
PB141179BS


Tgt Ion:284 Resp: 33170
Ion Ratio Lower Upper
284 100
142 25.6 21.0 31.4
249 29.4 23.9 35.9



#23
Atrazine
Concen: 0.276 ng
RT: 16.569 min Scan# 1345
Delta R.T. 0.012 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47

Tgt Ion:200 Resp: 16209
Ion Ratio Lower Upper
200 100
173 22.6 21.4 32.2
215 51.8 39.3 58.9









#24
 Pentachlorophenol
 Concen: 0.387 ng
 RT: 16.764 min Scan# 1361
 Delta R.T. 0.000 min
 Lab File: BN017782.D
 Acq: 08 Dec 2021 03:47




Instrument :
 BNA_N
 ClientSampleId :
 PB141179BS


Tgt Ion:	266	Resp:	3892
Ion Ratio	Lower	Upper	
266	100		
264	62.8	50.8	76.2
268	63.1	51.8	77.6



#25
 Phenanthrene
 Concen: 0.337 ng
 RT: 17.129 min Scan# 1391
 Delta R.T. -0.000 min
 Lab File: BN017782.D
 Acq: 08 Dec 2021 03:47

Tgt Ion:	178	Resp:	118974
Ion Ratio	Lower	Upper	
178	100		
176	18.7	15.0	22.6
179	15.5	12.5	18.7









#26
 Anthracene
 Concen: 0.319 ng
 RT: 17.227 min Scan# 1399
 Delta R.T. -0.000 min
 Lab File: BN017782.D
 Acq: 08 Dec 2021 03:47




Instrument :
 BNA_N
 ClientSampleId :
 PB141179BS


Tgt Ion:178 Resp: 99020
 Ion Ratio Lower Upper
 178 100
 176 18.1 14.5 21.7
 179 15.4 12.2 18.4



#27
 Fluoranthene-d10
 Concen: 0.331 ng
 RT: 19.130 min Scan# 1688
 Delta R.T. -0.000 min
 Lab File: BN017782.D
 Acq: 08 Dec 2021 03:47

Tgt Ion:212 Resp: 148684
 Ion Ratio Lower Upper
 212 100
 106 10.2 8.2 12.2
 104 5.7 4.6 6.8









#28
Fluoranthene
Concen: 0.344 ng
RT: 19.159 min Scan# 1694
Delta R.T. -0.000 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47




Instrument :
BNA_N
ClientSampleId :
PB141179BS


Tgt Ion:202 Resp: 147729
Ion Ratio Lower Upper
202 100
101 9.0 7.2 10.8
203 17.0 13.7 20.5



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.293 min Scan# 1986
Delta R.T. 0.011 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47

Tgt Ion:240 Resp: 132891
Ion Ratio Lower Upper
240 100
120 5.5 7.9 11.9#
236 25.9 21.4 32.2









#30
Pyrene
Concen: 0.357 ng
RT: 19.522 min Scan# 1767
Delta R.T. -0.000 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47




Instrument :
BNA_N
ClientSampleId :
PB141179BS


Tgt Ion:202 Resp: 150230
Ion Ratio Lower Upper
202 100
200 20.5 16.5 24.7
203 17.5 14.0 21.0



#31
Terphenyl-d14
Concen: 0.341 ng
RT: 19.735 min Scan# 1810
Delta R.T. -0.000 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47

Tgt Ion:244 Resp: 97445
Ion Ratio Lower Upper
244 100
212 7.5 5.8 8.8
122 9.7 7.1 10.7









#32
Benzo(a)anthracene
Concen: 0.332 ng
RT: 21.272 min Scan# 1984
Delta R.T. -0.000 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47




Instrument :
BNA_N
ClientSampleId :
PB141179BS


Tgt Ion:228 Resp: 135671
Ion Ratio Lower Upper
228 100
226 26.7 21.4 32.0
229 19.5 15.7 23.5



#33
Chrysene
Concen: 0.335 ng
RT: 21.325 min Scan# 1989
Delta R.T. -0.000 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47

Tgt Ion:228 Resp: 143272
Ion Ratio Lower Upper
228 100
226 29.5 23.6 35.4
229 19.8 15.6 23.4









#34
Bis(2-ethylhexyl)phthalate
Concen: 0.277 ng
RT: 21.219 min Scan# 1979
Delta R.T. -0.000 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47




Instrument :
BNA_N
ClientSampleId :
PB141179BS


Tgt Ion:149 Resp: 51344
Ion Ratio Lower Upper
149 100
167 29.2 23.5 35.3
279 4.8 3.8 5.6



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.586 min Scan# 2435
Delta R.T. 0.003 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47

Tgt Ion:264 Resp: 115098
Ion Ratio Lower Upper
264 100
260 25.1 20.0 30.0
265 24.6 20.5 30.7









#36
Indeno(1,2,3-cd)pyrene
Concen: 0.286 ng
RT: 25.944 min Scan# 3123
Delta R.T. 0.003 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47




Instrument :
BNA_N
ClientSampleId :
PB141179BS


Tgt Ion:276 Resp: 91265
Ion Ratio Lower Upper
276 100
138 19.0 15.0 22.4
277 24.8 19.8 29.8



#37
Benzo(b)fluoranthene
Concen: 0.332 ng
RT: 22.888 min Scan# 2231
Delta R.T. 0.000 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47

Tgt Ion:252 Resp: 129980
Ion Ratio Lower Upper
252 100
253 22.2 17.7 26.5
125 7.6 6.2 9.2









#38
Benzo(k)fluoranthene
Concen: 0.345 ng
RT: 22.936 min Scan# 21
Delta R.T. 0.003 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47




Instrument :
BNA_N
ClientSampleId :
PB141179BS


Tgt Ion:252 Resp: 127462
Ion Ratio Lower Upper
252 100
253 22.3 17.8 26.6
125 7.7 6.2 9.2



#39
Benzo(a)pyrene
Concen: 0.330 ng
RT: 23.483 min Scan# 2405
Delta R.T. -0.000 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47

Tgt Ion:252 Resp: 116075
Ion Ratio Lower Upper
252 100
253 22.5 18.1 27.1
125 9.1 7.0 10.4









#40
Dibenzo(a,h)anthracene
Concen: 0.262 ng
RT: 25.965 min Scan# 3128
Delta R.T. 0.007 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47

Instrument :
BNA_N
ClientSampleId :
PB141179BS

Tgt Ion:278 Resp: 64626
Ion Ratio Lower Upper
278 100
139 14.1 10.8 16.2
279 24.9 19.1 28.7



#41
Benzo(g,h,i)perylene
Concen: 0.301 ng
RT: 26.660 min Scan# 3333
Delta R.T. 0.003 min
Lab File: BN017782.D
Acq: 08 Dec 2021 03:47

Tgt Ion:276 Resp: 79984
Ion Ratio Lower Upper
276 100
277 23.9 19.1 28.7
138 16.5 13.2 19.8

