

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120721\
 Data File : BN017761.D
 Acq On : 07 Dec 2021 12:47
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Dec 08 00:22:32 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 08 00:20:56 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.715	152	21143	0.400	ng	0.00
7) Naphthalene-d8	10.494	136	91241	0.400	ng	0.00
13) Acenaphthene-d10	14.347	164	59374	0.400	ng	0.00
19) Phenanthrene-d10	17.093	188	127005	0.400	ng	0.00
29) Chrysene-d12	21.282	240	132067	0.400	ng	0.00
35) Perylene-d12	23.582	264	120687	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.327	112	22331	0.492	ng	0.00
5) Phenol-d6	6.903	99	23367	0.481	ng	0.00
8) Nitrobenzene-d5	8.859	82	25118	0.407	ng	0.00
11) 2-Methylnaphthalene-d10	12.090	152	68002	0.407	ng	0.00
14) 2,4,6-Tribromophenol	15.832	330	12611	0.443	ng	-0.01
15) 2-Fluorobiphenyl	12.964	172	88703	0.403	ng	0.00
27) Fluoranthene-d10	19.124	212	172797	0.404	ng	0.00
31) Terphenyl-d14	19.730	244	115404	0.341	ng	0.00
Target Compounds						
						Qvalue
3) n-Nitrosodimethylamine	3.602	42	9341	0.467	ng	98
6) bis(2-Chloroethyl)ether	7.143	93	23927	0.451	ng	100
9) Naphthalene	10.545	128	95641	0.423	ng	100
10) Hexachlorobutadiene	10.836	225	25809	0.354	ng	# 99
12) 2-Methylnaphthalene	12.161	142	64916	0.418	ng	100
16) Acenaphthylene	14.056	152	92752	0.407	ng	100
17) Acenaphthene	14.398	154	62884	0.407	ng	99
18) Fluorene	15.388	166	81101	0.410	ng	99
20) 4,6-Dinitro-2-methylph...	15.477	198	885	0.302	ng	# 81
21) 4-Bromophenyl-phenylether	16.289	248	30774	0.404	ng	96
22) Hexachlorobenzene	16.399	284	36559	0.392	ng	100
23) Atrazine	16.557	200	22298	0.423	ng	99
24) Pentachlorophenol	16.752	266	9676	0.507	ng	100
25) Phenanthrene	17.129	178	131952	0.394	ng	100
26) Anthracene	17.214	178	117578	0.402	ng	100
28) Fluoranthene	19.154	202	168163	0.415	ng	100
30) Pyrene	19.517	202	169907	0.342	ng	100
32) Benzo(a)anthracene	21.272	228	162114	0.409	ng	100
33) Chrysene	21.325	228	175470	0.416	ng	99
34) Bis(2-ethylhexyl)phtha...	21.219	149	77192	0.370	ng	100
36) Indeno(1,2,3-cd)pyrene	25.937	276	135295	0.536	ng	100
37) Benzo(b)fluoranthene	22.884	252	163912	0.386	ng	100
38) Benzo(k)fluoranthene	22.932	252	159606	0.408	ng	100
39) Benzo(a)pyrene	23.480	252	148457	0.433	ng	100
40) Dibenzo(a,h)anthracene	25.958	278	107113	0.608	ng	100
41) Benzo(g,h,i)perylene	26.656	276	111901	0.441	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120721\
Data File : BN017761.D
Acq On : 07 Dec 2021 12:47
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Dec 08 00:22:32 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 08 00:20:56 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.715 min Scan# 41
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 21143
Ion Ratio Lower Upper
152 100
150 154.7 123.0 184.4
115 57.8 45.3 67.9



#3
n-Nitrosodimethylamine
Concen: 0.467 ng
RT: 3.602 min Scan# 37
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47

Tgt Ion: 42 Resp: 9341
Ion Ratio Lower Upper
42 100
74 128.9 101.6 152.4
44 7.7 5.1 7.7









#4
2-Fluorophenol
Concen: 0.492 ng
RT: 5.327 min Scan# 211
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:112 Resp: 22331
Ion Ratio Lower Upper
112 100
64 62.2 49.3 73.9
63 33.0 25.8 38.8



#5
Phenol-d6
Concen: 0.481 ng
RT: 6.903 min Scan# 395
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47

Tgt Ion: 99 Resp: 23367
Ion Ratio Lower Upper
99 100
42 20.9 16.6 25.0
71 36.1 29.0 43.4









#6
bis(2-Chloroethyl)ether
Concen: 0.451 ng
RT: 7.143 min Scan# 41
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion: 93 Resp: 23927




Ion	Ratio	Lower	Upper
93	100		
63	77.4	62.0	93.0
95	32.4	26.4	39.6




#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.494 min Scan# 720
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47

Tgt Ion: 136 Resp: 91241

Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.8	13.2
54	6.9	5.9	8.9
68	5.9	5.0	7.6









#8
Nitrobenzene-d5
Concen: 0.407 ng
RT: 8.859 min Scan# 591
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 82 Resp: 25118
Ion Ratio Lower Upper
82 100
128 39.2 30.2 45.2
54 47.5 39.0 58.6



#9
Naphthalene
Concen: 0.423 ng
RT: 10.545 min Scan# 724
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47


Tgt Ion: 128 Resp: 95641
Ion Ratio Lower Upper
128 100
129 11.1 8.8 13.2
127 13.1 10.6 15.8








#10
Hexachlorobutadiene
Concen: 0.354 ng
RT: 10.836 min Scan# 747
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:225 Resp: 25809
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.5 50.9 76.3



#11
2-Methylnaphthalene-d10
Concen: 0.407 ng
RT: 12.090 min Scan# 929
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47

Tgt Ion:152 Resp: 68002
Ion Ratio Lower Upper
152 100
151 20.3 16.4 24.6









#12
2-Methylnaphthalene
Concen: 0.418 ng
RT: 12.161 min Scan# 945
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:142 Resp: 64916
Ion Ratio Lower Upper
142 100
141 88.4 70.5 105.7
115 35.3 28.0 42.0



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.347 min Scan# 1168
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47

Tgt Ion:164 Resp: 59374
Ion Ratio Lower Upper
164 100
162 98.1 79.0 118.6
160 42.6 34.6 52.0









#14
2,4,6-Tribromophenol
Concen: 0.443 ng
RT: 15.832 min Scan# 1285
Delta R.T. -0.013 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:330 Resp: 12611
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 27.4 21.5 32.3



#15
2-Fluorobiphenyl
Concen: 0.403 ng
RT: 12.964 min Scan# 1059
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47

Tgt Ion:172 Resp: 88703
Ion Ratio Lower Upper
172 100
171 35.5 28.1 42.1
170 23.8 18.9 28.3









#16
Acenaphthylene
Concen: 0.407 ng
RT: 14.056 min Scan# 1145
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 92752
Ion Ratio Lower Upper
152 100
151 19.4 15.7 23.5
153 13.0 10.4 15.6



#17
Acenaphthene
Concen: 0.407 ng
RT: 14.398 min Scan# 1172
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47

Tgt Ion:154 Resp: 62884
Ion Ratio Lower Upper
154 100
153 112.3 90.5 135.7
152 54.6 44.0 66.0









#18
Fluorene
Concen: 0.410 ng
RT: 15.388 min Scan# 1250
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:166 Resp: 81101
Ion Ratio Lower Upper
166 100
165 97.6 77.7 116.5
167 13.3 11.0 16.4



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.093 min Scan# 1388
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47

Tgt Ion:188 Resp: 127005
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 5.8 5.1 7.7









#20
4,6-Dinitro-2-methylphenol
Concen: 0.302 ng
RT: 15.477 min Scan# 1257
Delta R.T. -0.013 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:198 Resp: 885
Ion Ratio Lower Upper
198 100
182 41.4 44.2 66.2#
77 0.0 0.0 0.0



#21
4-Bromophenyl-phenylether
Concen: 0.404 ng
RT: 16.289 min Scan# 1322
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47

Tgt Ion:248 Resp: 30774
Ion Ratio Lower Upper
248 100
250 102.3 80.3 120.5
141 30.6 29.8 44.8









#22
Hexachlorobenzene
Concen: 0.392 ng
RT: 16.399 min Scan# 1331
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:284 Resp: 36559
Ion Ratio Lower Upper
284 100
142 26.0 20.9 31.3
249 29.7 23.8 35.8



#23
Atrazine
Concen: 0.423 ng
RT: 16.557 min Scan# 1344
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47

Tgt Ion:200 Resp: 22298
Ion Ratio Lower Upper
200 100
173 26.5 21.4 32.2
215 49.4 39.3 58.9









#24
Pentachlorophenol
Concen: 0.507 ng
RT: 16.752 min Scan# 1360
Delta R.T. -0.012 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	266	Resp:	9676
Ion Ratio	Lower	Upper	
266	100		
264	63.2	50.8	76.2
268	64.6	51.8	77.6



#25
Phenanthrene
Concen: 0.394 ng
RT: 17.129 min Scan# 1391
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47

Tgt Ion:	178	Resp:	131952
Ion Ratio	Lower	Upper	
178	100		
176	18.8	15.0	22.6
179	15.7	12.5	18.7









#26
 Anthracene
 Concen: 0.402 ng
 RT: 17.214 min Scan# 11
 Delta R.T. -0.012 min
 Lab File: BN017761.D
 Acq: 07 Dec 2021 12:47




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:178 Resp: 117578
 Ion Ratio Lower Upper
 178 100
 176 18.4 14.5 21.7
 179 15.3 12.2 18.4



#27
 Fluoranthene-d10
 Concen: 0.404 ng
 RT: 19.124 min Scan# 1687
 Delta R.T. -0.005 min
 Lab File: BN017761.D
 Acq: 07 Dec 2021 12:47

Tgt Ion:212 Resp: 172797
 Ion Ratio Lower Upper
 212 100
 106 10.3 8.2 12.2
 104 5.7 4.6 6.8









#28
Fluoranthene
Concen: 0.415 ng
RT: 19.154 min Scan# 1694
Delta R.T. -0.005 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:202 Resp: 168163
Ion Ratio Lower Upper
202 100
101 9.0 7.2 10.8
203 17.2 13.7 20.5



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.282 min Scan# 1985
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47

Tgt Ion:240 Resp: 132067
Ion Ratio Lower Upper
240 100
120 8.8 7.9 11.9
236 26.8 21.4 32.2









#30
Pyrene
Concen: 0.342 ng
RT: 19.517 min Scan# 11
Delta R.T. -0.005 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:202 Resp: 169907
Ion Ratio Lower Upper
202 100
200 20.5 16.5 24.7
203 17.5 14.0 21.0



#31
Terphenyl-d14
Concen: 0.341 ng
RT: 19.730 min Scan# 1809
Delta R.T. -0.005 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47

Tgt Ion:244 Resp: 115404
Ion Ratio Lower Upper
244 100
212 7.5 5.8 8.8
122 9.6 7.1 10.7









#32
Benzo(a)anthracene
Concen: 0.409 ng
RT: 21.272 min Scan# 1984
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:228 Resp: 162114
Ion Ratio Lower Upper
228 100
226 26.8 21.4 32.0
229 20.0 15.7 23.5



#33
Chrysene
Concen: 0.416 ng
RT: 21.325 min Scan# 1989
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47

Tgt Ion:228 Resp: 175470
Ion Ratio Lower Upper
228 100
226 28.5 23.6 35.4
229 19.8 15.6 23.4









#34
Bis(2-ethylhexyl)phthalate
Concen: 0.370 ng
RT: 21.219 min Scan# 1979
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:149 Resp: 77192
Ion Ratio Lower Upper
149 100
167 29.6 23.8 35.6
279 4.8 3.8 5.8



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.582 min Scan# 2434
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47

Tgt Ion:264 Resp: 120687
Ion Ratio Lower Upper
264 100
260 24.6 20.0 30.0
265 25.8 20.5 30.7









#36
Indeno(1,2,3-cd)pyrene
Concen: 0.536 ng
RT: 25.937 min Scan# 31
Delta R.T. -0.004 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:276 Resp: 135295
Ion Ratio Lower Upper
276 100
138 18.9 14.9 22.3
277 24.9 19.8 29.8



#37
Benzo(b)fluoranthene
Concen: 0.386 ng
RT: 22.884 min Scan# 2230
Delta R.T. -0.004 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47

Tgt Ion:252 Resp: 163912
Ion Ratio Lower Upper
252 100
253 22.2 17.7 26.5
125 7.9 6.2 9.2









#38
Benzo(k)fluoranthene
Concen: 0.408 ng
RT: 22.932 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 159606
Ion Ratio Lower Upper
252 100
253 22.3 17.8 26.6
125 7.7 6.2 9.2



#39
Benzo(a)pyrene
Concen: 0.433 ng
RT: 23.480 min Scan# 2404
Delta R.T. -0.004 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47

Tgt Ion:252 Resp: 148457
Ion Ratio Lower Upper
252 100
253 22.5 18.1 27.1
125 8.8 7.0 10.4









#40
Dibenzo(a,h)anthracene
Concen: 0.608 ng
RT: 25.958 min Scan# 3128
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:278 Resp: 107113
Ion Ratio Lower Upper
278 100
139 13.2 10.8 16.2
279 24.0 19.1 28.7



#41
Benzo(g,h,i)perylene
Concen: 0.441 ng
RT: 26.656 min Scan# 3332
Delta R.T. -0.000 min
Lab File: BN017761.D
Acq: 07 Dec 2021 12:47

Tgt Ion:276 Resp: 111901
Ion Ratio Lower Upper
276 100
277 23.8 19.1 28.7
138 16.0 13.2 19.8

