

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120722\
 Data File : BN023090.D
 Acq On : 07 Dec 2022 17:05
 Operator : CG/JU
 Sample : PB149143BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB149143BL

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 12/08/2022
 Supervised By :mohammad ahmed 12/08/2022

Quant Time: Dec 08 03:21:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 08 02:58:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.020	152	6436	0.400	ng	0.00
7) Naphthalene-d8	10.840	136	18011	0.400	ng	0.00
13) Acenaphthene-d10	14.666	164	9004	0.400	ng	0.00
19) Phenanthrene-d10	17.414	188	21354	0.400	ng	0.00
29) Chrysene-d12	21.598	240	15201	0.400	ng	# 0.00
35) Perylene-d12	24.062	264	12107	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.558	112	6166	0.414	ng	0.00
5) Phenol-d6	7.168	99	7603	0.406	ng	0.00
8) Nitrobenzene-d5	9.185	82	4705m	0.348	ng	0.00
11) 2-Methylnaphthalene-d10	12.427	152	11202	0.329	ng	0.00
14) 2,4,6-Tribromophenol	16.148	330	1346	0.354	ng	0.00
15) 2-Fluorobiphenyl	13.298	172	15780	0.395	ng	0.00
27) Fluoranthene-d10	19.439	212	19729	0.337	ng	0.00
31) Terphenyl-d14	20.031	244	13248	0.459	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120722\
Data File : BN023090.D
Acq On : 07 Dec 2022 17:05
Operator : CG/JU
Sample : PB149143BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB149143BL

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 12/08/2022
Supervised By :mohammad ahmed 12/08/2022

Quant Time: Dec 08 03:21:59 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 08 02:58:10 2022
Response via : Initial Calibration

