

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120821\
 Data File : BN017824.D
 Acq On : 09 Dec 2021 07:35
 Operator : CG/JU
 Sample : M4922-12
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DUP02-20211130

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 12/10/2021
 Supervised By :mohammad ahmed 12/15/2021

Quant Time: Dec 10 01:52:06 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 08 01:18:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.714	152	20974	0.400	ng	0.00
7) Naphthalene-d8	10.494	136	83810	0.400	ng	0.00
13) Acenaphthene-d10	14.347	164	58193	0.400	ng	0.00
19) Phenanthrene-d10	17.092	188	122187	0.400	ng	0.00
29) Chrysene-d12	21.293	240	102365	0.400	ng	# 0.00
35) Perylene-d12	23.589	264	74846	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.335	112	6024	0.119	ng	0.00
5) Phenol-d6	6.903	99	3247	0.068	ng	0.00
8) Nitrobenzene-d5	8.859	82	14927	0.268	ng	0.00
11) 2-Methylnaphthalene-d10	12.089	152	41957	0.251	ng	0.00
14) 2,4,6-Tribromophenol	15.844	330	2323	0.076	ng	0.00
15) 2-Fluorobiphenyl	12.964	172	64020	0.300	ng	0.00
27) Fluoranthene-d10	19.129	212	133707	0.327	ng	0.00
31) Terphenyl-d14	19.735	244	122187	0.556	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.270	88	30880m	4.000	ng	Qvalue
6) bis(2-Chloroethyl)ether	7.143	93	2835	0.052	ng	97
34) Bis(2-ethylhexyl)phtha...	21.218	149	19791	0.139	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120821\
Data File : BN017824.D
Acq On : 09 Dec 2021 07:35
Operator : CG/JU
Sample : M4922-12
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DUP02-20211130

Quant Time: Dec 10 01:52:06 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 08 01:18:09 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 12/10/2021
Supervised By :mohammad ahmed 12/15/2021

