

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120821\
 Data File : BN017831.D
 Acq On : 09 Dec 2021 11:46
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Dec 10 01:57:09 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 08 01:18:09 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.715	152	23775	0.400	ng	0.00
7) Naphthalene-d8	10.494	136	100286	0.400	ng	0.00
13) Acenaphthene-d10	14.347	164	64577	0.400	ng	0.00
19) Phenanthrene-d10	17.092	188	134428	0.400	ng	0.00
29) Chrysene-d12	21.293	240	122862	0.400	ng	# 0.01
35) Perylene-d12	23.589	264	95188	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.327	112	21701	0.378	ng	0.00
5) Phenol-d6	6.903	99	21742	0.400	ng	0.00
8) Nitrobenzene-d5	8.859	82	24407	0.366	ng	0.00
11) 2-Methylnaphthalene-d10	12.090	152	73941	0.370	ng	0.00
14) 2,4,6-Tribromophenol	15.844	330	7461	0.221	ng	0.00
15) 2-Fluorobiphenyl	12.977	172	94090	0.398	ng	0.01
27) Fluoranthene-d10	19.129	212	177069	0.394	ng	0.00
31) Terphenyl-d14	19.735	244	118374	0.449	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.270	88	2869	0.328	ng	# 100
3) n-Nitrosodimethylamine	3.602	42	9747	0.397	ng	98
6) bis(2-Chloroethyl)ether	7.143	93	25670	0.417	ng	99
9) Naphthalene	10.545	128	99764	0.380	ng	100
10) Hexachlorobutadiene	10.836	225	28901	0.379	ng	# 99
12) 2-Methylnaphthalene	12.165	142	70634	0.373	ng	100
16) Acenaphthylene	14.055	152	96836	0.382	ng	100
17) Acenaphthene	14.411	154	68727	0.407	ng	100
18) Fluorene	15.400	166	85514	0.396	ng	100
20) 4,6-Dinitro-2-methylph...	15.514	198	93	0.149	ng	# 85
21) 4-Bromophenyl-phenylether	16.289	248	30423	0.374	ng	# 93
22) Hexachlorobenzene	16.411	284	39558	0.409	ng	98
23) Atrazine	16.569	200	19858	0.338	ng	94
24) Pentachlorophenol	16.764	266	1685	0.326	ng	99
25) Phenanthrene	17.129	178	138279	0.392	ng	99
26) Anthracene	17.226	178	117713	0.379	ng	100
28) Fluoranthene	19.159	202	176010	0.409	ng	100
30) Pyrene	19.521	202	188155	0.483	ng	100
32) Benzo(a)anthracene	21.272	228	143009	0.378	ng	99
33) Chrysene	21.325	228	158468	0.401	ng	99
34) Bis(2-ethylhexyl)phtha...	21.219	149	58163	0.339	ng	100
36) Indeno(1,2,3-cd)pyrene	25.951	276	75114	0.285	ng	99
37) Benzo(b)fluoranthene	22.894	252	130166	0.402	ng	100
38) Benzo(k)fluoranthene	22.939	252	123150	0.403	ng	100
39) Benzo(a)pyrene	23.490	252	111272	0.383	ng	# 98
40) Dibenzo(a,h)anthracene	25.978	278	50788	0.249	ng	99
41) Benzo(g,h,i)perylene	26.670	276	67253	0.306	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN120821\
Data File : BN017831.D
Acq On : 09 Dec 2021 11:46
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Quant Time: Dec 10 01:57:09 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 08 01:18:09 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.715 min Scan# 49
Delta R.T. -0.000 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:152 Resp: 23775
Ion Ratio Lower Upper
152 100
150 154.4 123.0 184.4
115 56.2 45.3 67.9




#2
1,4-Dioxane
Concen: 0.328 ng
RT: 3.270 min Scan# 10
Delta R.T. -0.028 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46

Tgt Ion: 88 Resp: 2869
Ion Ratio Lower Upper
88 100
43 35.6 0.0 0.0#
58 80.2 0.0 0.0#







Abundance Scan 37 (3.603 min): BN017764.D\data.ms (-32)



#3
n-Nitrosodimethylamine
Concen: 0.397 ng
RT: 3.602 min Scan# 41
Delta R.T. -0.000 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 42 Resp: 9747
Ion Ratio Lower Upper
42 100
74 123.7 101.3 151.9
44 6.4 5.0 7.4



#4
2-Fluorophenol
Concen: 0.378 ng
RT: 5.327 min Scan# 233
Delta R.T. -0.000 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46

Tgt Ion: 112 Resp: 21701
Ion Ratio Lower Upper
112 100
64 61.7 49.3 73.9
63 33.1 25.8 38.8









#5
Phenol-d6
Concen: 0.400 ng
RT: 6.903 min Scan# 40
Delta R.T. -0.000 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 99 Resp: 21742
Ion Ratio Lower Upper
99 100
42 21.4 16.6 25.0
71 36.5 29.0 43.4



#6
bis(2-Chloroethyl)ether
Concen: 0.417 ng
RT: 7.143 min Scan# 430
Delta R.T. -0.000 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46

Tgt Ion: 93 Resp: 25670
Ion Ratio Lower Upper
93 100
63 77.1 62.0 93.0
95 32.4 26.4 39.6









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.494 min Scan# 71
Delta R.T. -0.000 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:136 Resp: 100286
Ion Ratio Lower Upper
136 100
137 11.0 8.8 13.2
54 7.2 5.9 8.9
68 5.9 5.0 7.6



#8
Nitrobenzene-d5
Concen: 0.366 ng
RT: 8.859 min Scan# 600
Delta R.T. -0.000 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46

Tgt Ion: 82 Resp: 24407
Ion Ratio Lower Upper
82 100
128 39.0 30.2 45.2
54 48.4 39.0 58.6









#9
Naphthalene
Concen: 0.380 ng
RT: 10.545 min Scan# 71
Delta R.T. -0.000 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:128 Resp: 99764
Ion Ratio Lower Upper
128 100
129 11.0 8.8 13.2
127 13.0 10.6 15.8



#10
Hexachlorobutadiene
Concen: 0.379 ng
RT: 10.836 min Scan# 756
Delta R.T. -0.000 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46

Tgt Ion:225 Resp: 28901
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.1 50.9 76.3





#11

2-Methylnaphthalene-d10

Concen: 0.370 ng

RT: 12.090 min Scan# 91

Delta R.T. -0.000 min

Lab File: BN017831.D

Acq: 09 Dec 2021 11:46

Instrument :

BNA_N

ClientSampleId :





SSTDCCC0.4EC

Tgt Ion:152 Resp: 73941

Ion Ratio Lower Upper

152 100

151 20.3 16.4 24.6



#12

2-Methylnaphthalene

Concen: 0.373 ng

RT: 12.165 min Scan# 955

Delta R.T. 0.004 min

Lab File: BN017831.D

Acq: 09 Dec 2021 11:46




Tgt Ion:142 Resp: 70634


Ion Ratio Lower Upper

142 100

141 88.0 70.5 105.7

115 35.0 28.0 42.0









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.347 min Scan# 1168
Delta R.T. -0.000 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:164 Resp: 64577
Ion Ratio Lower Upper
164 100
162 97.5 79.0 118.6
160 43.5 34.6 52.0



#14
2,4,6-Tribromophenol
Concen: 0.221 ng
RT: 15.844 min Scan# 1295
Delta R.T. -0.000 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46

Tgt Ion:330 Resp: 7461
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 25.6 21.5 32.3









#15
2-Fluorobiphenyl
Concen: 0.398 ng
RT: 12.977 min Scan# 1059
Delta R.T. 0.012 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	172	Resp:	94090
Ion Ratio	Lower	Upper	
172	100		
171	33.4	28.1	42.1
170	21.2	18.9	28.3



#16
Acenaphthylene
Concen: 0.382 ng
RT: 14.055 min Scan# 1154
Delta R.T. -0.000 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46

Tgt Ion:	152	Resp:	96836
Ion Ratio	Lower	Upper	
152	100		
151	19.6	15.7	23.5
153	13.1	10.4	15.6









#17
Acenaphthene
Concen: 0.407 ng
RT: 14.411 min Scan# 1172
Delta R.T. 0.012 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:154 Resp: 68727
Ion Ratio Lower Upper
154 100
153 112.4 90.1 135.1
152 54.5 43.7 65.5



#18
Fluorene
Concen: 0.396 ng
RT: 15.400 min Scan# 1260
Delta R.T. 0.012 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46

Tgt Ion:166 Resp: 85514
Ion Ratio Lower Upper
166 100
165 97.0 77.8 116.8
167 13.6 11.0 16.4









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.092 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:188 Resp: 134428
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.6 5.1 7.7



#20
4,6-Dinitro-2-methylphenol
Concen: 0.149 ng
RT: 15.514 min Scan# 1269
Delta R.T. 0.025 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46

Tgt Ion:198 Resp: 93
Ion Ratio Lower Upper
198 100
182 66.3 44.2 66.2#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.374 ng
RT: 16.289 min Scan# 1341
Delta R.T. -0.000 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:248 Resp: 30423
Ion Ratio Lower Upper
248 100
250 96.5 80.3 120.5
141 46.1 29.8 44.8#



#22
Hexachlorobenzene
Concen: 0.409 ng
RT: 16.411 min Scan# 1341
Delta R.T. 0.012 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46

Tgt Ion:284 Resp: 39558
Ion Ratio Lower Upper
284 100
142 25.8 21.0 31.4
249 28.7 23.9 35.9









#23
Atrazine
Concen: 0.338 ng
RT: 16.569 min Scan# 11
Delta R.T. 0.012 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:200 Resp: 19858
Ion Ratio Lower Upper
200 100
173 21.8 21.4 32.2
215 51.8 39.3 58.9



#24
Pentachlorophenol
Concen: 0.326 ng
RT: 16.764 min Scan# 1370
Delta R.T. -0.000 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46

Tgt Ion:266 Resp: 1685
Ion Ratio Lower Upper
266 100
264 64.6 50.8 76.2
268 65.1 51.8 77.6









#25
Phenanthrene
Concen: 0.392 ng
RT: 17.129 min Scan# 1400
Delta R.T. -0.000 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:178 Resp: 138279
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.6
179 16.2 12.5 18.7



#26
Anthracene
Concen: 0.379 ng
RT: 17.226 min Scan# 1408
Delta R.T. -0.000 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46

Tgt Ion:178 Resp: 117713
Ion Ratio Lower Upper
178 100
176 18.3 14.5 21.7
179 15.6 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.394 ng
RT: 19.129 min Scan# 1703
Delta R.T. -0.000 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:212 Resp: 177069
Ion Ratio Lower Upper
212 100
106 10.2 8.2 12.2
104 5.6 4.6 6.8



#28
Fluoranthene
Concen: 0.409 ng
RT: 19.159 min Scan# 1703
Delta R.T. -0.000 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46

Tgt Ion:202 Resp: 176010
Ion Ratio Lower Upper
202 100
101 9.2 7.2 10.8
203 17.2 13.7 20.5









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.293 min Scan# 1985
Delta R.T. 0.010 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:240 Resp: 122862
Ion Ratio Lower Upper
240 100
120 6.1 7.9 11.9#
236 25.8 21.4 32.2



#30
Pyrene
Concen: 0.483 ng
RT: 19.521 min Scan# 1776
Delta R.T. -0.000 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46

Tgt Ion:202 Resp: 188155
Ion Ratio Lower Upper
202 100
200 20.5 16.5 24.7
203 17.5 14.0 21.0









#31
 Terphenyl-d14
 Concen: 0.449 ng
 RT: 19.735 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: BN017831.D
 Acq: 09 Dec 2021 11:46




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:244 Resp: 118374
 Ion Ratio Lower Upper
 244 100
 212 7.5 5.8 8.8
 122 9.5 7.1 10.7



#32
 Benzo(a)anthracene
 Concen: 0.378 ng
 RT: 21.272 min Scan# 1993
 Delta R.T. -0.000 min
 Lab File: BN017831.D
 Acq: 09 Dec 2021 11:46

Tgt Ion:228 Resp: 143009
 Ion Ratio Lower Upper
 228 100
 226 27.2 21.4 32.0
 229 19.6 15.7 23.5









#33
Chrysene
Concen: 0.401 ng
RT: 21.325 min Scan# 1989
Delta R.T. -0.000 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:228 Resp: 158468
Ion Ratio Lower Upper
228 100
226 30.3 23.6 35.4
229 19.4 15.6 23.4



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.339 ng
RT: 21.219 min Scan# 1988
Delta R.T. -0.000 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46

Tgt Ion:149 Resp: 58163
Ion Ratio Lower Upper
149 100
167 29.5 23.5 35.3
279 4.8 3.8 5.6









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.589 min Scan# 24
Delta R.T. 0.007 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:264 Resp: 95188
Ion Ratio Lower Upper
264 100
260 25.0 20.0 30.0
265 22.7 20.5 30.7



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.285 ng
RT: 25.951 min Scan# 3135
Delta R.T. 0.010 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46

Tgt Ion:276 Resp: 75114
Ion Ratio Lower Upper
276 100
138 20.1 15.0 22.4
277 24.9 19.8 29.8









#37
Benzo(b)fluoranthene
Concen: 0.402 ng
RT: 22.894 min Scan# 211
Delta R.T. 0.006 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 130166
Ion Ratio Lower Upper
252 100
253 22.2 17.7 26.5
125 8.0 6.2 9.2



#38
Benzo(k)fluoranthene
Concen: 0.403 ng
RT: 22.939 min Scan# 2255
Delta R.T. 0.007 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46

Tgt Ion:252 Resp: 123150
Ion Ratio Lower Upper
252 100
253 22.2 17.8 26.6
125 7.9 6.2 9.2









#39
Benzo(a)pyrene
Concen: 0.383 ng
RT: 23.490 min Scan# 2405
Delta R.T. 0.006 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 111272
Ion Ratio Lower Upper
252 100
253 22.3 18.1 27.1
125 11.2 7.0 10.4#



#40
Dibenzo(a,h)anthracene
Concen: 0.249 ng
RT: 25.978 min Scan# 3143
Delta R.T. 0.020 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46


Tgt Ion:278 Resp: 50788
Ion Ratio Lower Upper
278 100
139 14.0 10.8 16.2
279 24.1 19.1 28.7





#41
Benzo(g,h,i)perylene
Concen: 0.306 ng
RT: 26.670 min Scan# 31
Delta R.T. 0.013 min
Lab File: BN017831.D
Acq: 09 Dec 2021 11:46

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:276 Resp: 67253
Ion Ratio Lower Upper
276 100
277 23.7 19.1 28.7
138 16.1 13.2 19.8

