




Data File : BN003789.D
Acq On : 10 Dec 2018 21:49
Operator : JU/SJ
Sample : J6077-11
Misc :
ALS Vial : 15 Sample Multiplier: 1

Quant Time: Dec 11 04:34:04 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 11 04:22:54 2018
Response via : Initial Calibration



TIC: BN003789.D

(15) Fluoranthene (C)

19.267min (-0.000) 3.27ng/ul m

response 348304

Ion	Exp%	Act%
202.00	100	100
101.00	10.70	11.29
100.00	8.10	8.36
0.00	0.00	0.00