




Data File : BN003791.D
Acq On : 10 Dec 2018 22:59
Operator : JU/SJ
Sample : J6077-12
Misc :
ALS Vial : 17 Sample Multiplier: 1

Quant Time: Dec 11 04:41:40 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 11 04:22:54 2018
Response via : Initial Calibration



TIC: BN003791.D

(15) Fluoranthene (C)

19.267min (-19.267) 0.00ng/ul

response 0

Ion	Exp%	Act%
202.00	100	0.00
101.00	10.70	0.00
100.00	8.10	0.00
0.00	0.00	0.00