




Data File : BN003801.D
Acq On : 11 Dec 2018 04:49
Operator : JU/SJ
Sample : J6077-13
Misc :
ALS Vial : 27 Sample Multiplier: 1

Quant Time: Dec 11 08:17:06 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 11 04:22:54 2018
Response via : Initial Calibration



TIC: BN003801.D

(13) Anthracene

17.411min (+0.063) 0.00ng/ul

response 196

Ion	Exp%	Act%
178.00	100	100
179.00	15.40	235.49#
176.00	18.50	39.59#
0.00	0.00	0.00