

Data File : BN003778.D
Acq On : 10 Dec 2018 14:47
Operator : JU/SJ
Sample : PB115353BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :

Quant Time: Dec 11 04:03:14 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 11 03:42:06 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	5418	0.40	ng/ul	0.00
2) Naphthalene-d8	10.63	136	22322	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	12565	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.21	188	26370m	0.40	ng/ul	0.00
16) Chrysene-d12	21.40	240	22164m	0.40	ng/ul	0.00
20) Perylene-d12	23.71	264	18563m	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.22	152	11146	0.32	ng/ul	0.00
14) Fluoranthene-d10	19.24	212	22942m	0.30	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BN003778.D
Acq On : 10 Dec 2018 14:47
Operator : JU/SJ
Sample : PB115353BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :

Quant Time: Dec 11 04:03:14 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 11 03:42:06 2018
Response via : Initial Calibration

