

Data File : BN003801.D
Acq On : 11 Dec 2018 04:49
Operator : JU/SJ
Sample : J6077-13
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :

Quant Time: Dec 11 08:21:02 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 11 04:22:54 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	5346	0.40	ng/ul	0.00
2) Naphthalene-d8	10.63	136	20841	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	10032	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.22	188	18793m	0.40	ng/ul	0.00
16) Chrysene-d12	21.40	240	14409m	0.40	ng/ul	0.00
20) Perylene-d12	23.71	264	15409m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.22	152	7342	0.23	ng/ul	0.00
14) Fluoranthene-d10	19.24	212	12016m	0.22	ng/ul	0.00
Target Compounds				Qvalue		
12) Phenanthrene	17.26	178	15721	0.255	ng/ul	98
13) Anthracene	17.35	178	1913m	0.038	ng/ul	
15) Fluoranthene	19.27	202	38257m	0.525	ng/ul	
17) Pyrene	19.64	202	32408	0.627	ng/ul	99
18) Benzo(a)anthracene	21.38	228	12824m	0.282	ng/ul	
19) Chrysene	21.43	228	16889m	0.324	ng/ul	
21) Benzo(b)fluoranthene	23.02	252	25195m	0.395	ng/ul	
22) Benzo(k)fluoranthene	23.06	252	8828m	0.142	ng/ul	
23) Benzo(a)pyrene	23.62	252	12517m	0.227	ng/ul	
24) Indeno(1,2,3-cd)pyrene	26.06	276	13777m	0.218	ng/ul	
25) Dibenzo(a,h)anthracene	26.07	278	3274m	0.064	ng/ul	
26) Benzo(g,h,i)perylene	26.79	276	11927m	0.217	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BN003801.D
Acq On : 11 Dec 2018 04:49
Operator : JU/SJ
Sample : J6077-13
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :

Quant Time: Dec 11 08:21:02 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 11 04:22:54 2018
Response via : Initial Calibration

