

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN121018\
 Data File : BN003775.D
 Acq On : 10 Dec 2018 11:50
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.494

Quant Time: Dec 11 03:42:24 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN111918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 08 00:32:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	5646	0.40	ng/ul	0.00
2) Naphthalene-d8	10.63	136	22710	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	11680	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.21	188	25209	0.40	ng/ul	0.00
16) Chrysene-d12	21.40	240	21912	0.40	ng/ul	0.00
20) Perylene-d12	23.71	264	19671	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.22	152	13816	0.40	ng/ul	0.00
14) Fluoranthene-d10	19.24	212	30644	0.42	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.68	128	26747	0.419	ng/ul	100
5) 2-Methylnaphthalene	12.29	142	18506	0.415	ng/ul	100
7) Acenaphthylene	14.19	152	22921	0.436	ng/ul	100
8) Acenaphthene	14.53	153	19550	0.424	ng/ul	100
9) Fluorene	15.52	166	21534	0.401	ng/ul	100
11) Pentachlorophenol	16.86	266	1853	0.408	ng/ul	99
12) Phenanthrene	17.25	178	35391	0.428	ng/ul	100
13) Anthracene	17.34	178	29319	0.431	ng/ul	100
15) Fluoranthene	19.27	202	42212	0.432	ng/ul	97
17) Pyrene	19.63	202	40653	0.517	ng/ul	99
18) Benzo(a)anthracene	21.38	228	32152	0.465	ng/ul	98
19) Chrysene	21.43	228	33302	0.420	ng/ul	100
21) Benzo(b)fluoranthene	23.01	252	33125	0.407	ng/ul	99
22) Benzo(k)fluoranthene	23.05	252	32117	0.405	ng/ul	98
23) Benzo(a)pyrene	23.61	252	29555	0.419	ng/ul	99
24) Indeno(1,2,3-cd)pyrene	26.05	276	31636	0.392	ng/ul	97
25) Dibenzo(a,h)anthracene	26.07	278	24169	0.372	ng/ul	98
26) Benzo(g,h,i)perylene	26.78	276	27568	0.393	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN121018\
Data File : BN003775.D
Acq On : 10 Dec 2018 11:50
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.494

Quant Time: Dec 11 03:42:24 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN11918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 08 00:32:32 2018
Response via : Initial Calibration

