

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN121018\
 Data File : BN003803.D
 Acq On : 11 Dec 2018 05:59
 Operator : JU/SJ
 Sample : J6077-17
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 F9Y37

Manual Integrations
 APPROVED

Sohil
 12/11/2018 6:12:40 PM

Quant Time: Dec 11 08:27:33 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN111918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 11 04:22:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	5800	0.40	ng/ul	0.00
2) Naphthalene-d8	10.63	136	23609	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	13964	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.22	188	24317m	0.40	ng/ul	0.00
16) Chrysene-d12	21.40	240	16115m	0.40	ng/ul	0.00
20) Perylene-d12	23.71	264	15860m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.22	152	9101m	0.25	ng/ul	0.00
14) Fluoranthene-d10	19.24	212	17597m	0.25	ng/ul	0.00
Target Compounds						
3) Naphthalene	10.68	128	6134	0.092	ng/ul#	Ovalue 1

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN121018\
Data File : BN003803.D
Acq On : 11 Dec 2018 05:59
Operator : JU/SJ
Sample : J6077-17
Misc :
ALS Vial : 29 Sample Multiplier: 1




Instrument :
BNA_N
ClientSampleId :
F9Y37

Quant Time: Dec 11 08:27:33 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 11 04:22:54 2018
Response via : Initial Calibration

Manual Integrations
APPROVED

Sohil
12/11/2018 6:12:40 PM





#3

Naphthalene

Concen: 0.092 ng/ul

RT: 10.68 min Scan# 2374

Delta R.T. 0.01 min

Lab File: BN003803.D

Acq: 11 Dec 2018 05:59

Instrument :

BNA_N

ClientSampleId :

F9Y37

Tot Ion:128 Resp: 6134

Ion Ratio Lower Upper

128 100

129 81.9 9.0 13.6#

127 49.1 10.5 15.7#

Manual Integrations
APPROVED

Sohil
12/11/2018 6:12:40 PM

