


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121018\
 Data File : BN003774.D
 Acq On : 10 Dec 2018 11:15
 Operator : JU/SJ
 Sample : DFTPP64
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP64

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN11918.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Dec 11 04:22:54 2018



AutoFind: Scans 2488, 2489, 2490; Background Corrected with Scan 2481

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	26.8	48256	PASS
68	69	0.00	2	1.9	1063	PASS
69	198	0.00	100	30.7	55216	PASS
70	69	0.00	2	0.4	246	PASS
127	198	10	80	42.8	77050	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	179997	PASS
199	198	5	9	6.7	12108	PASS
275	198	10	60	32.0	57597	PASS
365	198	1	100	4.1	7339	PASS
441	198	0.01	100	15.1	27098	PASS
442	198	50	100	93.5	168298	PASS
443	442	15	24	19.2	32330	PASS

m/z	Abundance
44.00	1978.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

m/z	Abundance
38.00	589.0
39.00	4427.0
40.00	331.0
44.00	1688.0
49.10	495.0
50.00	15414.0
51.00	61296.0
52.00	3347.0
55.00	456.0
56.00	1818.0
57.00	4617.0
61.00	793.0
62.00	989.0
63.00	2572.0
64.00	392.0
65.00	1416.0
68.10	1273.0
69.00	69104.0
70.00	387.0
73.00	715.0
74.00	6901.0
75.00	10833.0
76.00	3640.0
77.00	77912.0
78.00	5471.0
79.00	5402.0
80.00	3748.0
81.00	5534.0
82.00	1397.0
83.00	1434.0
85.00	1017.0
86.00	1683.0
87.00	715.0
91.00	1345.0
92.00	1447.0
93.00	8803.0
94.00	674.0
96.00	458.0
98.00	6992.0
99.00	5428.0
100.00	564.0
101.00	3255.0
103.00	1218.0
104.00	2111.0
105.00	1889.0
106.10	671.0
107.00	23752.0
108.00	3865.0
109.10	860.0
110.00	47768.0
111.00	7096.0
112.00	772.0
116.00	1378.0
117.00	18376.0
118.00	1382.0
120.00	341.0
122.00	1763.0
123.00	2486.0
124.00	1187.0
125.00	1135.0
127.00	92248.0
128.00	7252.0
129.00	34344.0
130.00	2986.0
130.90	660.0
132.00	352.0
134.00	950.0
135.00	2798.0
136.00	1160.0
137.00	1528.0
137.90	312.0
140.00	434.0
141.00	4698.0
142.00	1610.0
143.00	1024.0
144.00	314.0
145.90	851.0
147.00	2544.0
148.00	4884.0
149.00	1202.0
151.10	609.0
151.80	438.0
153.00	1523.0
154.00	1210.0
155.00	2584.0
156.00	3810.0
157.00	794.0
158.00	848.0
159.00	618.0
160.00	1385.0
161.00	2350.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

162.00	693.0
165.00	1751.0
166.00	1353.0
167.00	8915.0
168.00	4557.0
169.00	788.0
170.00	344.0
171.00	394.0
172.00	943.0
173.00	1072.0
174.00	1937.0
175.00	3659.0
176.00	1091.0
177.00	1695.0
178.00	550.0
179.00	6847.0
180.00	5172.0
181.00	2365.0
182.00	369.0
184.00	593.0
185.00	3477.0
186.00	27184.0
187.00	7855.0
188.00	837.0
189.00	1568.0
191.10	773.0
192.00	2247.0
193.00	2440.0
194.00	521.0
195.00	373.0
196.00	5896.0
198.00	204736.0
199.00	14451.0
200.00	894.0
201.40	876.0
203.00	1343.0
204.00	7037.0
205.10	11991.0
206.10	51368.0
207.10	7059.0
208.00	1672.0
209.00	555.0
210.30	888.0
211.10	1872.0
214.90	537.0
216.10	1066.0
217.00	12651.0
218.00	1662.0
221.10	9374.0
223.00	3013.0
224.00	27992.0
225.00	7191.0
226.10	860.0
227.00	11187.0
228.00	1716.0
229.00	2415.0
230.00	374.0
231.00	1052.0
234.00	684.0
235.00	871.0
235.90	512.0
236.90	1055.0
239.00	433.0
240.00	370.0
241.00	636.0
242.00	1485.0
243.00	1684.0
244.10	23120.0
245.00	3100.0
246.00	4285.0
247.00	946.0
249.00	929.0
253.00	521.0
255.00	122712.0
256.00	18288.0
257.00	1480.0
258.00	6735.0
259.00	1020.0
265.00	2606.0
266.00	429.0
272.00	307.0
273.00	3845.0
274.00	9984.0
275.00	60824.0
276.00	8180.0
277.00	4742.0
278.00	779.0
283.00	591.0
284.00	366.0
285.00	930.0
293.00	1045.0
294.00	305.0
296.00	15185.0
297.00	2173.0
303.00	1742.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

304.00	554.0
314.00	741.0
314.90	1712.0
316.00	1051.0
321.00	500.0
322.00	303.0
323.00	5516.0
324.00	1117.0
326.90	1034.0
328.00	571.0
332.00	404.0
333.00	517.0
334.00	3462.0
335.00	1000.0
341.00	590.0
345.90	1006.0
352.00	1602.0
353.00	1111.0
354.10	1864.0
365.00	7076.0
366.00	1050.0
371.00	455.0
372.00	2916.0
373.00	696.0
383.00	780.0
390.00	384.0
402.00	1198.0
403.00	1829.0
404.00	607.0
421.10	1405.0
422.10	1398.0
423.10	9677.0
424.00	2496.0
425.00	361.0
441.10	23776.0
442.10	145152.0
443.10	29208.0
444.10	2751.0

Instrument :

BNA_N

ClientSampleId :

DFTPP64

m/z	Abundance
38.00	652.0
39.00	3970.0
40.00	336.0
44.00	1717.0
49.10	380.0
50.00	13859.0
51.00	54728.0
52.00	2726.0
55.00	407.0
56.00	1849.0
57.00	4021.0
61.00	823.0
62.00	931.0
63.00	2312.0
64.00	376.0
65.00	1160.0
68.10	1321.0
69.00	63072.0
70.00	352.0
73.00	797.0
74.00	6392.0
75.00	10235.0
76.10	3443.0
77.00	71520.0
78.00	4898.0
79.00	4893.0
80.00	3650.0
81.00	5161.0
82.00	1294.0
83.00	1276.0
85.00	985.0
86.00	1455.0
87.00	702.0
91.00	1304.0
92.00	1308.0
93.00	8211.0
94.00	536.0
96.00	361.0
98.00	6631.0
99.00	5146.0
100.00	418.0
101.00	2999.0
103.00	1037.0
104.00	1962.0
105.00	1875.0
106.10	601.0
107.00	23328.0
108.00	3630.0
109.20	880.0
110.00	45624.0
111.00	6523.0
112.00	862.0
116.00	1429.0
117.00	17536.0
118.00	1307.0
122.00	1697.0
123.00	2587.0
124.00	1204.0
125.00	1176.0
127.00	89640.0
128.00	6636.0
129.00	33784.0
130.00	2987.0
131.00	582.0
134.00	1083.0
135.00	2748.0
136.00	1049.0
137.00	1569.0
137.90	366.0
140.00	423.0
141.00	4586.0
142.00	1450.0
142.90	1015.0
144.90	318.0
146.00	880.0
147.00	2387.0
148.00	5173.0
149.00	1195.0
151.00	647.0
151.80	392.0
153.00	1427.0
154.00	1181.0
155.00	2588.0
156.00	3894.0
157.10	915.0
157.90	882.0
159.00	650.0
160.00	1446.0
161.00	2143.0
162.00	613.0
165.00	1729.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

166.00	1487.0
167.00	9111.0
168.00	4380.0
169.00	805.0
170.00	380.0
170.90	482.0
172.00	825.0
173.00	1115.0
174.00	1948.0
175.00	3898.0
176.00	1024.0
177.00	1877.0
178.00	591.0
179.00	7337.0
180.00	5078.0
181.00	2399.0
182.00	404.0
184.00	644.0
185.00	3507.0
186.00	28176.0
187.00	8079.0
188.00	795.0
189.00	1752.0
191.00	890.0
192.00	2396.0
193.00	2525.0
194.00	567.0
196.00	6771.0
198.00	212352.0
199.00	14014.0
200.00	1018.0
201.50	1035.0
203.00	1371.0
204.00	7109.0
205.00	12217.0
206.10	54848.0
207.10	7364.0
208.00	1683.0
208.90	606.0
210.30	950.0
211.00	2140.0
215.00	551.0
216.00	1182.0
217.00	13661.0
218.00	1727.0
221.00	10067.0
223.00	3161.0
224.10	29768.0
225.00	7767.0
226.00	877.0
227.00	11669.0
228.00	1697.0
229.00	2573.0
230.00	320.0
231.00	1072.0
234.00	733.0
235.00	893.0
236.10	525.0
237.00	1023.0
239.00	494.0
240.00	421.0
241.00	706.0
242.00	1692.0
243.10	1771.0
244.00	25528.0
245.00	3342.0
246.00	4622.0
247.00	988.0
249.00	953.0
253.00	547.0
255.00	135808.0
256.00	19608.0
257.00	1531.0
258.00	7264.0
259.00	1129.0
265.00	2930.0
265.90	411.0
272.10	362.0
273.00	4366.0
274.00	11742.0
275.00	69400.0
276.00	8516.0
277.00	5165.0
278.00	937.0
283.00	580.0
284.00	360.0
285.00	955.0
292.00	302.0
293.00	1203.0
296.00	17272.0
297.00	2489.0
302.00	349.0
303.00	2161.0
304.10	647.0
314.00	930.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

315.00	2033.0
316.00	1283.0
321.00	643.0
323.00	6270.0
324.10	1051.0
327.00	1194.0
328.00	633.0
332.00	415.0
333.00	712.0
334.00	4332.0
335.00	1123.0
341.10	732.0
346.00	1421.0
352.00	2121.0
353.00	1466.0
354.10	2242.0
355.00	464.0
365.00	8594.0
366.00	1337.0
371.00	530.0
372.00	3744.0
373.00	985.0
383.00	1024.0
384.00	322.0
390.00	500.0
391.00	314.0
402.00	1475.0
403.00	2195.0
404.10	743.0
421.00	1889.0
422.10	1828.0
423.10	12833.0
424.10	3042.0
425.00	318.0
441.10	33792.0
442.10	206400.0
443.10	39584.0
444.10	3599.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

m/z	Abundance
38.00	336.0
39.00	2224.0
44.00	1967.0
50.00	7177.0
51.00	28744.0
52.00	1556.0
56.00	941.0
57.00	2114.0
60.90	365.0
61.90	402.0
63.00	1259.0
65.00	653.0
68.20	597.0
69.00	33472.0
73.00	763.0
74.00	3391.0
75.00	5467.0
76.00	1820.0
77.00	38656.0
78.00	2600.0
79.00	2727.0
80.00	1930.0
81.00	2929.0
82.00	725.0
83.00	748.0
84.90	520.0
86.00	767.0
87.00	364.0
91.00	717.0
92.00	775.0
93.00	4660.0
98.00	3545.0
99.00	2789.0
101.00	1669.0
103.00	629.0
104.00	992.0
105.00	995.0
106.00	340.0
107.00	12242.0
108.00	1861.0
109.10	468.0
110.00	24504.0
111.00	3477.0
112.00	415.0
116.00	768.0
117.00	9852.0
118.00	682.0
122.00	844.0
123.00	1284.0
124.00	568.0
125.00	592.0
127.00	49264.0
128.00	3847.0
129.00	18536.0
130.00	1491.0
130.90	341.0
134.00	505.0
135.00	1604.0
136.00	663.0
137.00	885.0
141.00	2558.0
142.00	820.0
142.90	640.0
146.00	464.0
147.00	1512.0
148.00	2855.0
149.00	623.0
150.90	325.0
153.00	810.0
153.90	622.0
155.00	1337.0
156.00	2269.0
157.00	507.0
157.90	544.0
159.00	452.0
160.00	714.0
161.00	1265.0
162.00	399.0
165.00	1035.0
166.00	748.0
167.00	5135.0
168.00	2672.0
169.00	479.0
171.90	435.0
173.00	688.0
174.00	1112.0
175.00	2194.0
176.00	567.0
176.90	1074.0
178.00	320.0
179.00	4066.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64

180.00	2831.0
181.00	1346.0
184.00	313.0
185.00	2057.0
186.00	16295.0
187.00	4563.0
188.00	499.0
189.00	1098.0
191.00	565.0
192.00	1369.0
193.00	1540.0
196.00	3762.0
198.00	122904.0
199.00	7860.0
200.00	589.0
201.50	609.0
203.00	789.0
204.00	4225.0
205.00	7376.0
206.10	31696.0
207.10	4167.0
208.00	1114.0
209.00	423.0
210.20	486.0
211.00	1308.0
215.00	354.0
216.00	720.0
217.00	8185.0
218.00	1046.0
221.10	6025.0
223.00	1913.0
224.10	18384.0
225.00	4705.0
226.10	514.0
227.00	7505.0
228.00	997.0
229.00	1597.0
231.00	699.0
234.00	502.0
235.00	557.0
236.00	357.0
237.00	608.0
241.00	465.0
242.00	1018.0
243.00	1204.0
244.10	15259.0
245.10	1900.0
246.00	2953.0
247.00	562.0
249.00	648.0
253.00	321.0
255.00	81872.0
256.00	11610.0
257.00	872.0
258.00	4654.0
259.00	662.0
265.00	1869.0
273.00	2619.0
274.00	6897.0
275.00	42568.0
276.00	5529.0
277.00	3200.0
278.00	556.0
282.90	402.0
285.00	567.0
293.00	809.0
296.00	11010.0
297.00	1649.0
303.00	1381.0
304.00	371.0
314.00	622.0
315.00	1320.0
316.00	896.0
321.00	401.0
323.10	4285.0
324.10	709.0
327.00	858.0
328.00	389.0
332.00	316.0
333.00	425.0
334.00	2908.0
335.00	644.0
341.00	590.0
346.00	969.0
352.00	1318.0
353.00	1021.0
354.00	1471.0
355.00	357.0
365.00	6347.0
366.00	829.0
371.00	406.0
372.10	2477.0
373.10	682.0
383.00	693.0
390.00	332.0

Instrument :

BNA_N

ClientSampleId :

DFTPP64

402.00	1079.0
403.10	1548.0
404.00	624.0
421.00	1311.0
422.00	1285.0
423.10	9342.0
424.00	1874.0
441.10	23728.0
442.10	153344.0
443.10	28200.0
444.10	2501.0

Instrument :
BNA_N
ClientSampleId :
DFTPP64