



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121021\
 Data File : BN017869.D
 Acq On : 10 Dec 2021 13:59
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN120721.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Dec 08 01:18:09 2021






AutoFind: Scans 2475, 2476, 2477; Background Corrected with Scan 2469

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.0	286024	PASS
68	69	0.00	2	1.8	5310	PASS
69	198	0.00	100	36.5	290411	PASS
70	69	0.00	2	0.6	1702	PASS
127	198	10	80	39.1	310955	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	795200	PASS
199	198	5	9	6.7	53256	PASS
275	198	10	60	27.4	218027	PASS
365	198	1	100	3.5	27671	PASS
441	198	0.01	100	11.0	87621	PASS
442	442	50	100	100.0	574336	PASS
443	442	15	24	18.9	108651	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121021\
Data File : BN017869.D
Acq On : 10 Dec 2021 13:59
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Dec 11 07:40:04 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 11 07:39:58 2021
Response via : Initial Calibration



TIC: BN017869.D\data.ms

(77) Benzidine

19.345min (0.000) 0.00 ng




response 3439254

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.80
183.00	13.20	11.39
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121021\
Data File : BN017869.D
Acq On : 10 Dec 2021 13:59
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Dec 11 07:40:04 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 11 07:39:58 2021
Response via : Initial Calibration



TIC: BN017869.D\data.ms

(70) Pentachlorophenol (C)

16.745min (0.000) 14903.35 ng

response 550988

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	62.93
264.00	61.60	61.24
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
12/10/2021	BNA_N	<u>BN017869.D</u>
Compound Name	Response	Retention Time
DDT	2235560	20.586
DDD	87000	20.145
DDE	4495	19.639
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
91495	2327055	3.93

Instrument :
BNA_N
ClientSampleId :
DFTPP