


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121022\
 Data File : BN023115.D
 Acq On : 10 Dec 2022 09:44
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120822.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Dec 12 05:18:26 2022






AutoFind: Scans 2508, 2509, 2510; Background Corrected with Scan 2501

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.1	212304	PASS
68	69	0.00	2	1.6	3614	PASS
69	198	0.00	100	42.2	229066	PASS
70	69	0.00	2	0.5	1230	PASS
127	198	10	80	50.6	274901	PASS
197	198	0.00	2	0.3	1414	PASS
198	198	100	100	100.0	543083	PASS
199	198	5	9	6.9	37448	PASS
275	198	10	60	25.5	138595	PASS
365	198	1	100	3.0	16503	PASS
441	198	0.01	100	12.9	70048	PASS
442	442	50	100	100.0	460864	PASS
443	442	15	24	19.0	87344	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121022\
 Data File : BN023115.D
 Acq On : 10 Dec 2022 09:44
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Quant Time: Dec 12 07:11:29 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 12 07:11:25 2022
 Response via : Initial Calibration



TIC: BN023115.D\data.ms

(70) Pentachlorophenol (C)

17.057min (0.000) 18995.35 ng




response 617302

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.57
264.00	61.60	60.89
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121022\
Data File : BN023115.D
Acq On : 10 Dec 2022 09:44
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Dec 12 07:11:29 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 12 07:11:25 2022
Response via : Initial Calibration



TIC: BN023115.D\data.ms

(77) Benzidine

19.651min (0.000) 0.00 ng

response 4209687

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.98
183.00	13.20	11.89
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
12/10/2022	BNA_N	<u>BN023115.D</u>
Compound Name	Response	Retention Time
DDT	1752487	20.886
DDD	14976	20.492
DDE	276	19.939
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
15252	1767739	0.86

Instrument :
BNA_N
ClientSampleId :
DFTPP