

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121120\
 Data File : BN013083.D
 Acq On : 12 Dec 2020 09:32
 Operator : CG/JU
 Sample : L4996-06
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE117D1-20201208

Manual Integrations
 APPROVED

mohammad
 12/15/2020 12:19:14 PM

Quant Time: Dec 14 03:06:14 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 11 01:53:51 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.370	152	904	0.40	ng	0.00
7) Naphthalene-d8	10.137	136	2909	0.40	ng	# 0.00
13) Acenaphthene-d10	14.029	164	1711	0.40	ng	0.00
19) Phenanthrene-d10	16.799	188	3746	0.40	ng	# 0.00
29) Chrysene-d12	21.024	240	4139	0.40	ng	# 0.01
36) Perylene-d12	23.113	264	4594	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	4.994	112	313	0.10	ng	0.00
5) Phenol-d6	6.565	99	228	0.06	ng	0.00
8) Nitrobenzene-d5	8.541	82	801	0.26	ng	0.01
11) 2-Methylnaphthalene-d10	11.746	152	1351	0.26	ng	0.00
14) 2,4,6-Tribromophenol	15.552	330	257	0.23	ng	0.00
15) 2-Fluorobiphenyl	12.646	172	2115	0.28	ng	0.00
27) Fluoranthene-d10	18.838	212	4122	0.33	ng	0.00
31) Terphenyl-d14	19.454	244	3787	0.37	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.965	88	357	0.11	ng	# 71
34) Bis(2-ethylhexyl)phtha...	20.950	149	1614m	0.17	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121120\
Data File : BN013083.D
Acq On : 12 Dec 2020 09:32
Operator : CG/JU
Sample : L4996-06
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE117D1-20201208

Manual Integrations
APPROVED

mohammad
12/15/2020 12:19:14 PM

Quant Time: Dec 14 03:06:14 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 11 01:53:51 2020
Response via : Initial Calibration

