

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121120\
 Data File : BN013063.D
 Acq On : 11 Dec 2020 11:45
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 11 12:49:42 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 11 01:53:51 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	82	0.00
2	1,4-Dioxane	0.949	1.034	-9.0	81	0.00
3	n-Nitrosodimethylamine	1.050	1.345	-28.1#	123	0.00
4 S	2-Fluorophenol	1.437	1.504	-4.7	95	0.00
5 S	Phenol-d6	1.646	1.674	-1.7	95	0.00
6	bis(2-Chloroethyl)ether	1.111	1.182	-6.4	99	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	86	0.00
8 S	Nitrobenzene-d5	0.424	0.400	5.7	91	0.01
9	Naphthalene	1.224	1.204	1.6	94	0.00
10	Hexachlorobutadiene	0.265	0.283	-6.8	95	-0.01
11 SURR	2-Methylnaphthalene-d10	0.722	0.697	3.5	92	0.00
12	2-Methylnaphthalene	0.839	0.792	5.6	90	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	85	0.00
14 S	2,4,6-Tribromophenol	0.263	0.231	12.2	85	0.00
15 S	2-Fluorobiphenyl	1.762	1.745	1.0	92	0.00
16	Acenaphthylene	2.132	2.007	5.9	90	0.00
17	Acenaphthene	1.345	1.337	0.6	92	0.00
18	Fluorene	1.795	1.766	1.6	93	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	82	0.00
20	4,6-Dinitro-2-methylphenol	0.076	0.077	-1.3	91	0.00
21	4-Bromophenyl-phenylether	0.093	0.195	-109.7#	112	0.00
22	Hexachlorobenzene	0.294	0.295	-0.3	90	0.00
23	Atrazine	0.238	0.226	5.0	91	0.00
24	Pentachlorophenol	0.123	0.108	12.2	88	0.01
25	Phenanthrene	1.314	1.258	4.3	89	0.00
26	Anthracene	1.167	1.084	7.1	89	0.00
27 SURR	Fluoranthene-d10	1.343	1.276	5.0	88	0.00
28	Fluoranthene	1.621	1.525	5.9	88	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	77	0.01
30	Pyrene	1.500	1.481	1.3	87	0.00
31 S	Terphenyl-d14	0.998	0.989	0.9	87	0.00
32	Benzo(a)anthracene	1.466	1.365	6.9	80	0.00
33	Chrysene	1.524	1.520	0.3	88	0.00
34	Bis(2-ethylhexyl)phthalate	0.893	0.674	24.5	66	0.00
35	Indeno(1,2,3-cd)pyrene	2.024	1.979	2.2	82	0.00
36 I	Perylene-d12	1.000	1.000	0.0	75	0.00
37	Benzo(b)fluoranthene	1.520	1.465	3.6	85	0.00
38	Benzo(k)fluoranthene	1.605	1.589	1.0	84	0.00
39 C	Benzo(a)pyrene	1.450	1.418	2.2	81	0.00
40	Dibenzo(a,h)anthracene	1.582	1.558	1.5	83	0.00
41	Benzo(g,h,i)perylene	1.673	1.706	-2.0	83	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4