

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121121\
 Data File : BN017884.D
 Acq On : 10 Dec 2021 23:41
 Operator : CG/JU
 Sample : PB141310BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB141310BL

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 12/13/2021
 Supervised By :Jagrut Upadhyay 12/13/2021

Quant Time: Dec 11 05:45:40 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 11 00:01:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.715	152	22692	0.400	ng	0.00
7) Naphthalene-d8	10.506	136	79103	0.400	ng	0.01
13) Acenaphthene-d10	14.347	164	49933	0.400	ng	0.01
19) Phenanthrene-d10	17.117	188	90552	0.400	ng	# 0.02
29) Chrysene-d12	21.303	240	62591	0.400	ng	# 0.01
35) Perylene-d12	23.596	264	45119	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.345	112	13101	0.280	ng	0.02
5) Phenol-d6	6.912	99	8843	0.190	ng	0.02
8) Nitrobenzene-d5	8.859	82	14924	0.289	ng	0.00
11) 2-Methylnaphthalene-d10	12.107	152	38927	0.280	ng	0.02
14) 2,4,6-Tribromophenol	15.895	330	1063m	0.372	ng	0.05
15) 2-Fluorobiphenyl	12.989	172	46322m	0.264	ng	0.03
27) Fluoranthene-d10	19.139	212	94611	0.322	ng	0.00
31) Terphenyl-d14	19.750	244	69147	0.448	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121121\
Data File : BN017884.D
Acq On : 10 Dec 2021 23:41
Operator : CG/JU
Sample : PB141310BL
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB141310BL

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 12/13/2021
Supervised By : Jagrut Upadhyay 12/13/2021

Quant Time: Dec 11 05:45:40 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 11 00:01:30 2021
Response via : Initial Calibration

