

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121121\
 Data File : BN017905.D
 Acq On : 11 Dec 2021 14:08
 Operator : CG/JU
 Sample : M4923-15DL 2X
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 TT101D1-20211201DL

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 12/13/2021
 Supervised By : Jagrut Upadhyay 12/13/2021

Quant Time: Dec 11 21:30:27 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 11 00:01:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.715	152	26514	0.400	ng	0.00
7) Naphthalene-d8	10.494	136	98675	0.400	ng	# 0.00
13) Acenaphthene-d10	14.347	164	67764	0.400	ng	0.01
19) Phenanthrene-d10	17.092	188	138868	0.400	ng	0.00
29) Chrysene-d12	21.293	240	94443	0.400	ng	# 0.00
35) Perylene-d12	23.596	264	67209	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.326	112	2952	0.054	ng	0.00
5) Phenol-d6	6.903	99	1465	0.027	ng	0.00
8) Nitrobenzene-d5	8.859	82	7539	0.117	ng	0.00
11) 2-Methylnaphthalene-d10	12.098	152	19102	0.110	ng	0.00
14) 2,4,6-Tribromophenol	15.857	330	1310	0.368	ng	0.01
15) 2-Fluorobiphenyl	12.977	172	27390	0.115	ng	0.01
27) Fluoranthene-d10	19.129	212	74171	0.165	ng	0.00
31) Terphenyl-d14	19.740	244	54004	0.232	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.270	88	22410m	2.785	ng	Qvalue
6) bis(2-Chloroethyl)ether	7.143	93	1677	0.025	ng	98
34) Bis(2-ethylhexyl)phtha...	21.219	149	5812	0.057	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121121\
Data File : BN017905.D
Acq On : 11 Dec 2021 14:08
Operator : CG/JU
Sample : M4923-15DL 2X
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
TT101D1-20211201DL

Quant Time: Dec 11 21:30:27 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 11 00:01:30 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 12/13/2021
Supervised By :Jagrut Upadhyay 12/13/2021

