




Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN121121\  
Data File : BN017875.D  
Acq On : 10 Dec 2021 17:39  
Operator : CG/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 11 07:36:00 2021  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Sat Dec 11 07:33:12 2021  
Response via : Initial Calibration



TIC: BN017875.D\data.ms

**(70) Pentachlorophenol (C)**




16.751min (-0.000) 10438.07 ng m

response 388934

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 62.20  | 65.39  |
| 264.00 | 61.60  | 64.27  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN121121\  
Data File : BN017875.D  
Acq On : 10 Dec 2021 17:39  
Operator : CG/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 11 07:36:00 2021  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Sat Dec 11 07:33:12 2021  
Response via : Initial Calibration



TIC: BN017875.D\data.ms

**(77) Benzidine**

19.345min (-0.000) 0.00 ng

response 3684339

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 15.50  | 13.93  |
| 183.00 | 13.20  | 11.64  |
| 0.00   | 0.00   | 0.00   |