




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121121\
Data File : BN017893.D
Acq On : 11 Dec 2021 06:18
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 11 07:33:16 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 11 07:33:12 2021
Response via : Initial Calibration



TIC: BN017893.D\data.ms

(70) Pentachlorophenol (C)




16.751min (0.000) 14086.35 ng

response 441079

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	62.24
264.00	61.60	60.37
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121121\
Data File : BN017893.D
Acq On : 11 Dec 2021 06:18
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 11 07:33:16 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 11 07:33:12 2021
Response via : Initial Calibration



TIC: BN017893.D\data.ms

(77) Benzidine

19.345min (0.000) 0.00 ng m

response 3895064

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.74
183.00	13.20	11.06
0.00	0.00	0.00