

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121121\
 Data File : BN017883.D
 Acq On : 10 Dec 2021 23:05
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN121121

Quant Time: Dec 11 00:10:47 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 11 00:01:30 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	107	0.00
2	1,4-Dioxane	0.400	0.400	0.0	113	0.00
3	n-Nitrosodimethylamine	0.400	0.389	2.8	106	0.00
4 S	2-Fluorophenol	0.400	0.389	2.8	108	0.00
5 S	Phenol-d6	0.400	0.381	4.8	108	0.00
6	bis(2-Chloroethyl)ether	0.400	0.404	-1.0	107	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	106	0.00
8 S	Nitrobenzene-d5	0.400	0.341	14.8	108	0.00
9	Naphthalene	0.400	0.390	2.5	104	0.00
10	Hexachlorobutadiene	0.400	0.394	1.5	106	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.398	0.5	105	0.00
12	2-Methylnaphthalene	0.400	0.403	-0.8	105	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	104	0.00
14 S	2,4,6-Tribromophenol	0.400	0.478	-19.5	117	0.00
15 S	2-Fluorobiphenyl	0.400	0.395	1.3	105	0.00
16	Acenaphthylene	0.400	0.377	5.8	106	0.00
17	Acenaphthene	0.400	0.398	0.5	105	0.00
18	Fluorene	0.400	0.398	0.5	105	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	104	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.000	100.0#	0	-15.48#
21	4-Bromophenyl-phenylether	0.400	0.377	5.8	107	0.00
22	Hexachlorobenzene	0.400	0.403	-0.8	105	0.00
23	Atrazine	0.400	0.379	5.3	112	0.00
24	Pentachlorophenol	0.400	0.384	4.0	132	0.00
25	Phenanthrene	0.400	0.395	1.3	105	0.00
26	Anthracene	0.400	0.377	5.8	105	0.00
27 SURR	Fluoranthene-d10	0.400	0.390	2.5	106	0.00
28	Fluoranthene	0.400	0.406	-1.5	106	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	105	0.00
30	Pyrene	0.400	0.417	-4.2	106	0.00
31 S	Terphenyl-d14	0.400	0.408	-2.0	106	0.00
32	Benzo(a)anthracene	0.400	0.363	9.3	107	-0.01
33	Chrysene	0.400	0.387	3.3	103	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.349	12.8	109	0.00
35 I	Perylene-d12	0.400	0.400	0.0	100	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.349	12.8	100	0.00
37	Benzo(b)fluoranthene	0.400	0.372	7.0	97	0.00
38	Benzo(k)fluoranthene	0.400	0.375	6.3	103	0.00
39 C	Benzo(a)pyrene	0.400	0.371	7.3	99	0.00
40	Dibenzo(a,h)anthracene	0.400	0.355	11.3	105	0.00
41	Benzo(g,h,i)perylene	0.400	0.355	11.3	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN121121