

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121121\
 Data File : BN017883.D
 Acq On : 10 Dec 2021 23:05
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN121121

Quant Time: Dec 11 00:10:47 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 11 00:01:30 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.121	0.121	0.0	113	0.00
3	n-Nitrosodimethylamine	0.367	0.357	2.7	106	0.00
4 S	2-Fluorophenol	0.825	0.803	2.7	108	0.00
5 S	Phenol-d6	0.819	0.779	4.9	108	0.00
6	bis(2-Chloroethyl)ether	1.004	1.013	-0.9	107	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	106	0.00
8 S	Nitrobenzene-d5	0.261	0.223	14.6	108	0.00
9	Naphthalene	0.995	0.970	2.5	104	0.00
10	Hexachlorobutadiene	0.283	0.279	1.4	106	0.00
11 SURR	2-Methylnaphthalene-d10	0.703	0.700	0.4	105	0.00
12	2-Methylnaphthalene	0.662	0.667	-0.8	105	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
14 S	2,4,6-Tribromophenol	0.129	0.079	38.8#	117	0.00
15 S	2-Fluorobiphenyl	1.408	1.390	1.3	105	0.00
16	Acenaphthylene	1.493	1.409	5.6	106	0.00
17	Acenaphthene	1.038	1.033	0.5	105	0.00
18	Fluorene	1.263	1.257	0.5	105	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
20	4,6-Dinitro-2-methylphenol	0.002	0.000	100.0#	0#	-15.48#
21	4-Bromophenyl-phenylether	0.229	0.216	5.7	107	0.00
22	Hexachlorobenzene	0.299	0.301	-0.7	105	0.00
23	Atrazine	0.142	0.141	0.7	112	0.00
24	Pentachlorophenol	0.014	0.006	57.1#	132	0.00
25	Phenanthrene	1.022	1.010	1.2	105	0.00
26	Anthracene	0.852	0.803	5.8	105	0.00
27 SURR	Fluoranthene-d10	1.297	1.264	2.5	106	0.00
28	Fluoranthene	1.258	1.276	-1.4	106	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	105	0.00
30	Pyrene	1.569	1.637	-4.3	106	0.00
31 S	Terphenyl-d14	0.986	1.005	-1.9	106	0.00
32	Benzo(a)anthracene	1.086	1.004	7.6	107	-0.01
33	Chrysene	1.314	1.270	3.3	103	0.00
34	Bis(2-ethylhexyl)phthalate	0.429	0.374	12.8	109	0.00
35 I	Perylene-d12	1.000	1.000	0.0	100	0.00
36	Indeno(1,2,3-cd)pyrene	0.605	0.549	9.3	100	0.00
37	Benzo(b)fluoranthene	1.321	1.228	7.0	97	0.00
38	Benzo(k)fluoranthene	1.288	1.206	6.4	103	0.00
39 C	Benzo(a)pyrene	1.175	1.088	7.4	99	0.00
40	Dibenzo(a,h)anthracene	0.397	0.364	8.3	105	0.00
41	Benzo(g,h,i)perylene	0.636	0.564	11.3	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN121121