

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121121\
 Data File : BN017894.D
 Acq On : 11 Dec 2021 06:54
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 11 07:25:41 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 11 00:01:30 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	127	0.00
2	1,4-Dioxane	0.121	0.126	-4.1	139	0.00
3	n-Nitrosodimethylamine	0.367	0.378	-3.0	134	0.00
4 S	2-Fluorophenol	0.825	0.813	1.5	130	0.00
5 S	Phenol-d6	0.819	0.810	1.1	134	0.00
6	bis(2-Chloroethyl)ether	1.004	1.022	-1.8	128	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	127	0.00
8 S	Nitrobenzene-d5	0.261	0.226	13.4	131	-0.01
9	Naphthalene	0.995	0.974	2.1	125	0.00
10	Hexachlorobutadiene	0.283	0.272	3.9	123	0.00
11 SURR	2-Methylnaphthalene-d10	0.703	0.695	1.1	125	0.00
12	2-Methylnaphthalene	0.662	0.664	-0.3	125	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	123	0.00
14 S	2,4,6-Tribromophenol	0.129	0.079	38.8#	139	0.00
15 S	2-Fluorobiphenyl	1.408	1.405	0.2	126	0.00
16	Acenaphthylene	1.493	1.421	4.8	127	0.00
17	Acenaphthene	1.038	1.030	0.8	124	0.00
18	Fluorene	1.263	1.263	0.0	125	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	122	0.00
20	4,6-Dinitro-2-methylphenol	0.002	0.001	50.0#	2300#	0.05
21	4-Bromophenyl-phenylether	0.229	0.215	6.1	125	0.00
22	Hexachlorobenzene	0.299	0.295	1.3	121	0.00
23	Atrazine	0.142	0.123	13.4	114	0.00
24	Pentachlorophenol	0.014	0.011	21.4	261#	-0.01
25	Phenanthrene	1.022	1.012	1.0	124	0.00
26	Anthracene	0.852	0.812	4.7	125	0.00
27 SURR	Fluoranthene-d10	1.297	1.242	4.2	122	0.00
28	Fluoranthene	1.258	1.268	-0.8	123	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	125	0.00
30	Pyrene	1.569	1.567	0.1	122	0.00
31 S	Terphenyl-d14	0.986	0.943	4.4	118	0.00
32	Benzo(a)anthracene	1.086	0.977	10.0	125	-0.01
33	Chrysene	1.314	1.311	0.2	127	0.00
34	Bis(2-ethylhexyl)phthalate	0.429	0.311	27.5#	109	0.01
35 I	Perylene-d12	1.000	1.000	0.0	136	0.00
36	Indeno(1,2,3-cd)pyrene	0.605	0.643	-6.3	159#	0.00
37	Benzo(b)fluoranthene	1.321	1.194	9.6	129	0.00
38	Benzo(k)fluoranthene	1.288	1.187	7.8	138	0.00
39 C	Benzo(a)pyrene	1.175	1.093	7.0	135	0.00
40	Dibenzo(a,h)anthracene	0.397	0.471	-18.6	185#	0.00
41	Benzo(g,h,i)perylene	0.636	0.618	2.8	150	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4