

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN121121\  
 Data File : BN017894.D  
 Acq On : 11 Dec 2021 06:54  
 Operator : CG/JU  
 Sample : SSTDCCC0.4  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 LabSampleId :  
 SSTDCCC0.4

Quant Time: Dec 11 07:25:41 2021  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN121121.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Sat Dec 11 00:01:30 2021  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|         | Compound                   | Amount | Calc. | %Dev   | Area% | Dev(min) |
|---------|----------------------------|--------|-------|--------|-------|----------|
| 1 I     | 1,4-Dichlorobenzene-d4     | 0.400  | 0.400 | 0.0    | 127   | 0.00     |
| 2       | 1,4-Dioxane                | 0.400  | 0.416 | -4.0   | 139   | 0.00     |
| 3       | n-Nitrosodimethylamine     | 0.400  | 0.412 | -3.0   | 134   | 0.00     |
| 4 S     | 2-Fluorophenol             | 0.400  | 0.394 | 1.5    | 130   | 0.00     |
| 5 S     | Phenol-d6                  | 0.400  | 0.396 | 1.0    | 134   | 0.00     |
| 6       | bis(2-Chloroethyl)ether    | 0.400  | 0.407 | -1.7   | 128   | 0.00     |
| 7 I     | Naphthalene-d8             | 0.400  | 0.400 | 0.0    | 127   | 0.00     |
| 8 S     | Nitrobenzene-d5            | 0.400  | 0.346 | 13.5   | 131   | -0.01    |
| 9       | Naphthalene                | 0.400  | 0.391 | 2.3    | 125   | 0.00     |
| 10      | Hexachlorobutadiene        | 0.400  | 0.384 | 4.0    | 123   | 0.00     |
| 11 SURR | 2-Methylnaphthalene-d10    | 0.400  | 0.395 | 1.3    | 125   | 0.00     |
| 12      | 2-Methylnaphthalene        | 0.400  | 0.402 | -0.5   | 125   | 0.00     |
| 13 I    | Acenaphthene-d10           | 0.400  | 0.400 | 0.0    | 123   | 0.00     |
| 14 S    | 2,4,6-Tribromophenol       | 0.400  | 0.478 | -19.5  | 139   | 0.00     |
| 15 S    | 2-Fluorobiphenyl           | 0.400  | 0.399 | 0.3    | 126   | 0.00     |
| 16      | Acenaphthylene             | 0.400  | 0.381 | 4.8    | 127   | 0.00     |
| 17      | Acenaphthene               | 0.400  | 0.397 | 0.8    | 124   | 0.00     |
| 18      | Fluorene                   | 0.400  | 0.400 | 0.0    | 125   | 0.00     |
| 19 I    | Phenanthrene-d10           | 0.400  | 0.400 | 0.0    | 122   | 0.00     |
| 20      | 4,6-Dinitro-2-methylphenol | 0.400  | 0.622 | -55.5# | 2300  | 0.05     |
| 21      | 4-Bromophenyl-phenylether  | 0.400  | 0.375 | 6.3    | 125   | 0.00     |
| 22      | Hexachlorobenzene          | 0.400  | 0.395 | 1.3    | 121   | 0.00     |
| 23      | Atrazine                   | 0.400  | 0.341 | 14.8   | 114   | 0.00     |
| 24      | Pentachlorophenol          | 0.400  | 0.522 | -30.5# | 261   | -0.01    |
| 25      | Phenanthrene               | 0.400  | 0.396 | 1.0    | 124   | 0.00     |
| 26      | Anthracene                 | 0.400  | 0.381 | 4.8    | 125   | 0.00     |
| 27 SURR | Fluoranthene-d10           | 0.400  | 0.383 | 4.3    | 122   | 0.00     |
| 28      | Fluoranthene               | 0.400  | 0.403 | -0.8   | 123   | 0.00     |
| 29 I    | Chrysene-d12               | 0.400  | 0.400 | 0.0    | 125   | 0.00     |
| 30      | Pyrene                     | 0.400  | 0.400 | 0.0    | 122   | 0.00     |
| 31 S    | Terphenyl-d14              | 0.400  | 0.382 | 4.5    | 118   | 0.00     |
| 32      | Benzo(a)anthracene         | 0.400  | 0.355 | 11.3   | 125   | -0.01    |
| 33      | Chrysene                   | 0.400  | 0.399 | 0.3    | 127   | 0.00     |
| 34      | Bis(2-ethylhexyl)phthalate | 0.400  | 0.290 | 27.5#  | 109   | 0.01     |
| 35 I    | Perylene-d12               | 0.400  | 0.400 | 0.0    | 136   | 0.00     |
| 36      | Indeno(1,2,3-cd)pyrene     | 0.400  | 0.393 | 1.8    | 159   | 0.00     |
| 37      | Benzo(b)fluoranthene       | 0.400  | 0.361 | 9.8    | 129   | 0.00     |
| 38      | Benzo(k)fluoranthene       | 0.400  | 0.369 | 7.8    | 138   | 0.00     |
| 39 C    | Benzo(a)pyrene             | 0.400  | 0.372 | 7.0    | 135   | 0.00     |
| 40      | Dibenzo(a,h)anthracene     | 0.400  | 0.429 | -7.2   | 185   | 0.00     |
| 41      | Benzo(g,h,i)perylene       | 0.400  | 0.389 | 2.8    | 150   | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

**Instrument :**  
BNA\_N  
**LabSampleId :**  
SSTDCCC0.4