



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121121\
 Data File : BN017911.D
 Acq On : 11 Dec 2021 18:23
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121121.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Sat Dec 11 00:01:30 2021






AutoFind: Scans 2475, 2476, 2477; Background Corrected with Scan 2468

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.9	283332	PASS
68	69	0.00	2	2.0	5478	PASS
69	198	0.00	100	34.3	278229	PASS
70	69	0.00	2	0.6	1648	PASS
127	198	10	80	37.6	305024	PASS
197	198	0.00	2	0.1	1035	PASS
198	198	100	100	100.0	811968	PASS
199	198	5	9	6.8	54933	PASS
275	198	10	60	26.5	215104	PASS
365	198	1	100	3.2	25700	PASS
441	198	0.01	100	11.5	93645	PASS
442	442	50	100	100.0	586773	PASS
443	442	15	24	19.9	116581	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121121\
Data File : BN017911.D
Acq On : 11 Dec 2021 18:23
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Dec 11 22:39:38 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 11 07:39:58 2021
Response via : Initial Calibration



TIC: BN017911.D\data.ms

(77) Benzidine

19.345min (-0.121) 0.00 ng




response 3177751

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.23
183.00	13.20	11.47
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121121\
Data File : BN017911.D
Acq On : 11 Dec 2021 18:23
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Dec 11 22:39:38 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 11 07:39:58 2021
Response via : Initial Calibration



TIC: BN017911.D\data.ms

(70) Pentachlorophenol (C)

16.751min (-0.121) 10536.04 ng

response 436490

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	62.92
264.00	61.60	56.84
0.00	0.00	0.00

Instrument :
BNA_N
ClientSampleId :
DFTPP

DDT Breakdown

Date	Instrument Name	DFTPP Data File
12/11/2021	BNA_N	<u>BN017911.D</u>
Compound Name	Response	Retention Time
DDT	1895671	20.586
DDD	101855	20.145
DDE	4940	19.639
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
106795	2002466	5.33

Instrument :
BNA_N
ClientSampleId :
DFTPP