

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121321\
 Data File : BN017964.D
 Acq On : 13 Dec 2021 10:05
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Manual Integrations
 APPROVED

Reviewed By :Christian
 Giraldo

Quant Time: Dec 14 03:40:02 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 11 00:01:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
Internal Standards							12/14/2021
1) 1,4-Dichlorobenzene-d4	7.705	152	26059	0.400	ng	0.00	Supervised By :Jagrut padhyay
7) Naphthalene-d8	10.481	136	103690	0.400	ng	#-0.01	
13) Acenaphthene-d10	14.334	164	61345	0.400	ng	0.00	
19) Phenanthrene-d10	17.092	188	117337	0.400	ng	0.00	
29) Chrysene-d12	21.293	240	75026	0.400	ng	# 0.00	
35) Perylene-d12	23.596	264	40750	0.400	ng	0.00	12/14/2021
System Monitoring Compounds							
4) 2-Fluorophenol	5.317	112	20972	0.390	ng	0.00	
5) Phenol-d6	6.885	99	20047	0.376	ng	0.00	
8) Nitrobenzene-d5	8.846	82	21740	0.321	ng	-0.01	
11) 2-Methylnaphthalene-d10	12.085	152	68401	0.375	ng	0.00	
14) 2,4,6-Tribromophenol	15.844	330	3769	0.445	ng	0.00	
15) 2-Fluorobiphenyl	12.964	172	83518	0.387	ng	0.00	
27) Fluoranthene-d10	19.129	212	135841	0.357	ng	0.00	
31) Terphenyl-d14	19.740	244	76714	0.415	ng	0.00	
Target Compounds							Qvalue
2) 1,4-Dioxane	3.261	88	3002m	0.380	ng		
3) n-Nitrosodimethylamine	3.602	42	9280	0.388	ng		95
6) bis(2-Chloroethyl)ether	7.134	93	26073	0.399	ng		100
9) Naphthalene	10.532	128	98911	0.384	ng		100
10) Hexachlorobutadiene	10.836	225	27062	0.368	ng	#	100
12) 2-Methylnaphthalene	12.161	142	66107	0.385	ng		98
16) Acenaphthylene	14.055	152	83172	0.363	ng		100
17) Acenaphthene	14.398	154	62418	0.392	ng		99
18) Fluorene	15.387	166	74135	0.383	ng		100
20) 4,6-Dinitro-2-methylph...	15.578	198	50m	0.463	ng		
21) 4-Bromophenyl-phenylether	16.289	248	24127	0.359	ng		99
22) Hexachlorobenzene	16.399	284	34431	0.393	ng		99
23) Atrazine	16.569	200	13673	0.328	ng		98
24) Pentachlorophenol	16.776	266	984	0.451	ng		96
25) Phenanthrene	17.129	178	114229	0.381	ng		100
26) Anthracene	17.226	178	88921	0.356	ng		99
28) Fluoranthene	19.159	202	138442	0.375	ng		100
30) Pyrene	19.521	202	135124	0.459	ng		100
32) Benzo(a)anthracene	21.282	228	60539	0.302	ng		99
33) Chrysene	21.335	228	101553	0.412	ng		99
34) Bis(2-ethylhexyl)phtha...	21.219	149	24775	0.308	ng		100
36) Indeno(1,2,3-cd)pyrene	25.985	276	18938m	0.309	ng		
37) Benzo(b)fluoranthene	22.901	252	47542	0.353	ng		99
38) Benzo(k)fluoranthene	22.945	252	40979	0.312	ng		100
39) Benzo(a)pyrene	23.503	252	41086	0.343	ng	#	94
40) Dibenzo(a,h)anthracene	26.023	278	12203m	0.309	ng		
41) Benzo(g,h,i)perylene	26.693	276	21371	0.330	ng		100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121321\
Data File : BN017964.D
Acq On : 13 Dec 2021 10:05
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo

Quant Time: Dec 14 03:40:02 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 11 00:01:30 2021
Response via : Initial Calibration






#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.705 min Scan# 49
Delta R.T. -0.010 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05

Instrument :

BNA_N

ClientSampleId :


SSTDCCC0.4



Tgt Ion:152 Resp: 2605
Ion Ratio Lower Upper
152 100
150 154.7 121.9 182.9
115 55.9 44.1 66.1


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






12/14/2021
Supervised By :Jagrut
Upadhyay


12/14/2021



#2
1,4-Dioxane
Concen: 0.380 ng m
RT: 3.261 min Scan# 9
Delta R.T. -0.009 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05

Tgt Ion: 88 Resp: 3002
Ion Ratio Lower Upper
88 100
43 30.7 28.0 42.0
58 78.4 65.4 98.0






#3
n-Nitrosodimethylamine
Concen: 0.388 ng
RT: 3.602 min Scan# 40
Delta R.T. -0.000 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05

Instrument :

BNA_N

ClientSampleId :



SSTDCCC0.4



Tgt Ion: 42 Resp: 9280
Ion Ratio Lower Upper
42 100
74 131.1 100.6 150.8
44 7.2 5.2 7.8


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






12/14/2021
Supervised By :Jagrut
Upadhyay


12/14/2021



#4
2-Fluorophenol
Concen: 0.390 ng
RT: 5.317 min Scan# 232
Delta R.T. -0.009 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05

Tgt Ion:112 Resp: 20972
Ion Ratio Lower Upper
112 100
64 60.1 49.0 73.4
63 31.6 26.2 39.2







#5
Phenol-d6
Concen: 0.376 ng
RT: 6.885 min Scan# 403
Delta R.T. -0.009 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion: 99 Resp: 2004
Ion Ratio Lower Upper
99 100
42 20.8 17.0 25.4
71 33.9 28.0 42.0



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






12/14/2021
Supervised By :Jagrut
Upadhyay


12/14/2021



#6
bis(2-Chloroethyl)ether
Concen: 0.399 ng
RT: 7.134 min Scan# 429
Delta R.T. -0.009 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05

Tgt Ion: 93 Resp: 26073
Ion Ratio Lower Upper
93 100
63 77.7 62.0 93.0
95 32.6 26.2 39.4





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.481 min Scan# 729
Delta R.T. -0.013 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05

Instrument :

BNA_N



ClientSampleId :

SSTDCCC0.4

Tgt Ion:136 Resp: 103690
Ion Ratio Lower Upper
136 100
137 11.0 8.8 13.2
54 10.0 5.1 7.7
68 8.0 4.4 6.6



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






12/14/2021
Supervised By :Jagrut
Upadhyay


12/14/2021



#8
Nitrobenzene-d5
Concen: 0.321 ng
RT: 8.846 min Scan# 599
Delta R.T. -0.013 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05

Tgt Ion: 82 Resp: 21740
Ion Ratio Lower Upper
82 100
128 33.7 32.2 48.4
54 52.5 37.6 56.4







#9
Naphthalene
Concen: 0.384 ng
RT: 10.532 min Scan# 71
Delta R.T. -0.013 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:128 Resp: 9891
Ion Ratio Lower Upper
128 100
129 11.1 8.8 13.2
127 13.4 10.5 15.7



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






12/14/2021
Supervised By :Jagrut
Upadhyay


12/14/2021



#10
Hexachlorobutadiene
Concen: 0.368 ng
RT: 10.836 min Scan# 756
Delta R.T. -0.000 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05


Tgt Ion:225 Resp: 27062
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.0 51.1 76.7





#11
2-Methylnaphthalene-d10
Concen: 0.375 ng
RT: 12.085 min Scan# 911
Delta R.T. -0.005 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:152 Resp: 6840
Ion Ratio Lower Upper
152 100
151 20.5 16.4 24.6


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






12/14/2021
Supervised By :Jagrut
Upadhyay


12/14/2021



#12
2-Methylnaphthalene
Concen: 0.385 ng
RT: 12.161 min Scan# 954
Delta R.T. -0.000 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05


Tgt Ion:142 Resp: 66107
Ion Ratio Lower Upper
142 100
141 86.5 70.6 105.8
115 33.7 27.8 41.6





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.334 min Scan# 1176
Delta R.T. -0.000 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:164 Resp: 6134
Ion Ratio Lower Upper
164 100
162 101.7 82.1 123.1
160 47.6 38.2 57.2


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




12/14/2021
Supervised By :Jagrut
Upadhyay



12/14/2021




#14
2,4,6-Tribromophenol
Concen: 0.445 ng
RT: 15.844 min Scan# 1295
Delta R.T. 0.000 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05



Tgt Ion:330 Resp: 3769
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 27.0 19.7 29.5





#15
2-Fluorobiphenyl
Concen: 0.387 ng
RT: 12.964 min Scan# 1068
Delta R.T. -0.000 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05

Instrument :

BNA_N



ClientSampleId :

SSTDCCC0.4

Tgt Ion:172 Resp: 83518
Ion Ratio Lower Upper
172 100
171 34.9 28.3 42.5
170 23.5 19.1 28.7



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






12/14/2021
Supervised By :Jagrut
Upadhyay


12/14/2021



#16
Acenaphthylene
Concen: 0.363 ng
RT: 14.055 min Scan# 1154
Delta R.T. -0.000 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05


Tgt Ion:152 Resp: 83172
Ion Ratio Lower Upper
152 100
151 19.7 15.6 23.4
153 13.1 10.5 15.7





#17
Acenaphthene
Concen: 0.392 ng
RT: 14.398 min Scan# 1181
Delta R.T. -0.000 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:154 Resp: 62418
Ion Ratio Lower Upper
154 100
153 113.0 89.8 134.8
152 56.0 44.2 66.2


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




12/14/2021
Supervised By :Jagrut
Upadhyay



12/14/2021




#18
Fluorene
Concen: 0.383 ng
RT: 15.387 min Scan# 1259
Delta R.T. -0.000 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05




Tgt Ion:166 Resp: 74135
Ion Ratio Lower Upper
166 100
165 97.2 77.5 116.3
167 13.5 10.8 16.2





#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.092 min Scan# 11733
Delta R.T. -0.000 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:188 Resp: 11733
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.8 6.2 9.2


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




12/14/2021
Supervised By :Jagrut
Upadhyay



12/14/2021




#20
4,6-Dinitro-2-methylphenol
Concen: 0.463 ng m
RT: 15.578 min Scan# 1274
Delta R.T. 0.102 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05



Tgt Ion:198 Resp: 50
Ion Ratio Lower Upper
198 100
182 79.4 70.8 106.2
77 0.0 0.0 0.0






#21
4-Bromophenyl-phenylether
Concen: 0.359 ng
RT: 16.289 min Scan# 1331
Delta R.T. -0.000 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05

Instrument :

BNA_N

ClientSampleId :



SSTDCCC0.4



Tgt Ion:248 Resp: 2412
Ion Ratio Lower Upper
248 100
250 97.5 78.6 117.8
141 42.8 34.6 51.8


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




12/14/2021
Supervised By :Jagrut
Upadhyay



12/14/2021




#22
Hexachlorobenzene
Concen: 0.393 ng
RT: 16.399 min Scan# 1340
Delta R.T. -0.000 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05




Tgt Ion:284 Resp: 34431
Ion Ratio Lower Upper
284 100
142 27.0 21.2 31.8
249 29.2 23.2 34.8





#23
Atrazine
Concen: 0.328 ng
RT: 16.569 min Scan# 1354
Delta R.T. 0.000 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:200 Resp: 1367
Ion Ratio Lower Upper
200 100
173 23.6 18.2 27.2
215 51.2 41.8 62.8


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




12/14/2021
Supervised By :Jagrut
Upadhyay



12/14/2021




#24
Pentachlorophenol
Concen: 0.451 ng
RT: 16.776 min Scan# 1371
Delta R.T. -0.000 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05




Tgt Ion:266 Resp: 984
Ion Ratio Lower Upper
266 100
264 61.7 50.6 76.0
268 62.1 53.7 80.5





#25
Phenanthrene
Concen: 0.381 ng
RT: 17.129 min Scan# 1400
Delta R.T. -0.000 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:178 Resp: 114229
Ion Ratio Lower Upper
178 100
176 19.0 15.0 22.6
179 15.4 12.4 18.6


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




12/14/2021
Supervised By :Jagrut
Upadhyay



12/14/2021




#26
Anthracene
Concen: 0.356 ng
RT: 17.226 min Scan# 1408
Delta R.T. -0.000 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05



Tgt Ion:178 Resp: 88921
Ion Ratio Lower Upper
178 100
176 18.3 14.3 21.5
179 15.3 12.2 18.4





#27
Fluoranthene-d10
Concen: 0.357 ng
RT: 19.129 min Scan# 1697
Delta R.T. -0.000 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05

Instrument :

BNA_N



ClientSampleId :

SSTDCCC0.4

Tgt Ion:212 Resp: 13584
Ion Ratio Lower Upper
212 100
106 10.3 8.3 12.5
104 5.7 4.6 6.8



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






12/14/2021
Supervised By :Jagrut
Upadhyay


12/14/2021



#28
Fluoranthene
Concen: 0.375 ng
RT: 19.159 min Scan# 1703
Delta R.T. -0.000 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05

Tgt Ion:202 Resp: 138442
Ion Ratio Lower Upper
202 100
101 9.0 7.3 10.9
203 17.1 13.7 20.5






#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.293 min Scan# 1995
Delta R.T. -0.000 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05

Instrument :

BNA_N

ClientSampleId :



SSTDCCC0.4



Tgt Ion:240 Resp: 75020
Ion Ratio Lower Upper
240 100
120 9.3 5.8 8.8
236 26.9 21.0 31.6


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




12/14/2021
Supervised By :Jagrut
Upadhyay



12/14/2021




#30
Pyrene
Concen: 0.459 ng
RT: 19.521 min Scan# 1776
Delta R.T. -0.000 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05



Tgt Ion:202 Resp: 135124
Ion Ratio Lower Upper
202 100
200 20.8 16.5 24.7
203 17.4 14.0 21.0






#31
Terphenyl-d14
Concen: 0.415 ng
RT: 19.740 min Scan# 1820
Delta R.T. -0.000 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05

Instrument :

BNA_N

ClientSampleId :


SSTDCCC0.4



Tgt Ion:244 Resp: 76714
Ion Ratio Lower Upper
244 100
212 7.8 6.2 9.2
122 10.0 7.5 11.3


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






12/14/2021
Supervised By :Jagrut
Upadhyay


12/14/2021



#32
Benzo(a)anthracene
Concen: 0.302 ng
RT: 21.282 min Scan# 1994
Delta R.T. 0.000 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05


Tgt Ion:228 Resp: 60539
Ion Ratio Lower Upper
228 100
226 27.2 21.0 31.4
229 19.6 15.8 23.8





#33
Chrysene
Concen: 0.412 ng
RT: 21.335 min Scan# 1999
Delta R.T. 0.010 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:228 Resp: 10155
Ion Ratio Lower Upper
228 100
226 29.3 23.8 35.8
229 19.7 15.6 23.4


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




12/14/2021
Supervised By :Jagrut
Upadhyay



12/14/2021




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.308 ng
RT: 21.219 min Scan# 1988
Delta R.T. 0.011 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05



Tgt Ion:149 Resp: 24775
Ion Ratio Lower Upper
149 100
167 29.0 23.4 35.0
279 4.2 3.6 5.4






#35
Perylene-d12
Concen: 0.400 ng
RT: 23.596 min Scan# 2445
Delta R.T. 0.007 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05

Instrument :

BN017964.D

ClientSampleId :


SSTDCCC0.4



Tgt Ion:264 Resp: 40750
Ion Ratio Lower Upper
264 100
260 25.6 20.2 30.4
265 22.8 18.2 27.4


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






12/14/2021
Supervised By :Jagrut
Upadhyay


12/14/2021



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.309 ng m
RT: 25.985 min Scan# 3145
Delta R.T. 0.024 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05

Tgt Ion:276 Resp: 18938
Ion Ratio Lower Upper
276 100
138 6.0 15.2 22.8#
277 19.2 19.8 29.6#






#37
Benzo(b)fluoranthene
Concen: 0.353 ng
RT: 22.901 min Scan# 21
Delta R.T. 0.007 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05

Instrument :

BN017964.D

ClientSampleId :



SSTDCCC0.4



Tgt Ion:252 Resp: 4754
Ion Ratio Lower Upper
252 100
253 22.4 17.9 26.9
125 7.4 6.8 10.2


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






12/14/2021
Supervised By :Jagrut
Upadhyay


12/14/2021



#38
Benzo(k)fluoranthene
Concen: 0.312 ng
RT: 22.945 min Scan# 2257
Delta R.T. 0.007 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05

Tgt Ion:252 Resp: 40979
Ion Ratio Lower Upper
252 100
253 22.1 17.8 26.6
125 7.4 6.1 9.1






#39
Benzo(a)pyrene
Concen: 0.343 ng
RT: 23.503 min Scan# 2417
Delta R.T. 0.010 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05

Instrument :

BN017964.D

ClientSampleId :



SSTDCCC0.4



Tgt Ion:252 Resp: 41080
Ion Ratio Lower Upper
252 100
253 23.3 18.8 28.2
125 9.5 12.2 18.2


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




12/14/2021
Supervised By :Jagrut
Upadhyay



12/14/2021




#40
Dibenzo(a,h)anthracene
Concen: 0.309 ng m
RT: 26.023 min Scan# 3156
Delta R.T. 0.035 min
Lab File: BN017964.D
Acq: 13 Dec 2021 10:05




Tgt Ion:278 Resp: 12203
Ion Ratio Lower Upper
278 100
139 15.8 12.0 18.0
279 28.3 20.8 31.2





#41
 Benzo(g,h,i)perylene
 Concen: 0.330 ng
 RT: 26.693 min Scan# 31
 Delta R.T. 0.023 min
 Lab File: BN017964.D
 Acq: 13 Dec 2021 10:05

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4





Tgt Ion: 276 Resp: 2137

Ion	Ratio	Lower	Upper
276	100		
277	24.8	19.6	29.4
138	17.4	13.8	20.8

Manual Integrations
 APPROVED

Reviewed By :Christian
 Giraldo



12/14/2021
 Supervised By :Jagrut
 Upadhyay

12/14/2021