

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121322\
 Data File : BN023158.D
 Acq On : 13 Dec 2022 10:58
 Operator : CG/JU
 Sample : PB149494BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK494

Quant Time: Dec 13 23:58:19 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN112622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 13 23:52:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.015	152	9196	0.400	ng/u1	0.00
4) Naphthalene-d8	10.836	136	30202	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.662	164	18146	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.406	188	41700	0.400	ng/u1	0.00
17) Chrysene-d12	21.590	240	34916	0.400	ng/u1	0.00
23) Perylene-d12	24.045	264	34026	0.400	ng/u1	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.361	96	48558	4.400	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.420	152	11997	0.261	ng/u1	0.00
18) Fluoranthene-d10	19.433	212	28725	0.223	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121322\
Data File : BN023158.D
Acq On : 13 Dec 2022 10:58
Operator : CG/JU
Sample : PB149494BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK494

Quant Time: Dec 13 23:58:19 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN112622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 13 23:52:47 2022
Response via : Initial Calibration

