

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121322\
 Data File : BN023174.D
 Acq On : 13 Dec 2022 21:23
 Operator : CG/JU
 Sample : PB149496BL
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK496

Quant Time: Dec 14 00:00:13 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN112622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 13 23:52:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.015	152	6008	0.400	ng/u1	0.00
4) Naphthalene-d8	10.836	136	19928	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.662	164	11770	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.406	188	27014	0.400	ng/u1	0.00
17) Chrysene-d12	21.587	240	23989	0.400	ng/u1	0.00
23) Perylene-d12	24.045	264	22679	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.357	96	30508	4.231	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.420	152	7610	0.251	ng/u1	0.00
18) Fluoranthene-d10	19.428	212	17744	0.200	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121322\
Data File : BN023174.D
Acq On : 13 Dec 2022 21:23
Operator : CG/JU
Sample : PB149496BL
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK496

Quant Time: Dec 14 00:00:13 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN112622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 13 23:52:47 2022
Response via : Initial Calibration

