

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121324\
 Data File : BN035600.D
 Acq On : 13 Dec 2024 17:08
 Operator : RC/JU
 Sample : PB165482BL
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK482

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/14/2024
 Supervised By :mohammad ahmed 12/14/2024

Quant Time: Dec 13 21:06:07 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN111624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 10 16:10:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.293	152	1962	0.400	ng/u1	0.00
4) Naphthalene-d8	10.042	136	5946	0.400	ng/u1	0.00
9) Acenaphthene-d10	13.958	164	4103	0.400	ng/u1	0.00
13) Phenanthrene-d10	16.726	188	9154m	0.400	ng/u1	0.00
17) Chrysene-d12	20.970	240	7431	0.400	ng/u1	# 0.00
23) Perylene-d12	23.062	264	7792m	0.400	ng/u1	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.959	96	9239	5.152	ng/u1	0.00
6) 2-Methylnaphthalene-d10	11.642	152	2306	0.265	ng/u1	0.00
18) Fluoranthene-d10	18.777	212	6179	0.301	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121324\
Data File : BN035600.D
Acq On : 13 Dec 2024 17:08
Operator : RC/JU
Sample : PB165482BL
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK482

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/14/2024
Supervised By :mohammad ahmed 12/14/2024

Quant Time: Dec 13 21:06:07 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN111624.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 10 16:10:57 2024
Response via : Initial Calibration

