

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121421\
 Data File : BN017991.D
 Acq On : 14 Dec 2021 05:36
 Operator : CG/JU
 Sample : M4957-07DL 2X
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE125D1-20211206DL

Quant Time: Dec 14 07:27:34 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 14 01:08:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.688	152	33059	0.400	ng	0.00
7) Naphthalene-d8	10.470	136	133590	0.400	ng	0.00
13) Acenaphthene-d10	14.319	164	85464	0.400	ng	0.00
19) Phenanthrene-d10	17.066	188	179592	0.400	ng	0.00
29) Chrysene-d12	21.270	240	183325	0.400	ng	0.00
35) Perylene-d12	23.549	264	183481	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.300	112	3778	0.053	ng	0.00
5) Phenol-d6	6.877	99	1841	0.026	ng	0.00
8) Nitrobenzene-d5	8.835	82	13040	0.138	ng	0.00
11) 2-Methylnaphthalene-d10	12.069	152	26492	0.120	ng	0.00
14) 2,4,6-Tribromophenol	15.816	330	4981	0.281	ng	0.00
15) 2-Fluorobiphenyl	12.949	172	40440	0.136	ng	0.00
27) Fluoranthene-d10	19.102	212	97613	0.163	ng	0.00
31) Terphenyl-d14	19.708	244	81249	0.213	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.244	88	63753	3.094	ng	56
34) Bis(2-ethylhexyl)phtha...	21.195	149	360290	1.331	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121421\
Data File : BN017991.D
Acq On : 14 Dec 2021 05:36
Operator : CG/JU
Sample : M4957-07DL 2X
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE125D1-20211206DL

Quant Time: Dec 14 07:27:34 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 14 01:08:04 2021
Response via : Initial Calibration

