


2,4,6-Tribromophenol

Response Ratio



$$\text{Response} = 2.529 \times 10^{-1} \times \text{Amt} - 1.192 \times 10^{-1}$$


Coef of Det (r^2) = 0.999558 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN121421.M

Calibration Table Last Updated: Tue Dec 14 01:08:04 2021

4,6-Dinitro-2-methylphenol


Response Ratio



$R = 3.653e-003 A^2 + 4.113e-002 A - 2.234e-002$
 Coef of Det (r^2) = 0.998967 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN121421.M
 Calibration Table Last Updated: Tue Dec 14 01:08:04 2021

Atrazine

Response Ratio



$$\text{Response} = 2.084\text{e-}001 * \text{Amt} - 4.199\text{e-}002$$


Coef of Det (r^2) = 0.999173 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN121421.M

Calibration Table Last Updated: Tue Dec 14 01:08:04 2021

Pentachlorophenol

Response Ratio



$R = 4.080e-003 A^2 + 7.177e-002 A - 4.506e-002$
 Coef of Det (r^2) = 0.998923 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN121421.M
 Calibration Table Last Updated: Tue Dec 14 01:08:04 2021