

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121421\
 Data File : BN018009.D
 Acq On : 14 Dec 2021 18:18
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 15 02:20:47 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 14 01:08:04 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	125	0.00
2	1,4-Dioxane	0.400	0.385	3.8	123	0.00
3	n-Nitrosodimethylamine	0.400	0.371	7.3	121	0.00
4 S	2-Fluorophenol	0.400	0.429	-7.2	140	0.00
5 S	Phenol-d6	0.400	0.439	-9.7	148	0.00
6	bis(2-Chloroethyl)ether	0.400	0.403	-0.8	129	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	128	0.00
8 S	Nitrobenzene-d5	0.400	0.348	13.0	124	-0.01
9	Naphthalene	0.400	0.380	5.0	125	-0.01
10	Hexachlorobutadiene	0.400	0.378	5.5	124	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.391	2.3	131	0.00
12	2-Methylnaphthalene	0.400	0.397	0.8	131	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	125	0.00
14 S	2,4,6-Tribromophenol	0.400	0.481	-20.2	185	0.00
15 S	2-Fluorobiphenyl	0.400	0.401	-0.3	133	0.00
16	Acenaphthylene	0.400	0.356	11.0	117	-0.01
17	Acenaphthene	0.400	0.384	4.0	125	0.00
18	Fluorene	0.400	0.393	1.8	128	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	129	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.344	14.0	143	0.00
21	4-Bromophenyl-phenylether	0.400	0.373	6.8	129	0.00
22	Hexachlorobenzene	0.400	0.374	6.5	124	0.00
23	Atrazine	0.400	0.371	7.3	142	0.00
24	Pentachlorophenol	0.400	0.596	-49.0#	323	-0.01
25	Phenanthrene	0.400	0.395	1.3	133	0.00
26	Anthracene	0.400	0.375	6.3	128	0.00
27 SURR	Fluoranthene-d10	0.400	0.381	4.8	127	0.00
28	Fluoranthene	0.400	0.396	1.0	129	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	135	0.00
30	Pyrene	0.400	0.375	6.3	126	0.00
31 S	Terphenyl-d14	0.400	0.393	1.8	136	0.00
32	Benzo(a)anthracene	0.400	0.397	0.8	148	-0.01
33	Chrysene	0.400	0.381	4.8	129	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.360	10.0	139	0.00
35 I	Perylene-d12	0.400	0.400	0.0	131	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.381	4.8	132	-0.01
37	Benzo(b)fluoranthene	0.400	0.410	-2.5	144	0.00
38	Benzo(k)fluoranthene	0.400	0.419	-4.7	144	0.00
39 C	Benzo(a)pyrene	0.400	0.385	3.8	131	0.00
40	Dibenzo(a,h)anthracene	0.400	0.384	4.0	142	-0.01
41	Benzo(g,h,i)perylene	0.400	0.348	13.0	118	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4