

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121620\
 Data File : BN013162.D
 Acq On : 16 Dec 2020 10:43
 Operator : CG/JU
 Sample : L4996-21
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE135D1-20201208

Quant Time: Dec 16 12:37:34 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 15 16:49:20 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.346	152	675	0.40	ng	0.00
7) Naphthalene-d8	10.125	136	2104	0.40	ng	# 0.01
13) Acenaphthene-d10	14.016	164	1373	0.40	ng	0.00
19) Phenanthrene-d10	16.788	188	2943	0.40	ng	# 0.01
29) Chrysene-d12	21.016	240	3140	0.40	ng	# 0.01
36) Perylene-d12	23.102	264	3693	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	4.978	112	329	0.12	ng	0.00
5) Phenol-d6	6.549	99	157	0.06	ng	0.00
8) Nitrobenzene-d5	8.541	82	622	0.29	ng	0.03
11) 2-Methylnaphthalene-d10	11.746	152	1082	0.29	ng	0.02
14) 2,4,6-Tribromophenol	15.552	330	222	0.26	ng	0.01
15) 2-Fluorobiphenyl	12.633	172	1625	0.28	ng	0.00
27) Fluoranthene-d10	18.826	212	3952	0.41	ng	0.00
31) Terphenyl-d14	19.442	244	3459	0.43	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.941	88	223	0.19	ng	# 73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121620\
Data File : BN013162.D
Acq On : 16 Dec 2020 10:43
Operator : CG/JU
Sample : L4996-21
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE135D1-20201208

Quant Time: Dec 16 12:37:34 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121620.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 15 16:49:20 2020
Response via : Initial Calibration

