

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121620\
 Data File : BN013146.D
 Acq On : 15 Dec 2020 16:16
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN121620

Quant Time: Dec 15 17:52:18 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 15 16:49:20 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	120	0.00
2	1,4-Dioxane	0.400	0.370	7.5	111	0.00
3	n-Nitrosodimethylamine	0.400	0.386	3.5	106	-0.02
4 S	2-Fluorophenol	0.400	0.434	-8.5	113	0.00
5 S	Phenol-d6	0.400	0.417	-4.2	125	0.00
6	bis(2-Chloroethyl)ether	0.400	0.416	-4.0	141	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	118	-0.01
8 S	Nitrobenzene-d5	0.400	0.393	1.8	121	-0.01
9	Naphthalene	0.400	0.394	1.5	115	-0.01
10	Hexachlorobutadiene	0.400	0.411	-2.7	119	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.401	-0.3	119	-0.01
12	2-Methylnaphthalene	0.400	0.394	1.5	118	-0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	114	0.00
14 S	2,4,6-Tribromophenol	0.400	0.375	6.3	118	-0.01
15 S	2-Fluorobiphenyl	0.400	0.407	-1.7	118	-0.01
16	Acenaphthylene	0.400	0.394	1.5	114	0.00
17	Acenaphthene	0.400	0.405	-1.3	117	0.00
18	Fluorene	0.400	0.397	0.8	118	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	114	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.394	1.5	114	-0.03
21	4-Bromophenyl-phenylether	0.400	0.296	26.0#	92	0.00
22	Hexachlorobenzene	0.400	0.389	2.8	107	0.00
23	Atrazine	0.400	0.363	9.3	113	-0.01
24	Pentachlorophenol	0.400	0.368	8.0	115	-0.01
25	Phenanthrene	0.400	0.384	4.0	115	-0.01
26	Anthracene	0.400	0.379	5.3	117	-0.01
27 SURR	Fluoranthene-d10	0.400	0.383	4.3	114	0.00
28	Fluoranthene	0.400	0.381	4.8	113	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	116	0.01
30	Pyrene	0.400	0.395	1.3	116	0.00
31 S	Terphenyl-d14	0.400	0.402	-0.5	117	0.00
32	Benzo(a)anthracene	0.400	0.382	4.5	112	0.00
33	Chrysene	0.400	0.408	-2.0	122	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.342	14.5	122	0.01
35	Indeno(1,2,3-cd)pyrene	0.400	0.419	-4.7	112	0.01
36 I	Perylene-d12	0.400	0.400	0.0	110	0.00
37	Benzo(b)fluoranthene	0.400	0.404	-1.0	112	0.00
38	Benzo(k)fluoranthene	0.400	0.390	2.5	114	0.00
39 C	Benzo(a)pyrene	0.400	0.406	-1.5	111	0.00
40	Dibenzo(a,h)anthracene	0.400	0.394	1.5	109	0.00
41	Benzo(g,h,i)perylene	0.400	0.415	-3.7	113	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN121620