

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121620\
 Data File : BN013165.D
 Acq On : 16 Dec 2020 12:32
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 16 13:17:26 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 15 16:49:20 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	101	0.00
2	1,4-Dioxane	0.400	0.424	-6.0	107	0.00
3	n-Nitrosodimethylamine	0.400	0.426	-6.5	99	0.00
4 S	2-Fluorophenol	0.400	0.460	-15.0	101	0.00
5 S	Phenol-d6	0.400	0.397	0.8	100	0.00
6	bis(2-Chloroethyl)ether	0.400	0.314	21.5	90	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	94	0.00
8 S	Nitrobenzene-d5	0.400	0.411	-2.7	101	0.00
9	Naphthalene	0.400	0.399	0.3	93	0.00
10	Hexachlorobutadiene	0.400	0.433	-8.2	100	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.392	2.0	92	0.00
12	2-Methylnaphthalene	0.400	0.392	2.0	93	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	91	0.00
14 S	2,4,6-Tribromophenol	0.400	0.395	1.3	100	0.00
15 S	2-Fluorobiphenyl	0.400	0.409	-2.2	94	0.00
16	Acenaphthylene	0.400	0.394	1.5	91	0.00
17	Acenaphthene	0.400	0.408	-2.0	94	0.01
18	Fluorene	0.400	0.399	0.3	94	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	93	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.385	3.8	89	-0.01
21	4-Bromophenyl-phenylether	0.400	0.385	3.8	98	0.00
22	Hexachlorobenzene	0.400	0.396	1.0	89	0.00
23	Atrazine	0.400	0.362	9.5	92	0.00
24	Pentachlorophenol	0.400	0.395	1.3	101	-0.01
25	Phenanthrene	0.400	0.383	4.3	93	-0.01
26	Anthracene	0.400	0.369	7.8	93	-0.01
27 SURR	Fluoranthene-d10	0.400	0.372	7.0	90	0.00
28	Fluoranthene	0.400	0.382	4.5	92	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	92	0.00
30	Pyrene	0.400	0.398	0.5	92	0.00
31 S	Terphenyl-d14	0.400	0.402	-0.5	93	0.00
32	Benzo(a)anthracene	0.400	0.399	0.3	93	0.00
33	Chrysene	0.400	0.399	0.3	94	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.140	65.0#	64	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.447	-11.7	95	0.00
36 I	Perylene-d12	0.400	0.400	0.0	90	0.00
37	Benzo(b)fluoranthene	0.400	0.424	-6.0	97	0.00
38	Benzo(k)fluoranthene	0.400	0.406	-1.5	98	0.00
39 C	Benzo(a)pyrene	0.400	0.419	-4.7	94	0.00
40	Dibenzo(a,h)anthracene	0.400	0.425	-6.2	97	0.00
41	Benzo(g,h,i)perylene	0.400	0.436	-9.0	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4