

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN121719\
 Data File : BN009008.D
 Acq On : 17 Dec 2019 10:05
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 18 12:21:42 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN120919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 09 17:02:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	119	-0.02
2	1,4-Dioxane	0.375	0.378	-0.8	101	0.00
3	n-Nitrosodimethylamine	0.518	0.461	11.0	109	0.00
4 S	2-Fluorophenol	1.017	1.069	-5.1	115	-0.02
5 S	Phenol-d6	1.406	1.518	-8.0	120	-0.02
6	bis(2-Chloroethyl)ether	1.325	1.399	-5.6	113	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	126	-0.01
8 S	Nitrobenzene-d5	0.339	0.336	0.9	120	-0.01
9	Naphthalene	1.091	1.159	-6.2	120	-0.01
10	Hexachlorobutadiene	0.209	0.217	-3.8	117	-0.03
11 SURR	2-Methylnaphthalene-d10	0.684	0.725	-6.0	120	-0.02
12	2-Methylnaphthalene	0.813	0.882	-8.5	121	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	124	0.00
14 S	2,4,6-Tribromophenol	0.192	0.192	0.0	120	-0.02
15 S	2-Fluorobiphenyl	1.519	1.527	-0.5	117	-0.02
16	Acenaphthylene	1.830	1.997	-9.1	122	-0.02
17	Acenaphthene	1.215	1.286	-5.8	117	-0.02
18	Fluorene	1.634	1.719	-5.2	115	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	120	-0.01
20	4-Bromophenyl-phenylether	0.239	0.248	-3.8	114	-0.01
21	Hexachlorobenzene	0.273	0.286	-4.8	115	-0.01
22	Pentachlorophenol	0.098	0.096	2.0	121	-0.01
23	Phenanthrene	1.259	1.302	-3.4	113	-0.01
24	Anthracene	1.109	1.136	-2.4	114	-0.01
25 SURR	Fluoranthene-d10	1.211	1.190	1.7	110	-0.01
26	Fluoranthene	1.498	1.506	-0.5	110	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
28	Pvrene	1.402	1.671	-19.2	110	0.00
29 S	Terphenyl-d14	0.913	1.025	-12.3	109	0.00
30	Benzo(a)anthracene	1.418	1.548	-9.2	104	0.00
31	Chrysene	1.470	1.554	-5.7	98	-0.02
32	Bis(2-ethylhexyl)phthalate	0.664	0.811	-22.1	132	0.00
33	Indeno(1,2,3-cd)pyrene	1.556	1.289	17.2	86	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	94	-0.02
35	Benzo(b)fluoranthene	1.432	1.540	-7.5	87	-0.02
36	Benzo(k)fluoranthene	1.442	1.569	-8.8	90	-0.01
37 C	Benzo(a)pyrene	1.286	1.352	-5.1	88	-0.02
38	Dibenzo(a,h)anthracene	1.206	1.184	1.8	85	-0.02
39	Benzo(a,h,i)perylene	1.191	1.203	-1.0	86	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN121719\
Data File : BN009008.D
Acq On : 17 Dec 2019 10:05
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 18 12:21:42 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN120919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 09 17:02:53 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			