

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN121719\
 Data File : BN009018.D
 Acq On : 17 Dec 2019 18:54
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 18 12:41:09 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN120919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 09 17:02:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	137	-0.02
2	1,4-Dioxane	0.375	0.356	5.1	110	0.00
3	n-Nitrosodimethylamine	0.518	0.502	3.1	137	0.00
4 S	2-Fluorophenol	1.017	1.578	-55.2#	197#	0.00
5 S	Phenol-d6	1.406	1.868	-32.9#	171#	-0.02
6	bis(2-Chloroethyl)ether	1.325	1.385	-4.5	129	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	151#	-0.01
8 S	Nitrobenzene-d5	0.339	0.339	0.0	146	-0.01
9	Naphthalene	1.091	1.154	-5.8	143	-0.01
10	Hexachlorobutadiene	0.209	0.221	-5.7	143	-0.03
11 SURR	2-Methylnaphthalene-d10	0.684	0.737	-7.7	146	-0.02
12	2-Methylnaphthalene	0.813	0.893	-9.8	147	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	153#	0.00
14 S	2,4,6-Tribromophenol	0.192	0.182	5.2	139	-0.02
15 S	2-Fluorobiphenyl	1.519	1.513	0.4	143	-0.02
16	Acenaphthylene	1.830	2.047	-11.9	154#	-0.02
17	Acenaphthene	1.215	1.296	-6.7	145	-0.02
18	Fluorene	1.634	1.713	-4.8	141	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	151#	-0.01
20	4-Bromophenyl-phenylether	0.239	0.241	-0.8	140	-0.01
21	Hexachlorobenzene	0.273	0.281	-2.9	142	-0.01
22	Pentachlorophenol	0.098	0.062	36.7#	99	0.00
23	Phenanthrene	1.259	1.273	-1.1	139	-0.01
24	Anthracene	1.109	1.151	-3.8	145	0.00
25 SURR	Fluoranthene-d10	1.211	1.205	0.5	140	-0.01
26	Fluoranthene	1.498	1.500	-0.1	138	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	134	0.00
28	Pvrene	1.402	1.700	-21.3	139	0.00
29 S	Terphenyl-d14	0.913	1.043	-14.2	138	0.00
30	Benzo(a)anthracene	1.418	1.566	-10.4	130	0.00
31	Chrysene	1.470	1.556	-5.9	121	0.00
32	Bis(2-ethylhexyl)phthalate	0.664	0.982	-47.9#	198#	0.00
33	Indeno(1,2,3-cd)pyrene	1.556	1.384	11.1	114	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	122	-0.01
35	Benzo(b)fluoranthene	1.432	1.531	-6.9	111	-0.01
36	Benzo(k)fluoranthene	1.442	1.549	-7.4	115	0.00
37 C	Benzo(a)pyrene	1.286	1.365	-6.1	115	-0.01
38	Dibenzo(a,h)anthracene	1.206	1.230	-2.0	114	-0.02
39	Benzo(a,h,i)perylene	1.191	1.244	-4.5	115	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN121719\
Data File : BN009018.D
Acq On : 17 Dec 2019 18:54
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 18 12:41:09 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN120919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 09 17:02:53 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			