

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN121818\
 Data File : BN004089.D
 Acq On : 18 Dec 2018 22:34
 Operator : JU/SJ
 Sample : J6271-05
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 F9F12

Manual Integrations
 APPROVED

Sohil
 12/19/2018 4:16:43 PM

Quant Time: Dec 19 04:28:04 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN111918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 19 04:08:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	1645	0.40	ng/ul	0.00
2) Naphthalene-d8	10.63	136	7952	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	5096	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.22	188	12860m	0.40	ng/ul	0.00
16) Chrysene-d12	21.40	240	13448m	0.40	ng/ul	0.00
20) Perylene-d12	23.72	264	12437m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.22	152	3112	0.25	ng/ul	0.00
14) Fluoranthene-d10	19.24	212	10165m	0.27	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.68	128	9370	0.419	ng/ul	96
5) 2-Methylnaphthalene	12.29	142	17584	1.127	ng/ul	100
8) Acenaphthene	14.53	153	1488	0.074	ng/ul#	87
9) Fluorene	15.52	166	2215	0.094	ng/ul#	89
12) Phenanthrene	17.26	178	117765	2.790	ng/ul	100
13) Anthracene	17.35	178	9389m	0.271	ng/ul	
15) Fluoranthene	19.27	202	21446m	0.430	ng/ul	
17) Pyrene	19.64	202	114604	2.376	ng/ul	95
18) Benzo(a)anthracene	21.39	228	37049m	0.872	ng/ul	
19) Chrysene	21.44	228	73572m	1.514	ng/ul	
21) Benzo(b)fluoranthene	23.02	252	19516m	0.379	ng/ul	
23) Benzo(a)pyrene	23.62	252	23723m	0.532	ng/ul	
24) Indeno(1,2,3-cd)pyrene	26.08	276	11199m	0.219	ng/ul	
25) Dibenzo(a,h)anthracene	26.08	278	8020m	0.195	ng/ul	
26) Benzo(g,h,i)perylene	26.82	276	24812m	0.559	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN121818\
Data File : BN004089.D
Acq On : 18 Dec 2018 22:34
Operator : JU/SJ
Sample : J6271-05
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
F9F12

Quant Time: Dec 19 04:28:04 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 19 04:08:55 2018
Response via : Initial Calibration

Manual Integrations
APPROVED

Sohil
12/19/2018 4:16:43 PM

