

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121924\
 Data File : BN035752.D
 Acq On : 20 Dec 2024 21:21
 Operator : RC/JU
 Sample : P5261-09DL 2X
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE134D3-20241210DL

Quant Time: Dec 21 01:43:50 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 19 23:06:19 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 12/23/2024
 Supervised By :mohammad ahmed 12/24/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.271	152	3092	0.400	ng	0.00
7) Naphthalene-d8	10.020	136	8022	0.400	ng	# 0.00
13) Acenaphthene-d10	13.924	164	5392	0.400	ng	-0.01
19) Phenanthrene-d10	16.698	188	11029	0.400	ng	# 0.00
29) Chrysene-d12	20.955	240	9477	0.400	ng	# 0.00
35) Perylene-d12	23.044	264	9623	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	4.938	112	457	0.059	ng	0.00
5) Phenol-d6	6.491	99	227	0.024	ng	0.00
8) Nitrobenzene-d5	8.418	82	903m	0.184	ng	0.01
11) 2-Methylnaphthalene-d10	11.620	152	2145	0.171	ng	0.00
14) 2,4,6-Tribromophenol	15.453	330	384	0.100	ng	0.00
15) 2-Fluorobiphenyl	12.543	172	3190	0.157	ng	0.00
27) Fluoranthene-d10	18.761	212	6495	0.208	ng	0.00
31) Terphenyl-d14	19.388	244	5542	0.296	ng	0.00
Target Compounds						
2) 1,4-Dioxane	2.981	88	9199	3.112	ng	99
34) Bis(2-ethylhexyl)phtha...	20.911	149	586	0.045	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121924\
Data File : BN035752.D
Acq On : 20 Dec 2024 21:21
Operator : RC/JU
Sample : P5261-09DL 2X
Misc :
ALS Vial : 43 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE134D3-20241210DL

Quant Time: Dec 21 01:43:50 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 19 23:06:19 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 12/23/2024
Supervised By :mohammad ahmed 12/24/2024

