

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121924\
 Data File : BN035761.D
 Acq On : 21 Dec 2024 02:46
 Operator : RC/JU
 Sample : P5317-14
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RW9-MW01D1-20241217

Quant Time: Dec 21 03:33:35 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 19 23:06:19 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 12/23/2024
 Supervised By :mohammad ahmed 12/24/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.264	152	2847	0.400	ng	0.00
7) Naphthalene-d8	10.009	136	7595	0.400	ng	-0.01
13) Acenaphthene-d10	13.925	164	5108	0.400	ng	-0.01
19) Phenanthrene-d10	16.698	188	9580	0.400	ng	# 0.00
29) Chrysene-d12	20.956	240	8327	0.400	ng	0.00
35) Perylene-d12	23.038	264	8882	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	4.932	112	790	0.111	ng	0.00
5) Phenol-d6	6.484	99	470	0.055	ng	0.00
8) Nitrobenzene-d5	8.408	82	1644m	0.354	ng	0.00
11) 2-Methylnaphthalene-d10	11.616	152	4037	0.340	ng	0.00
14) 2,4,6-Tribromophenol	15.453	330	681	0.188	ng	0.00
15) 2-Fluorobiphenyl	12.539	172	6051	0.313	ng	0.00
27) Fluoranthene-d10	18.757	212	12187	0.449	ng	0.00
31) Terphenyl-d14	19.384	244	9337	0.568	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.982	88	798	0.293	ng	# 91
34) Bis(2-ethylhexyl)phtha...	20.902	149	1486	0.129	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121924\
Data File : BN035761.D
Acq On : 21 Dec 2024 02:46
Operator : RC/JU
Sample : P5317-14
Misc :
ALS Vial : 52 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW9-MW01D1-20241217

Quant Time: Dec 21 03:33:35 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 19 23:06:19 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 12/23/2024
Supervised By :mohammad ahmed 12/24/2024

