

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122022\
 Data File : BN023290.D
 Acq On : 20 Dec 2022 09:32
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020376

Quant Time: Dec 21 04:12:46 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN121522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 21 04:11:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.998	152	312482	20.000	ng/u1	0.00
20) Naphthalene-d8	10.816	136	1386651	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.645	164	879208	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.392	188	1952797	20.000	ng/u1	0.00
79) Chrysene-d12	21.580	240	1737939	20.000	ng/u1	0.00
88) Perylene-d12	24.027	264	1726416	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.346	96	57569	7.622	ng/uL	0.00
4) Pyridine-d5	3.775	84	431463	19.352	ng/u1	0.00
7) Phenol-d5	7.151	99	547519	18.839	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.322	67	329301	19.471	ng/u1	0.00
11) 2-Chlorophenol-d4	7.522	132	443290	20.242	ng/u1	0.00
15) 4-Methylphenol-d8	8.710	113	442981	18.984	ng/u1	0.00
21) Nitrobenzene-d5	9.169	128	221750	20.698	ng/u1	0.00
24) 2-Nitrophenol-d4	9.892	143	240998	21.357	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.434	165	449868	21.118	ng/u1	0.00
31) 4-Chloroaniline-d4	10.957	131	689499	20.116	ng/u1	0.00
46) Dimethylphthalate-d6	14.057	166	1335642	20.272	ng/u1	0.00
49) Acenaphthylene-d8	14.339	160	1586620	21.473	ng/u1	0.00
54) 4-Nitrophenol-d4	14.839	143	268731	18.262	ng/u1	0.00
60) Fluorene-d10	15.639	176	1182624	21.071	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.757	200	218282	18.989	ng/u1	0.00
73) Anthracene-d10	17.492	188	1858911	21.241	ng/u1	0.00
81) Pyrene-d10	19.786	212	2169716	22.950	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.868	264	1752955	20.715	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.381	88	60713	7.756	ng/uL	95
5) Pyridine	3.793	79	438806	19.262	ng/u1	98
6) Benzaldehyde	7.128	77	215237	20.675	ng/u1	99
8) Phenol	7.181	94	563334	19.189	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.416	93	441122	19.346	ng/u1	99
12) 2-Chlorophenol	7.557	128	454044	20.187	ng/u1	97
13) 2-Methylphenol	8.445	108	425328	19.042	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.528	45	570299	19.748	ng/u1	99
16) Acetophenone	8.828	105	688384	19.498	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.816	70	347267	19.643	ng/u1	99
18) 4-Methylphenol	8.775	108	476770	19.234	ng/u1	99
19) Hexachloroethane	9.087	117	180248	20.807	ng/u1	97
22) Nitrobenzene	9.210	77	533355	20.762	ng/u1	99
23) Isophorone	9.739	82	975381	20.129	ng/u1	99
25) 2-Nitrophenol	9.928	139	261275	21.230	ng/u1	99
26) 2,4-Dimethylphenol	9.986	107	516350	20.727	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.228	93	619411	19.983	ng/u1	100
29) 2,4-Dichlorophenol	10.463	162	450194	21.357	ng/u1	98
30) Naphthalene	10.869	128	1534464	20.714	ng/u1	100
32) 4-Chloroaniline	10.981	127	691221	20.213	ng/u1	99
33) Hexachlorobutadiene	11.151	225	266042	22.139	ng/u1	98
34) Caprolactam	11.757	113	99574	12.149	ng/u1	95
35) 4-Chloro-3-methylphenol	12.104	107	492225	20.616	ng/u1	95
36) 2-Methylnaphthalene	12.475	142	1038458	20.799	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122022\
 Data File : BN023290.D
 Acq On : 20 Dec 2022 09:32
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SSTD020376


Quant Time: Dec 21 04:12:46 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN121522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 21 04:11:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.692	142	1067914	20.756	ng/ul	96
39) 1,2,4,5-Tetrachloroben...	12.839	216	536202	21.562	ng/ul	97
40) Hexachlorocyclopentadiene	12.822	237	283503	18.953	ng/ul	90
41) 2,4,6-Trichlorophenol	13.080	196	366993	21.455	ng/ul	99
42) 2,4,5-Trichlorophenol	13.151	196	402778	21.276	ng/ul	100
43) 1,1'-Biphenyl	13.486	154	1395438	21.004	ng/ul	99
44) 2-Chloronaphthalene	13.527	162	1109094	21.438	ng/ul	98
45) 2-Nitroaniline	13.727	65	310087	20.683	ng/ul	98
47) Dimethylphthalate	14.104	163	1342275	20.189	ng/ul	99
48) 2,6-Dinitrotoluene	14.227	165	273982	21.285	ng/ul	96
50) Acenaphthylene	14.369	152	1738784	21.399	ng/ul	99
51) 3-Nitroaniline	14.551	138	259005	19.761	ng/ul	98
52) Acenaphthene	14.710	153	1175159	20.803	ng/ul	98
53) 2,4-Dinitrophenol	14.757	184	131677	16.054	ng/ul	95
55) 4-Nitrophenol	14.857	109	204709	18.809	ng/ul	91
56) Dibenzofuran	15.045	168	1645979	20.928	ng/ul	99
57) 2,4-Dinitrotoluene	15.010	165	405835	21.359	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.269	232	342176	20.877	ng/ul	97
59) Diethylphthalate	15.469	149	1363789	20.620	ng/ul	98
61) Fluorene	15.692	166	1320523	20.734	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.686	204	641528	21.016	ng/ul	99
63) 4-Nitroaniline	15.716	138	221344	17.341	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.769	198	218918	19.507	ng/ul	96
67) N-Nitrosodiphenylamine	15.898	169	1163081	21.521	ng/ul	96
68) 4-Bromophenyl-phenylether	16.580	248	418441	22.207	ng/ul	88
69) Hexachlorobenzene	16.698	284	480157	21.795	ng/ul	99
70) Atrazine	16.851	200	409586	20.739	ng/ul	97
71) Pentachlorophenol	17.039	266	290696	19.983	ng/ul	94
72) Phenanthrene	17.439	178	2165921	20.961	ng/ul	100
74) Anthracene	17.527	178	2175972	21.105	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.445	216	539132	22.200	ng/uL	98
76) Pentachlorobenzene	14.963	250	545320	22.160	ng/uL	99
77) Carbazole	17.798	167	1974823	20.801	ng/ul	100
78) Di-n-butylphthalate	18.357	149	2385027	21.157	ng/ul	99
80) Fluoranthene	19.451	202	2551965	23.210	ng/ul	98
82) Pyrene	19.815	202	2646936	22.917	ng/ul	97
83) Butylbenzylphthalate	20.709	149	1112598	23.277	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.498	252	791406	20.722	ng/ul	99
85) Benzo(a)anthracene	21.562	228	2490295	21.387	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.486	149	1594855	23.598	ng/ul	99
87) Chrysene	21.621	228	2308353	21.007	ng/ul	98
89) Di-n-octyl phthalate	22.427	149	2591154	24.784	ng/ul	100
90) Benzo(b)fluoranthene	23.274	252	2304730	21.166	ng/ul	99
91) Benzo(k)fluoranthene	23.327	252	2245906	21.673	ng/ul	98
93) Benzo(a)pyrene	23.921	252	1980780	21.167	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.591	276	2260955	19.072	ng/ul	99
95) Dibenzo(a,h)anthracene	26.603	278	1905454	19.713	ng/ul	96
96) Benzo(g,h,i)perylene	27.380	276	1780556	18.983	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Dec 21 04:12:46 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN121522.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 21 04:11:00 2022
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 7.756 ng/uL
 RT: 3.381 min Scan# 50
 Delta R.T. 0.000 min
 Lab File: BN023290.D
 Acq: 20 Dec 2022 09:32




Instrument :
 BNA_N
 ClientSampleId :
 SST020376


Tgt Ion: 88 Resp: 60713
 Ion Ratio Lower Upper
 88 100
 43 27.0 23.0 34.6
 58 65.0 55.8 83.6



#5
 Pyridine
 Concen: 19.262 ng/uL
 RT: 3.793 min Scan# 120
 Delta R.T. 0.000 min
 Lab File: BN023290.D
 Acq: 20 Dec 2022 09:32

Tgt Ion: 79 Resp: 438806
 Ion Ratio Lower Upper
 79 100
 52 70.0 56.7 85.1
 51 33.9 26.0 39.0









#6
Benzaldehyde
Concen: 20.675 ng/ul
RT: 7.128 min Scan# 687
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion: 77 Resp: 215237
Ion Ratio Lower Upper
77 100
105 96.5 78.0 117.0
106 92.4 73.2 109.8



#8
Phenol
Concen: 19.189 ng/ul
RT: 7.181 min Scan# 696
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32

Tgt Ion: 94 Resp: 563334
Ion Ratio Lower Upper
94 100
65 26.9 21.8 32.6
66 35.8 28.4 42.6









#10
Bis(2-Chloroethyl)ether
Concen: 19.346 ng/ul
RT: 7.416 min Scan# 736
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion: 93 Resp: 441122
Ion Ratio Lower Upper
93 100
63 74.5 58.9 88.3
95 32.4 26.2 39.4



#12
2-Chlorophenol
Concen: 20.187 ng/ul
RT: 7.557 min Scan# 760
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32

Tgt Ion:128 Resp: 454044
Ion Ratio Lower Upper
128 100
64 44.5 34.0 51.0
130 32.5 25.0 37.4





#13

2-Methylphenol

Concen: 19.042 ng/ul

RT: 8.445 min Scan# 911

Delta R.T. 0.000 min

Lab File: BN023290.D

Acq: 20 Dec 2022 09:32

Instrument :

BNA_N

ClientSampleId :





SSTD020376

Tgt Ion:108 Resp: 425328

Ion Ratio Lower Upper

108 100

107 90.9 71.2 106.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 19.748 ng/ul

RT: 8.528 min Scan# 925

Delta R.T. 0.000 min

Lab File: BN023290.D

Acq: 20 Dec 2022 09:32




Tgt Ion: 45 Resp: 570299


Ion Ratio Lower Upper

45 100

77 15.9 12.6 19.0

79 11.7 8.8 13.2









#16
Acetophenone
Concen: 19.498 ng/ul
RT: 8.828 min Scan# 976
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion:105 Resp: 688384
Ion Ratio Lower Upper
105 100
77 76.9 61.4 92.0
51 26.0 20.0 30.0



#17
N-Nitroso-di-n-propylamine
Concen: 19.643 ng/ul
RT: 8.816 min Scan# 974
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32

Tgt Ion: 70 Resp: 347267
Ion Ratio Lower Upper
70 100
42 56.7 45.3 67.9
101 9.6 8.1 12.1
130 21.6 17.9 26.9









#18
4-Methylphenol
Concen: 19.234 ng/ul
RT: 8.775 min Scan# 967
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion:108 Resp: 476770
Ion Ratio Lower Upper
108 100
107 111.0 87.6 131.4



#19
Hexachloroethane
Concen: 20.807 ng/ul
RT: 9.087 min Scan# 1020
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32

Tgt Ion:117 Resp: 180248
Ion Ratio Lower Upper
117 100
201 95.7 74.6 111.8
199 61.2 46.3 69.5









#22
Nitrobenzene
Concen: 20.762 ng/ul
RT: 9.210 min Scan# 1041
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion: 77 Resp: 533355
Ion Ratio Lower Upper
77 100
123 45.3 37.1 55.7
65 12.9 10.1 15.1



#23
Isophorone
Concen: 20.129 ng/ul
RT: 9.739 min Scan# 1131
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32

Tgt Ion: 82 Resp: 975381
Ion Ratio Lower Upper
82 100
95 7.1 5.4 8.2
138 17.7 14.5 21.7





#25

2-Nitrophenol

Concen: 21.230 ng/ul

RT: 9.928 min Scan# 1163

Delta R.T. 0.000 min

Lab File: BN023290.D

Acq: 20 Dec 2022 09:32

Instrument :

BNA_N

ClientSampleId :

SSTD020376





Tgt Ion:139 Resp: 261275

Ion Ratio Lower Upper

139 100

65 47.9 37.7 56.5

109 28.8 22.6 34.0



#26

2,4-Dimethylphenol

Concen: 20.727 ng/ul

RT: 9.986 min Scan# 1173

Delta R.T. 0.000 min

Lab File: BN023290.D

Acq: 20 Dec 2022 09:32




Tgt Ion:107 Resp: 516350


Ion Ratio Lower Upper

107 100

121 51.5 41.3 61.9

122 89.0 72.1 108.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.983 ng/ul
 RT: 10.228 min Scan# 1214
 Delta R.T. 0.000 min
 Lab File: BN023290.D
 Acq: 20 Dec 2022 09:32




Instrument :
 BNA_N
 ClientSampleId :
 SST020376


Tgt Ion: 93 Resp: 619411
 Ion Ratio Lower Upper
 93 100
 95 32.1 25.8 38.6
 123 11.8 9.0 13.4



#29
 2,4-Dichlorophenol
 Concen: 21.357 ng/ul
 RT: 10.463 min Scan# 1254
 Delta R.T. 0.000 min
 Lab File: BN023290.D
 Acq: 20 Dec 2022 09:32

Tgt Ion: 162 Resp: 450194
 Ion Ratio Lower Upper
 162 100
 164 63.9 52.1 78.1
 98 34.8 29.3 43.9









#30
Naphthalene
Concen: 20.714 ng/ul
RT: 10.869 min Scan# 1323
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion:128 Resp: 1534464
Ion Ratio Lower Upper
128 100
129 10.7 8.7 13.1
127 12.9 10.3 15.5



#32
4-Chloroaniline
Concen: 20.213 ng/ul
RT: 10.981 min Scan# 1342
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32

Tgt Ion:127 Resp: 691221
Ion Ratio Lower Upper
127 100
129 32.1 26.2 39.4









#33
Hexachlorobutadiene
Concen: 22.139 ng/ul
RT: 11.151 min Scan# 1371
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion:225 Resp: 266042
Ion Ratio Lower Upper
225 100
223 61.9 48.9 73.3
227 61.9 51.0 76.4



#34
Caprolactam
Concen: 12.149 ng/ul
RT: 11.757 min Scan# 1474
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32

Tgt Ion:113 Resp: 99574
Ion Ratio Lower Upper
113 100
55 147.7 123.7 185.5
56 109.4 91.7 137.5









#35
 4-Chloro-3-methylphenol
 Concen: 20.616 ng/ul
 RT: 12.104 min Scan# 1533
 Delta R.T. 0.000 min
 Lab File: BN023290.D
 Acq: 20 Dec 2022 09:32




Instrument :
 BNA_N
 ClientSampleId :
 SST020376


Tgt Ion:107 Resp: 492225
 Ion Ratio Lower Upper
 107 100
 144 27.3 21.8 32.6
 142 83.8 62.6 94.0



#36
 2-Methylnaphthalene
 Concen: 20.799 ng/ul
 RT: 12.475 min Scan# 1596
 Delta R.T. 0.000 min
 Lab File: BN023290.D
 Acq: 20 Dec 2022 09:32

Tgt Ion:142 Resp: 1038458
 Ion Ratio Lower Upper
 142 100
 141 87.7 69.9 104.9









#37
1-Methylnaphthalene
Concen: 20.756 ng/ul
RT: 12.692 min Scan# 1633
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion:142 Resp: 1067914
Ion Ratio Lower Upper
142 100
141 92.2 70.9 106.3
116 3.7 3.0 4.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 21.562 ng/ul
RT: 12.839 min Scan# 1658
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32

Tgt Ion:216 Resp: 536202
Ion Ratio Lower Upper
216 100
214 79.9 61.1 91.7
179 20.4 16.6 24.8
108 18.7 15.4 23.2





#40

Hexachlorocyclopentadiene

Concen: 18.953 ng/ul

RT: 12.822 min Scan# 1655

Delta R.T. 0.000 min

Lab File: BN023290.D

Acq: 20 Dec 2022 09:32

Instrument :

BNA_N

ClientSampleId :

SSTD020376





Tgt Ion:237 Resp: 283503

Ion Ratio Lower Upper

237 100

235 69.5 48.1 72.1

272 14.0 10.8 16.2



#41

2,4,6-Trichlorophenol

Concen: 21.455 ng/ul

RT: 13.080 min Scan# 1699

Delta R.T. 0.000 min

Lab File: BN023290.D

Acq: 20 Dec 2022 09:32




Tgt Ion:196 Resp: 366993


Ion Ratio Lower Upper

196 100

198 95.3 76.9 115.3

200 31.4 23.8 35.6









#42
2,4,5-Trichlorophenol
Concen: 21.276 ng/ul
RT: 13.151 min Scan# 1711
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion:196 Resp: 402778
Ion Ratio Lower Upper
196 100
198 93.4 74.8 112.2
200 29.3 24.2 36.2



#43
1,1'-Biphenyl
Concen: 21.004 ng/ul
RT: 13.486 min Scan# 1768
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32

Tgt Ion:154 Resp: 1395438
Ion Ratio Lower Upper
154 100
153 39.6 32.0 48.0
76 12.5 11.0 16.4





#44

2-Chloronaphthalene

Concen: 21.438 ng/ul

RT: 13.527 min Scan# 11

Delta R.T. 0.000 min

Lab File: BN023290.D

Acq: 20 Dec 2022 09:32

Instrument :

BNA_N

ClientSampleId :

SSTD020376





Tgt Ion:162 Resp: 1109094

Ion Ratio Lower Upper

162 100

164 31.7 25.4 38.0

127 37.8 31.8 47.8



#45

2-Nitroaniline

Concen: 20.683 ng/ul

RT: 13.727 min Scan# 1809

Delta R.T. 0.000 min

Lab File: BN023290.D

Acq: 20 Dec 2022 09:32




Tgt Ion: 65 Resp: 310087


Ion Ratio Lower Upper

65 100

92 69.0 57.4 86.2

138 110.9 89.0 133.6





#47

Dimethylphthalate

Concen: 20.189 ng/ul

RT: 14.104 min Scan# 1873

Delta R.T. 0.000 min

Lab File: BN023290.D

Acq: 20 Dec 2022 09:32

Instrument :

BNA_N

ClientSampleId :

SSTD020376





Tgt Ion:163 Resp: 1342275

Ion Ratio Lower Upper

163 100

194 3.9 3.3 4.9

164 9.9 8.3 12.5



#48

2,6-Dinitrotoluene

Concen: 21.285 ng/ul

RT: 14.227 min Scan# 1894

Delta R.T. 0.000 min

Lab File: BN023290.D

Acq: 20 Dec 2022 09:32




Tgt Ion:165 Resp: 273982


Ion Ratio Lower Upper

165 100

89 55.8 47.7 71.5

121 21.6 17.8 26.8









#50
Acenaphthylene
Concen: 21.399 ng/ul
RT: 14.369 min Scan# 1918
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion:152 Resp: 1738784
Ion Ratio Lower Upper
152 100
151 20.4 16.2 24.4
153 13.3 11.4 17.0



#51
3-Nitroaniline
Concen: 19.761 ng/ul
RT: 14.551 min Scan# 1949
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32

Tgt Ion:138 Resp: 259005
Ion Ratio Lower Upper
138 100
108 10.3 8.3 12.5
92 113.6 92.5 138.7









#52
Acenaphthene
Concen: 20.803 ng/ul
RT: 14.710 min Scan# 1976
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion:153 Resp: 1175159
Ion Ratio Lower Upper
153 100
152 48.7 38.9 58.3
154 91.5 70.5 105.7



#53
2,4-Dinitrophenol
Concen: 16.054 ng/ul
RT: 14.757 min Scan# 1984
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32

Tgt Ion:184 Resp: 131677
Ion Ratio Lower Upper
184 100
63 64.4 57.1 85.7
154 59.9 47.7 71.5









#55
4-Nitrophenol
Concen: 18.809 ng/ul
RT: 14.857 min Scan# 2001
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion:109 Resp: 204709
Ion Ratio Lower Upper
109 100
139 132.5 115.2 172.8
65 126.7 109.0 163.4



#56
Dibenzofuran
Concen: 20.928 ng/ul
RT: 15.045 min Scan# 2033
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32

Tgt Ion:168 Resp: 1645979
Ion Ratio Lower Upper
168 100
139 37.1 29.2 43.8





#57

2,4-Dinitrotoluene

Concen: 21.359 ng/ul

RT: 15.010 min Scan# 2027

Delta R.T. 0.000 min

Lab File: BN023290.D

Acq: 20 Dec 2022 09:32

Instrument :

BNA_N

ClientSampleId :

SSTD020376





Tgt Ion:165 Resp: 405835

Ion Ratio Lower Upper

165 100

63 42.5 36.6 54.8

182 2.8 2.5 3.7



#58

2,3,4,6-Tetrachlorophenol

Concen: 20.877 ng/ul

RT: 15.269 min Scan# 2071

Delta R.T. 0.000 min

Lab File: BN023290.D

Acq: 20 Dec 2022 09:32

Tgt Ion:232 Resp: 342176




Ion Ratio Lower Upper


232 100

131 48.5 37.9 56.9

130 2.6 1.9 2.9

166 33.1 23.8 35.8





#59

Diethylphthalate

Concen: 20.620 ng/ul

RT: 15.469 min Scan# 2105

Delta R.T. 0.000 min

Lab File: BN023290.D

Acq: 20 Dec 2022 09:32

Instrument :





BNA_N

ClientSampleId :

SSTD020376

Tgt Ion:149 Resp: 1363789

Ion	Ratio	Lower	Upper
149	100		
177	22.3	17.0	25.4
150	11.9	10.0	15.0



#61

Fluorene

Concen: 20.734 ng/ul

RT: 15.692 min Scan# 2143




Delta R.T. 0.000 min


Lab File: BN023290.D

Acq: 20 Dec 2022 09:32

Tgt Ion:166 Resp: 1320523

Ion	Ratio	Lower	Upper
166	100		
165	96.4	77.8	116.8
167	13.7	10.6	15.8









#62
4-Chlorophenyl-phenylether
Concen: 21.016 ng/ul
RT: 15.686 min Scan# 2142
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion:204 Resp: 641528
Ion Ratio Lower Upper
204 100
206 33.0 26.1 39.1
141 60.5 48.2 72.2



#63
4-Nitroaniline
Concen: 17.341 ng/ul
RT: 15.716 min Scan# 2147
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32

Tgt Ion:138 Resp: 221344
Ion Ratio Lower Upper
138 100
92 50.8 39.4 59.2
108 75.6 59.5 89.3









#66
4,6-Dinitro-2-methylphenol
Concen: 19.507 ng/ul
RT: 15.769 min Scan# 2156
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion:198 Resp: 218918
Ion Ratio Lower Upper
198 100
182 4.4 3.4 5.0
77 26.7 19.7 29.5



#67
N-Nitrosodiphenylamine
Concen: 21.521 ng/ul
RT: 15.898 min Scan# 2178
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32

Tgt Ion:169 Resp: 1163081
Ion Ratio Lower Upper
169 100
168 64.2 54.7 82.1
167 35.1 29.2 43.8









#68
4-Bromophenyl-phenylether
Concen: 22.207 ng/ul
RT: 16.580 min Scan# 2294
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion:248 Resp: 418441
Ion Ratio Lower Upper
248 100
250 84.4 81.0 121.4
141 70.7 60.8 91.2



#69
Hexachlorobenzene
Concen: 21.795 ng/ul
RT: 16.698 min Scan# 2314
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32

Tgt Ion:284 Resp: 480157
Ion Ratio Lower Upper
284 100
142 35.9 29.0 43.4
249 30.9 23.7 35.5





#70

Atrazine

Concen: 20.739 ng/ul

RT: 16.851 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN023290.D

Acq: 20 Dec 2022 09:32

Instrument :

BNA_N

ClientSampleId :

SSTD020376





Tgt Ion:200 Resp: 409586

Ion Ratio Lower Upper

200 100

173 28.3 21.3 31.9

215 48.4 39.8 59.8



#71

Pentachlorophenol

Concen: 19.983 ng/ul

RT: 17.039 min Scan# 2372

Delta R.T. 0.000 min

Lab File: BN023290.D

Acq: 20 Dec 2022 09:32




Tgt Ion:266 Resp: 290696


Ion Ratio Lower Upper

266 100

264 70.9 50.5 75.7

268 64.7 52.6 78.8









#72
 Phenanthrene
 Concen: 20.961 ng/ul
 RT: 17.439 min Scan# 2440
 Delta R.T. 0.000 min
 Lab File: BN023290.D
 Acq: 20 Dec 2022 09:32

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020376

Tgt Ion:178 Resp: 2165921




Ion	Ratio	Lower	Upper
178	100		
179	15.2	12.3	18.5
176	19.2	15.4	23.0




#74
 Anthracene
 Concen: 21.105 ng/ul
 RT: 17.527 min Scan# 2455
 Delta R.T. 0.000 min
 Lab File: BN023290.D
 Acq: 20 Dec 2022 09:32

Tgt Ion:178 Resp: 2175972

Ion	Ratio	Lower	Upper
178	100		
179	15.6	11.8	17.6
176	18.3	14.7	22.1





#75

1,2,3,4-Tetrachlorobenzene

Concen: 22.200 ng/uL

RT: 13.445 min Scan# 11

Delta R.T. 0.000 min

Lab File: BN023290.D

Acq: 20 Dec 2022 09:32

Instrument :

BNA_N

ClientSampleId :

SSTD020376





Tgt Ion:216 Resp: 539132

Ion Ratio Lower Upper

216 100

214 79.5 61.9 92.9

218 48.0 39.4 59.2



#76

Pentachlorobenzene

Concen: 22.160 ng/uL

RT: 14.963 min Scan# 2019

Delta R.T. 0.000 min

Lab File: BN023290.D

Acq: 20 Dec 2022 09:32




Tgt Ion:250 Resp: 545320


Ion Ratio Lower Upper

250 100

248 63.4 49.4 74.0

252 65.8 52.9 79.3









#77
 Carbazole
 Concen: 20.801 ng/ul
 RT: 17.798 min Scan# 2501
 Delta R.T. 0.000 min
 Lab File: BN023290.D
 Acq: 20 Dec 2022 09:32




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020376


Tgt Ion:167 Resp: 1974823
 Ion Ratio Lower Upper
 167 100
 166 21.9 17.3 25.9
 139 13.2 10.6 16.0



#78
 Di-n-butylphthalate
 Concen: 21.157 ng/ul
 RT: 18.357 min Scan# 2596
 Delta R.T. 0.000 min
 Lab File: BN023290.D
 Acq: 20 Dec 2022 09:32

Tgt Ion:149 Resp: 2385027
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.4 11.2
 104 5.0 4.1 6.1









#80
Fluoranthene
Concen: 23.210 ng/ul
RT: 19.451 min Scan# 2782
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion:202 Resp: 2551965
Ion Ratio Lower Upper
202 100
101 12.1 10.4 15.6
100 9.2 7.6 11.4



#82
Pyrene
Concen: 22.917 ng/ul
RT: 19.815 min Scan# 2844
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32

Tgt Ion:202 Resp: 2646936
Ion Ratio Lower Upper
202 100
101 13.4 12.0 18.0
100 11.2 9.5 14.3









#83
Butylbenzylphthalate
Concen: 23.277 ng/ul
RT: 20.709 min Scan# 2996
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion:149 Resp: 1112598
Ion Ratio Lower Upper
149 100
91 72.7 57.9 86.9
206 20.6 15.6 23.4



#84
3,3'-Dichlorobenzidine
Concen: 20.722 ng/ul
RT: 21.498 min Scan# 3130
Delta R.T. 0.000 min
Lab File: BN023290.D
Acq: 20 Dec 2022 09:32

Tgt Ion:252 Resp: 791406
Ion Ratio Lower Upper
252 100
254 63.9 50.3 75.5
126 11.1 9.4 14.0





#85

Benzo(a)anthracene

Concen: 21.387 ng/ul

RT: 21.562 min Scan# 3141

Delta R.T. 0.000 min

Lab File: BN023290.D

Acq: 20 Dec 2022 09:32

Instrument :

BNA_N

ClientSampleId :

SSTD020376





Tgt Ion:228 Resp: 2490295

Ion Ratio Lower Upper

228 100

229 18.9 15.5 23.3

226 26.6 21.4 32.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 23.598 ng/ul

RT: 21.486 min Scan# 3128

Delta R.T. 0.000 min

Lab File: BN023290.D

Acq: 20 Dec 2022 09:32




Tgt Ion:149 Resp: 1594855


Ion Ratio Lower Upper

149 100

167 27.7 21.9 32.9

279 4.4 3.4 5.0





#87

Chrysene

Concen: 21.007 ng/ul

RT: 21.621 min Scan# 3151

Delta R.T. 0.000 min

Lab File: BN023290.D

Acq: 20 Dec 2022 09:32

Instrument :

BNA_N

ClientSampleId :

SSTD020376





Tgt Ion:228 Resp: 2308353

Ion Ratio Lower Upper

228 100

226 29.6 24.6 36.8

229 19.6 16.5 24.7



#89

Di-n-octyl phthalate

Concen: 24.784 ng/ul




RT: 22.427 min Scan# 3288


Delta R.T. 0.000 min

Lab File: BN023290.D

Acq: 20 Dec 2022 09:32

Tgt Ion:149 Resp: 2591154





#90

Benzo(b)fluoranthene

Concen: 21.166 ng/ul

RT: 23.274 min Scan# 3432

Delta R.T. 0.000 min

Lab File: BN023290.D

Acq: 20 Dec 2022 09:32

Instrument :

BNA_N

ClientSampleId :

SSTD020376





Tgt Ion:252 Resp: 2304730

Ion Ratio Lower Upper

252 100

253 21.4 16.8 25.2

125 11.1 8.8 13.2



#91

Benzo(k)fluoranthene

Concen: 21.673 ng/ul

RT: 23.327 min Scan# 3441

Delta R.T. 0.000 min

Lab File: BN023290.D

Acq: 20 Dec 2022 09:32




Tgt Ion:252 Resp: 2245906


Ion Ratio Lower Upper

252 100

253 21.7 17.9 26.9

125 10.7 9.0 13.6





#93

Benzo(a)pyrene

Concen: 21.167 ng/ul

RT: 23.921 min Scan# 3542

Delta R.T. 0.000 min

Lab File: BN023290.D

Acq: 20 Dec 2022 09:32

Instrument :

BNA_N

ClientSampleId :

SSTD020376





Tgt Ion:252 Resp: 1980780

Ion Ratio Lower Upper

252 100

253 21.7 17.6 26.4

125 11.2 9.4 14.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.072 ng/ul

RT: 26.591 min Scan# 3996

Delta R.T. 0.000 min

Lab File: BN023290.D

Acq: 20 Dec 2022 09:32




Tgt Ion:276 Resp: 2260955


Ion Ratio Lower Upper

276 100

138 25.2 20.6 30.8

277 22.9 18.2 27.2









#95
 Dibenzo(a,h)anthracene
 Concen: 19.713 ng/ul
 RT: 26.603 min Scan# 3998
 Delta R.T. 0.000 min
 Lab File: BN023290.D
 Acq: 20 Dec 2022 09:32

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020376

Tgt Ion:278 Resp: 1905454
 Ion Ratio Lower Upper
 278 100
 139 16.8 13.3 19.9
 279 20.3 18.6 27.8



#96
 Benzo(g,h,i)perylene
 Concen: 18.983 ng/ul
 RT: 27.380 min Scan# 4130
 Delta R.T. 0.000 min
 Lab File: BN023290.D
 Acq: 20 Dec 2022 09:32

Tgt Ion:276 Resp: 1780556
 Ion Ratio Lower Upper
 276 100
 138 22.1 19.2 28.8
 277 23.0 19.3 28.9

