

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122022\
 Data File : BN023303.D
 Acq On : 20 Dec 2022 17:56
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020377

Quant Time: Dec 21 04:15:01 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN121522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 21 04:11:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.999	152	347974	20.000	ng/u1	0.00
20) Naphthalene-d8	10.816	136	1632698	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.651	164	1064110	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.392	188	2364938	20.000	ng/u1	0.00
79) Chrysene-d12	21.586	240	2074419	20.000	ng/u1	0.00
88) Perylene-d12	24.033	264	2100627	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.340	96	57918	6.886	ng/uL	0.00
4) Pyridine-d5	3.769	84	456466	18.386	ng/u1	0.00
7) Phenol-d5	7.157	99	628174	19.410	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.322	67	367757	19.527	ng/u1	0.00
11) 2-Chlorophenol-d4	7.522	132	500315	20.516	ng/u1	0.00
15) 4-Methylphenol-d8	8.716	113	514964	19.818	ng/u1	0.00
21) Nitrobenzene-d5	9.169	128	261681	20.744	ng/u1	0.00
24) 2-Nitrophenol-d4	9.898	143	291933	21.972	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.434	165	520260	20.742	ng/u1	0.00
31) 4-Chloroaniline-d4	10.957	131	777131	19.256	ng/u1	0.00
46) Dimethylphthalate-d6	14.057	166	1562429	19.593	ng/u1	0.00
49) Acenaphthylene-d8	14.345	160	1825967	20.418	ng/u1	0.00
54) 4-Nitrophenol-d4	14.845	143	332081	18.645	ng/u1	0.00
60) Fluorene-d10	15.639	176	1354771	19.944	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.757	200	280936	20.180	ng/u1	0.00
73) Anthracene-d10	17.492	188	2120721	20.009	ng/u1	0.00
81) Pyrene-d10	19.786	212	2505758	22.206	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.874	264	2156618	20.946	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.381	88	61539	7.060	ng/uL	95
5) Pyridine	3.793	79	459819	18.125	ng/u1	97
6) Benzaldehyde	7.128	77	261221	22.533	ng/u1	98
8) Phenol	7.181	94	632070	19.335	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.416	93	492158	19.383	ng/u1	99
12) 2-Chlorophenol	7.557	128	507393	20.258	ng/u1	99
13) 2-Methylphenol	8.446	108	487589	19.603	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.528	45	627525	19.514	ng/u1	99
16) Acetophenone	8.828	105	785872	19.989	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.816	70	400605	20.348	ng/u1	99
18) 4-Methylphenol	8.781	108	546460	19.797	ng/u1	97
19) Hexachloroethane	9.087	117	198248	20.551	ng/u1	94
22) Nitrobenzene	9.210	77	596570	19.723	ng/u1	99
23) Isophorone	9.740	82	1152880	20.207	ng/u1	100
25) 2-Nitrophenol	9.928	139	306998	21.186	ng/u1	98
26) 2,4-Dimethylphenol	9.987	107	586302	19.988	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.228	93	695989	19.070	ng/u1	98
29) 2,4-Dichlorophenol	10.463	162	508727	20.497	ng/u1	98
30) Naphthalene	10.869	128	1695836	19.442	ng/u1	100
32) 4-Chloroaniline	10.981	127	775048	19.249	ng/u1	98
33) Hexachlorobutadiene	11.151	225	301250	21.291	ng/u1	97
34) Caprolactam	11.763	113	130552	13.528	ng/u1	95
35) 4-Chloro-3-methylphenol	12.104	107	564031	20.063	ng/u1	97
36) 2-Methylnaphthalene	12.475	142	1171659	19.930	ng/u1	98


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122022\
 Data File : BN023303.D
 Acq On : 20 Dec 2022 17:56
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020377

Quant Time: Dec 21 04:15:01 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN121522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 21 04:11:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.698	142	1202151	19.844	ng/ul	96
39) 1,2,4,5-Tetrachloroben...	12.845	216	613253	20.375	ng/ul	97
40) Hexachlorocyclopentadiene	12.822	237	332468	18.364	ng/ul	96
41) 2,4,6-Trichlorophenol	13.081	196	426322	20.593	ng/ul	97
42) 2,4,5-Trichlorophenol	13.151	196	472179	20.608	ng/ul	96
43) 1,1'-Biphenyl	13.487	154	1569600	19.521	ng/ul	98
44) 2-Chloronaphthalene	13.528	162	1227569	19.605	ng/ul	98
45) 2-Nitroaniline	13.734	65	373700	20.595	ng/ul	97
47) Dimethylphthalate	14.104	163	1562971	19.424	ng/ul	100
48) 2,6-Dinitrotoluene	14.228	165	339405	21.786	ng/ul	95
50) Acenaphthylene	14.369	152	1943416	19.762	ng/ul	99
51) 3-Nitroaniline	14.557	138	321695	20.280	ng/ul	97
52) Acenaphthene	14.716	153	1331229	19.471	ng/ul	99
53) 2,4-Dinitrophenol	14.757	184	194423	19.585	ng/ul	93
55) 4-Nitrophenol	14.863	109	243826	18.510	ng/ul	92
56) Dibenzofuran	15.045	168	1858128	19.520	ng/ul	99
57) 2,4-Dinitrotoluene	15.010	165	479442	20.848	ng/ul	94
58) 2,3,4,6-Tetrachlorophenol	15.269	232	408589	20.597	ng/ul	98
59) Diethylphthalate	15.469	149	1555482	19.432	ng/ul	99
61) Fluorene	15.698	166	1491758	19.352	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.686	204	728034	19.706	ng/ul	99
63) 4-Nitroaniline	15.716	138	317788	20.571	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.775	198	284751	20.951	ng/ul	97
67) N-Nitrosodiphenylamine	15.904	169	1313499	20.068	ng/ul	96
68) 4-Bromophenyl-phenylether	16.580	248	463974	20.333	ng/ul	97
69) Hexachlorobenzene	16.698	284	553432	20.744	ng/ul	98
70) Atrazine	16.857	200	490000	20.487	ng/ul	97
71) Pentachlorophenol	17.045	266	361588	20.525	ng/ul	98
72) Phenanthrene	17.439	178	2465218	19.700	ng/ul	99
74) Anthracene	17.528	178	2491221	19.952	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.451	216	607530	20.657	ng/ul	99
76) Pentachlorobenzene	14.963	250	611195	20.508	ng/ul	94
77) Carbazole	17.798	167	2373163	20.640	ng/ul	99
78) Di-n-butylphthalate	18.363	149	2842307	20.820	ng/ul	100
80) Fluoranthene	19.451	202	2928963	22.318	ng/ul	98
82) Pyrene	19.816	202	3043390	22.075	ng/ul	97
83) Butylbenzylphthalate	20.710	149	1366877	23.959	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.498	252	945488	20.741	ng/ul	97
85) Benzo(a)anthracene	21.563	228	2901195	20.875	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.486	149	1949631	24.169	ng/ul	100
87) Chrysene	21.621	228	2730913	20.821	ng/ul	97
89) Di-n-octyl phthalate	22.427	149	3299498	25.937	ng/ul	100
90) Benzo(b)fluoranthene	23.280	252	2712384	20.472	ng/ul	99
91) Benzo(k)fluoranthene	23.333	252	2621294	20.790	ng/ul	99
93) Benzo(a)pyrene	23.927	252	2321752	20.390	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.598	276	2643724	18.328	ng/ul	97
95) Dibenzo(a,h)anthracene	26.615	278	2185766	18.584	ng/ul	98
96) Benzo(g,h,i)perylene	27.392	276	2064177	18.086	ng/ul	96





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 7.060 ng/uL
RT: 3.381 min Scan# 50
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56




Instrument :
BNA_N
ClientSampleId :
SSTD020377


Tgt Ion: 88 Resp: 61539
Ion Ratio Lower Upper
88 100
43 26.6 23.0 34.6
58 65.9 55.8 83.6



#5
Pyridine
Concen: 18.125 ng/uL
RT: 3.793 min Scan# 120
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56

Tgt Ion: 79 Resp: 459819
Ion Ratio Lower Upper
79 100
52 67.5 56.7 85.1
51 32.6 26.0 39.0









#6
Benzaldehyde
Concen: 22.533 ng/ul
RT: 7.128 min Scan# 687
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56




Instrument :
BNA_N
ClientSampleId :
SSTD020377


Tgt Ion: 77 Resp: 261221
Ion Ratio Lower Upper
77 100
105 98.3 78.0 117.0
106 93.8 73.2 109.8



#8
Phenol
Concen: 19.335 ng/ul
RT: 7.181 min Scan# 696
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56

Tgt Ion: 94 Resp: 632070
Ion Ratio Lower Upper
94 100
65 28.0 21.8 32.6
66 36.8 28.4 42.6









#10
Bis(2-Chloroethyl)ether
Concen: 19.383 ng/ul
RT: 7.416 min Scan# 736
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56




Instrument :
BNA_N
ClientSampleId :
SSTD020377


Tgt Ion: 93 Resp: 492158
Ion Ratio Lower Upper
93 100
63 74.7 58.9 88.3
95 32.6 26.2 39.4



#12
2-Chlorophenol
Concen: 20.258 ng/ul
RT: 7.557 min Scan# 760
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56

Tgt Ion:128 Resp: 507393
Ion Ratio Lower Upper
128 100
64 43.2 34.0 51.0
130 31.3 25.0 37.4





#13

2-Methylphenol

Concen: 19.603 ng/ul

RT: 8.446 min Scan# 911

Delta R.T. 0.000 min

Lab File: BN023303.D

Acq: 20 Dec 2022 17:56

Instrument :

BNA_N

ClientSampleId :





SSTD020377

Tgt Ion:108 Resp: 487589

Ion Ratio Lower Upper

108 100

107 91.1 71.2 106.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 19.514 ng/ul

RT: 8.528 min Scan# 925

Delta R.T. 0.000 min

Lab File: BN023303.D

Acq: 20 Dec 2022 17:56




Tgt Ion: 45 Resp: 627525


Ion Ratio Lower Upper

45 100

77 15.4 12.6 19.0

79 11.7 8.8 13.2









#16
Acetophenone
Concen: 19.989 ng/ul
RT: 8.828 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56




Instrument :
BNA_N
ClientSampleId :
SSTD020377


Tgt Ion:105 Resp: 785872
Ion Ratio Lower Upper
105 100
77 77.7 61.4 92.0
51 25.3 20.0 30.0



#17
N-Nitroso-di-n-propylamine
Concen: 20.348 ng/ul
RT: 8.816 min Scan# 974
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56

Tgt Ion: 70 Resp: 400605
Ion Ratio Lower Upper
70 100
42 57.4 45.3 67.9
101 9.9 8.1 12.1
130 22.2 17.9 26.9









#18
4-Methylphenol
Concen: 19.797 ng/ul
RT: 8.781 min Scan# 968
Delta R.T. 0.006 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56




Instrument :
BNA_N
ClientSampleId :
SSTD020377


Tgt Ion:108 Resp: 546460
Ion Ratio Lower Upper
108 100
107 106.7 87.6 131.4



#19
Hexachloroethane
Concen: 20.551 ng/ul
RT: 9.087 min Scan# 1020
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56

Tgt Ion:117 Resp: 198248
Ion Ratio Lower Upper
117 100
201 99.1 74.6 111.8
199 62.0 46.3 69.5









#22
Nitrobenzene
Concen: 19.723 ng/ul
RT: 9.210 min Scan# 1041
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56




Instrument :
BNA_N
ClientSampleId :
SSTD020377


Tgt Ion: 77 Resp: 596570
Ion Ratio Lower Upper
77 100
123 46.9 37.1 55.7
65 13.1 10.1 15.1



#23
Isophorone
Concen: 20.207 ng/ul
RT: 9.740 min Scan# 1131
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56

Tgt Ion: 82 Resp: 1152880
Ion Ratio Lower Upper
82 100
95 7.0 5.4 8.2
138 17.9 14.5 21.7









#25
2-Nitrophenol
Concen: 21.186 ng/ul
RT: 9.928 min Scan# 1163
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56




Instrument :
BNA_N
ClientSampleId :
SSTD020377


Tgt Ion:139 Resp: 306998
Ion Ratio Lower Upper
139 100
65 46.2 37.7 56.5
109 27.1 22.6 34.0



#26
2,4-Dimethylphenol
Concen: 19.988 ng/ul
RT: 9.987 min Scan# 1173
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56

Tgt Ion:107 Resp: 586302
Ion Ratio Lower Upper
107 100
121 52.7 41.3 61.9
122 89.1 72.1 108.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.070 ng/ul
 RT: 10.228 min Scan# 1214
 Delta R.T. 0.000 min
 Lab File: BN023303.D
 Acq: 20 Dec 2022 17:56




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020377


Tgt Ion: 93 Resp: 695989
 Ion Ratio Lower Upper
 93 100
 95 33.3 25.8 38.6
 123 11.7 9.0 13.4



#29
 2,4-Dichlorophenol
 Concen: 20.497 ng/ul
 RT: 10.463 min Scan# 1254
 Delta R.T. 0.000 min
 Lab File: BN023303.D
 Acq: 20 Dec 2022 17:56

Tgt Ion: 162 Resp: 508727
 Ion Ratio Lower Upper
 162 100
 164 67.2 52.1 78.1
 98 36.2 29.3 43.9





#30

Naphthalene

Concen: 19.442 ng/ul

RT: 10.869 min Scan# 1323

Delta R.T. 0.000 min

Lab File: BN023303.D

Acq: 20 Dec 2022 17:56

Instrument :

BNA_N

ClientSampleId :

SSTD020377





Tgt Ion:128 Resp: 1695836

Ion Ratio Lower Upper

128 100

129 11.1 8.7 13.1

127 12.9 10.3 15.5



#32

4-Chloroaniline

Concen: 19.249 ng/ul

RT: 10.981 min Scan# 1342

Delta R.T. 0.000 min

Lab File: BN023303.D




Acq: 20 Dec 2022 17:56


Tgt Ion:127 Resp: 775048

Ion Ratio Lower Upper

127 100

129 31.9 26.2 39.4





#33

Hexachlorobutadiene

Concen: 21.291 ng/ul

RT: 11.151 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN023303.D

Acq: 20 Dec 2022 17:56

Instrument :

BNA_N

ClientSampleId :

SSTD020377





Tgt Ion:225 Resp: 301250

Ion Ratio Lower Upper

225 100

223 63.4 48.9 73.3

227 60.9 51.0 76.4



#34

Caprolactam

Concen: 13.528 ng/ul

RT: 11.763 min Scan# 1475

Delta R.T. 0.006 min

Lab File: BN023303.D

Acq: 20 Dec 2022 17:56




Tgt Ion:113 Resp: 130552


Ion Ratio Lower Upper

113 100

55 146.9 123.7 185.5

56 109.6 91.7 137.5









#35
4-Chloro-3-methylphenol
Concen: 20.063 ng/ul
RT: 12.104 min Scan# 1533
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56




Instrument :
BNA_N
ClientSampleId :
SSTD020377


Tgt Ion:107 Resp: 564031
Ion Ratio Lower Upper
107 100
144 27.1 21.8 32.6
142 82.3 62.6 94.0



#36
2-Methylnaphthalene
Concen: 19.930 ng/ul
RT: 12.475 min Scan# 1596
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56

Tgt Ion:142 Resp: 1171659
Ion Ratio Lower Upper
142 100
141 89.5 69.9 104.9









#37
1-Methylnaphthalene
Concen: 19.844 ng/ul
RT: 12.698 min Scan# 1633
Delta R.T. 0.006 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56




Instrument :
BNA_N
ClientSampleId :
SSTD020377


Tgt Ion:142 Resp: 1202151
Ion Ratio Lower Upper
142 100
141 92.6 70.9 106.3
116 3.7 3.0 4.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.375 ng/ul
RT: 12.845 min Scan# 1659
Delta R.T. 0.006 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56

Tgt Ion:216 Resp: 613253
Ion Ratio Lower Upper
216 100
214 78.8 61.1 91.7
179 20.0 16.6 24.8
108 17.4 15.4 23.2





#40

Hexachlorocyclopentadiene

Concen: 18.364 ng/ul

RT: 12.822 min Scan# 1655

Delta R.T. 0.000 min

Lab File: BN023303.D

Acq: 20 Dec 2022 17:56

Instrument :

BNA_N

ClientSampleId :

SSTD020377





Tgt Ion:237 Resp: 332468

Ion Ratio Lower Upper

237 100

235 64.0 48.1 72.1

272 14.0 10.8 16.2



#41

2,4,6-Trichlorophenol

Concen: 20.593 ng/ul

RT: 13.081 min Scan# 1699

Delta R.T. 0.000 min

Lab File: BN023303.D

Acq: 20 Dec 2022 17:56




Tgt Ion:196 Resp: 426322


Ion Ratio Lower Upper

196 100

198 91.9 76.9 115.3

200 29.9 23.8 35.6









#42
2,4,5-Trichlorophenol
Concen: 20.608 ng/ul
RT: 13.151 min Scan# 1711
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56




Instrument :
BNA_N
ClientSampleId :
SSTD020377


Tgt Ion:196 Resp: 472179
Ion Ratio Lower Upper
196 100
198 97.9 74.8 112.2
200 31.0 24.2 36.2



#43
1,1'-Biphenyl
Concen: 19.521 ng/ul
RT: 13.487 min Scan# 1768
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56

Tgt Ion:154 Resp: 1569600
Ion Ratio Lower Upper
154 100
153 39.3 32.0 48.0
76 12.4 11.0 16.4









#44
2-Chloronaphthalene
Concen: 19.605 ng/ul
RT: 13.528 min Scan# 1775
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56




Instrument :
BNA_N
ClientSampleId :
SSTD020377


Tgt Ion:162 Resp: 1227569
Ion Ratio Lower Upper
162 100
164 33.3 25.4 38.0
127 39.1 31.8 47.8



#45
2-Nitroaniline
Concen: 20.595 ng/ul
RT: 13.734 min Scan# 1810
Delta R.T. 0.006 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56

Tgt Ion: 65 Resp: 373700
Ion Ratio Lower Upper
65 100
92 70.8 57.4 86.2
138 116.6 89.0 133.6









#47
Dimethylphthalate
Concen: 19.424 ng/ul
RT: 14.104 min Scan# 1873
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56




Instrument :
BNA_N
ClientSampleId :
SSTD020377


Tgt Ion:163 Resp: 1562971
Ion Ratio Lower Upper
163 100
194 4.2 3.3 4.9
164 10.4 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 21.786 ng/ul
RT: 14.228 min Scan# 1894
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56

Tgt Ion:165 Resp: 339405
Ion Ratio Lower Upper
165 100
89 55.2 47.7 71.5
121 21.6 17.8 26.8









#50
Acenaphthylene
Concen: 19.762 ng/ul
RT: 14.369 min Scan# 1918
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56




Instrument :
BNA_N
ClientSampleId :
SSTD020377


Tgt Ion:152 Resp: 1943416
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.4
153 13.5 11.4 17.0



#51
3-Nitroaniline
Concen: 20.280 ng/ul
RT: 14.557 min Scan# 1950
Delta R.T. 0.006 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56

Tgt Ion:138 Resp: 321695
Ion Ratio Lower Upper
138 100
108 10.4 8.3 12.5
92 111.5 92.5 138.7









#52
 Acenaphthene
 Concen: 19.471 ng/ul
 RT: 14.716 min Scan# 1977
 Delta R.T. 0.006 min
 Lab File: BN023303.D
 Acq: 20 Dec 2022 17:56




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020377


Tgt Ion:153 Resp: 1331229
 Ion Ratio Lower Upper
 153 100
 152 47.3 38.9 58.3
 154 87.5 70.5 105.7



#53
 2,4-Dinitrophenol
 Concen: 19.585 ng/ul
 RT: 14.757 min Scan# 1984
 Delta R.T. 0.000 min
 Lab File: BN023303.D
 Acq: 20 Dec 2022 17:56

Tgt Ion:184 Resp: 194423
 Ion Ratio Lower Upper
 184 100
 63 61.6 57.1 85.7
 154 60.4 47.7 71.5









#55
4-Nitrophenol
Concen: 18.510 ng/ul
RT: 14.863 min Scan# 2001
Delta R.T. 0.006 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56




Instrument :
BNA_N
ClientSampleId :
SSTD020377


Tgt Ion:109 Resp: 243826
Ion Ratio Lower Upper
109 100
139 138.6 115.2 172.8
65 122.9 109.0 163.4



#56
Dibenzofuran
Concen: 19.520 ng/ul
RT: 15.045 min Scan# 2033
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56

Tgt Ion:168 Resp: 1858128
Ion Ratio Lower Upper
168 100
139 37.0 29.2 43.8









#57
2,4-Dinitrotoluene
Concen: 20.848 ng/ul
RT: 15.010 min Scan# 2027
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56




Instrument :
BNA_N
ClientSampleId :
SSTD020377


Tgt Ion:165 Resp: 479442
Ion Ratio Lower Upper
165 100
63 41.3 36.6 54.8
182 2.7 2.5 3.7



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.597 ng/ul
RT: 15.269 min Scan# 2071
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56

Tgt Ion:232 Resp: 408589
Ion Ratio Lower Upper
232 100
131 47.7 37.9 56.9
130 2.5 1.9 2.9
166 31.6 23.8 35.8









#59
Diethylphthalate
Concen: 19.432 ng/ul
RT: 15.469 min Scan# 2105
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56




Instrument :
BNA_N
ClientSampleId :
SSTD020377


Tgt Ion:149 Resp: 1555482
Ion Ratio Lower Upper
149 100
177 22.1 17.0 25.4
150 12.5 10.0 15.0



#61
Fluorene
Concen: 19.352 ng/ul
RT: 15.698 min Scan# 2144
Delta R.T. 0.006 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56

Tgt Ion:166 Resp: 1491758
Ion Ratio Lower Upper
166 100
165 97.4 77.8 116.8
167 13.8 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 19.706 ng/ul
RT: 15.686 min Scan# 2142
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56




Instrument :
BNA_N
ClientSampleId :
SSTD020377


Tgt Ion:204 Resp: 728034
Ion Ratio Lower Upper
204 100
206 32.8 26.1 39.1
141 60.6 48.2 72.2



#63
4-Nitroaniline
Concen: 20.571 ng/ul
RT: 15.716 min Scan# 2147
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56

Tgt Ion:138 Resp: 317788
Ion Ratio Lower Upper
138 100
92 49.2 39.4 59.2
108 74.0 59.5 89.3









#66
4,6-Dinitro-2-methylphenol
Concen: 20.951 ng/ul
RT: 15.775 min Scan# 2156
Delta R.T. 0.006 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56




Instrument :
BNA_N
ClientSampleId :
SSTD020377


Tgt Ion:198 Resp: 284751
Ion Ratio Lower Upper
198 100
182 4.0 3.4 5.0
77 22.9 19.7 29.5



#67
N-Nitrosodiphenylamine
Concen: 20.068 ng/ul
RT: 15.904 min Scan# 2179
Delta R.T. 0.006 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56

Tgt Ion:169 Resp: 1313499
Ion Ratio Lower Upper
169 100
168 64.1 54.7 82.1
167 35.5 29.2 43.8









#68
 4-Bromophenyl-phenylether
 Concen: 20.333 ng/ul
 RT: 16.580 min Scan# 2114
 Delta R.T. 0.000 min
 Lab File: BN023303.D
 Acq: 20 Dec 2022 17:56




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020377


Tgt Ion:	248	Resp:	463974
Ion Ratio	Lower	Upper	
248	100		
250	103.5	81.0	121.4
141	79.0	60.8	91.2



#69
 Hexachlorobenzene
 Concen: 20.744 ng/ul
 RT: 16.698 min Scan# 2314
 Delta R.T. 0.000 min
 Lab File: BN023303.D
 Acq: 20 Dec 2022 17:56

Tgt Ion:	284	Resp:	553432
Ion Ratio	Lower	Upper	
284	100		
142	35.9	29.0	43.4
249	31.3	23.7	35.5









#70
Atrazine
Concen: 20.487 ng/ul
RT: 16.857 min Scan# 21
Delta R.T. 0.006 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56




Instrument :
BNA_N
ClientSampleId :
SSTD020377


Tgt Ion:200 Resp: 490000
Ion Ratio Lower Upper
200 100
173 28.4 21.3 31.9
215 51.3 39.8 59.8



#71
Pentachlorophenol
Concen: 20.525 ng/ul
RT: 17.045 min Scan# 2373
Delta R.T. 0.006 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56

Tgt Ion:266 Resp: 361588
Ion Ratio Lower Upper
266 100
264 60.4 50.5 75.7
268 65.6 52.6 78.8





#72

Phenanthrene

Concen: 19.700 ng/ul

RT: 17.439 min Scan# 2440

Delta R.T. 0.000 min

Lab File: BN023303.D

Acq: 20 Dec 2022 17:56

Instrument :

BNA_N

ClientSampleId :

SSTD020377





Tgt Ion:178 Resp: 2465218

Ion Ratio Lower Upper

178 100

179 15.8 12.3 18.5

176 19.3 15.4 23.0



#74

Anthracene

Concen: 19.952 ng/ul

RT: 17.528 min Scan# 2455

Delta R.T. 0.000 min

Lab File: BN023303.D

Acq: 20 Dec 2022 17:56




Tgt Ion:178 Resp: 2491221


Ion Ratio Lower Upper

178 100

179 15.5 11.8 17.6

176 18.9 14.7 22.1





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.657 ng/uL

RT: 13.451 min Scan# 1

Delta R.T. 0.006 min

Lab File: BN023303.D

Acq: 20 Dec 2022 17:56

Instrument :

BNA_N

ClientSampleId :

SSTD020377





Tgt Ion:216 Resp: 607530

Ion Ratio Lower Upper

216 100

214 78.6 61.9 92.9

218 48.2 39.4 59.2



#76

Pentachlorobenzene

Concen: 20.508 ng/uL

RT: 14.963 min Scan# 2019

Delta R.T. 0.000 min

Lab File: BN023303.D

Acq: 20 Dec 2022 17:56




Tgt Ion:250 Resp: 611195


Ion Ratio Lower Upper

250 100

248 62.3 49.4 74.0

252 58.0 52.9 79.3









#77
 Carbazole
 Concen: 20.640 ng/ul
 RT: 17.798 min Scan# 2501
 Delta R.T. 0.000 min
 Lab File: BN023303.D
 Acq: 20 Dec 2022 17:56




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020377


Tgt Ion:167 Resp: 2373163
 Ion Ratio Lower Upper
 167 100
 166 20.8 17.3 25.9
 139 13.0 10.6 16.0



#78
 Di-n-butylphthalate
 Concen: 20.820 ng/ul
 RT: 18.363 min Scan# 2597
 Delta R.T. 0.006 min
 Lab File: BN023303.D
 Acq: 20 Dec 2022 17:56

Tgt Ion:149 Resp: 2842307
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.4 11.2
 104 5.0 4.1 6.1









#80
Fluoranthene
Concen: 22.318 ng/ul
RT: 19.451 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56




Instrument :
BNA_N
ClientSampleId :
SSTD020377


Tgt Ion:202 Resp: 2928963
Ion Ratio Lower Upper
202 100
101 12.2 10.4 15.6
100 9.2 7.6 11.4



#82
Pyrene
Concen: 22.075 ng/ul
RT: 19.816 min Scan# 2844
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56

Tgt Ion:202 Resp: 3043390
Ion Ratio Lower Upper
202 100
101 13.6 12.0 18.0
100 11.0 9.5 14.3









#83
Butylbenzylphthalate
Concen: 23.959 ng/ul
RT: 20.710 min Scan# 2996
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56




Instrument :
BNA_N
ClientSampleId :
SSTD020377


Tgt Ion:149 Resp: 1366877
Ion Ratio Lower Upper
149 100
91 71.8 57.9 86.9
206 20.6 15.6 23.4



#84
3,3'-Dichlorobenzidine
Concen: 20.741 ng/ul
RT: 21.498 min Scan# 3130
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56

Tgt Ion:252 Resp: 945488
Ion Ratio Lower Upper
252 100
254 59.8 50.3 75.5
126 11.6 9.4 14.0









#85
Benzo(a)anthracene
Concen: 20.875 ng/ul
RT: 21.563 min Scan# 3141
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56




Instrument :
BNA_N
ClientSampleId :
SSTD020377


Tgt Ion:228 Resp: 2901195
Ion Ratio Lower Upper
228 100
229 20.0 15.5 23.3
226 26.5 21.4 32.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 24.169 ng/ul
RT: 21.486 min Scan# 3128
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56

Tgt Ion:149 Resp: 1949631
Ion Ratio Lower Upper
149 100
167 27.5 21.9 32.9
279 4.2 3.4 5.0









#87
Chrysene
Concen: 20.821 ng/ul
RT: 21.621 min Scan# 3151
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56




Instrument :
BNA_N
ClientSampleId :
SSTD020377


Tgt Ion:228 Resp: 2730913
Ion Ratio Lower Upper
228 100
226 28.7 24.6 36.8
229 19.1 16.5 24.7



#89
Di-n-octyl phthalate
Concen: 25.937 ng/ul
RT: 22.427 min Scan# 3288
Delta R.T. 0.000 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56

Tgt Ion:149 Resp: 3299498





#90

Benzo(b)fluoranthene

Concen: 20.472 ng/ul

RT: 23.280 min Scan# 3432

Delta R.T. 0.006 min

Lab File: BN023303.D

Acq: 20 Dec 2022 17:56

Instrument :

BNA_N

ClientSampleId :

SSTD020377





Tgt Ion:252 Resp: 2712384

Ion Ratio Lower Upper

252 100

253 21.6 16.8 25.2

125 10.6 8.8 13.2



#91

Benzo(k)fluoranthene

Concen: 20.790 ng/ul

RT: 23.333 min Scan# 3442

Delta R.T. 0.006 min

Lab File: BN023303.D

Acq: 20 Dec 2022 17:56




Tgt Ion:252 Resp: 2621294


Ion Ratio Lower Upper

252 100

253 22.3 17.9 26.9

125 10.6 9.0 13.6









#93
Benzo(a)pyrene
Concen: 20.390 ng/ul
RT: 23.927 min Scan# 31
Delta R.T. 0.006 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56




Instrument :
BNA_N
ClientSampleId :
SSTD020377


Tgt Ion:252 Resp: 2321752
Ion Ratio Lower Upper
252 100
253 21.7 17.6 26.4
125 12.1 9.4 14.2



#94
Indeno(1,2,3-cd)pyrene
Concen: 18.328 ng/ul
RT: 26.598 min Scan# 3997
Delta R.T. 0.006 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56

Tgt Ion:276 Resp: 2643724
Ion Ratio Lower Upper
276 100
138 26.2 20.6 30.8
277 25.0 18.2 27.2









#95
Dibenzo(a,h)anthracene
Concen: 18.584 ng/ul
RT: 26.615 min Scan# 4000
Delta R.T. 0.012 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56

Instrument :
BNA_N
ClientSampleId :
SSTD020377

Tgt Ion:278 Resp: 2185766
Ion Ratio Lower Upper
278 100
139 17.2 13.3 19.9
279 24.0 18.6 27.8



#96
Benzo(g,h,i)perylene
Concen: 18.086 ng/ul
RT: 27.392 min Scan# 4132
Delta R.T. 0.012 min
Lab File: BN023303.D
Acq: 20 Dec 2022 17:56

Tgt Ion:276 Resp: 2064177
Ion Ratio Lower Upper
276 100
138 22.1 19.2 28.8
277 21.9 19.3 28.9

