

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122022\
 Data File : BN023295.D
 Acq On : 20 Dec 2022 13:01
 Operator : CG/JU
 Sample : N6003-17
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DBXT1

Quant Time: Dec 21 04:13:53 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN121522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 21 04:11:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)

Internal Standards					
1) 1,4-Dichlorobenzene-d4	7.999	152	282249	20.000 ng/u1	0.00
20) Naphthalene-d8	10.816	136	1305746	20.000 ng/u1	0.00
38) Acenaphthene-d10	14.645	164	837374	20.000 ng/u1	0.00
64) Phenanthrene-d10	17.392	188	1843115	20.000 ng/u1	0.00
79) Chrysene-d12	21.580	240	1721858	20.000 ng/u1	0.00
88) Perylene-d12	24.033	264	1648775	20.000 ng/u1	0.00
System Monitoring Compounds					
3) 1,4-Dioxane-d8	3.340	96	28902	4.237 ng/uL	0.00
4) Pyridine-d5	3.775	84	252845	12.556 ng/u1	0.00
7) Phenol-d5	7.152	99	655433	24.968 ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.322	67	397731	26.036 ng/u1	0.00
11) 2-Chlorophenol-d4	7.522	132	536783	27.137 ng/u1	0.00
15) 4-Methylphenol-d8	8.710	113	554154	26.292 ng/u1	0.00
21) Nitrobenzene-d5	9.169	128	278921	27.647 ng/u1	0.00
24) 2-Nitrophenol-d4	9.893	143	308563	29.039 ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.434	165	558324	27.834 ng/u1	0.00
31) 4-Chloroaniline-d4	10.952	131	748881	23.202 ng/u1	0.00
46) Dimethylphthalate-d6	14.057	166	1875850	29.893 ng/u1	0.00
49) Acenaphthylene-d8	14.340	160	2156096	30.638 ng/u1	0.00
54) 4-Nitrophenol-d4	14.840	143	327247	23.349 ng/u1	0.00
60) Fluorene-d10	15.640	176	1647968	30.829 ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.751	200	271417	25.016 ng/u1	0.00
73) Anthracene-d10	17.492	188	2554150	30.921 ng/u1	0.00
81) Pyrene-d10	19.786	212	3052540	32.590 ng/u1	0.00
92) Benzo(a)pyrene-d12	23.874	264	2570585	31.808 ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122022\
Data File : BN023295.D
Acq On : 20 Dec 2022 13:01
Operator : CG/JU
Sample : N6003-17
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DBXT1

Quant Time: Dec 21 04:13:53 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN121522.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 21 04:11:00 2022
Response via : Initial Calibration

