


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122022\
 Data File : BN023289.D
 Acq On : 20 Dec 2022 08:19
 Operator : CG/JU
 Sample : DFTPP264
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP264

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN121522.M
 Title : SVOA CALIBRATION
 Last Update : Wed Dec 21 04:11:00 2022



AutoFind: Scans 2505, 2506, 2507; Background Corrected with Scan 2499

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.3	226493	PASS
68	69	0.00	2	1.8	4360	PASS
69	198	0.00	100	38.7	248299	PASS
70	69	0.00	2	0.6	1457	PASS
127	198	10	80	48.0	307499	PASS
197	198	0.00	2	0.4	2552	PASS
198	198	100	100	100.0	641045	PASS
199	198	5	9	6.9	43971	PASS
275	198	10	60	25.7	164541	PASS
365	198	1	100	3.3	20862	PASS
441	198	0.01	100	14.2	90832	PASS
442	198	50	100	93.1	596800	PASS
443	442	15	24	20.3	121424	PASS

m/z	Abundance
36.00	213.0
38.90	171.0
40.00	995.0
41.00	203.0
41.90	155.0
43.00	223.0
44.00	1216.0
44.90	241.0
48.90	258.0
50.90	174.0
57.00	166.0
59.00	168.0
73.00	1257.0
75.00	244.0
76.90	463.0
83.80	237.0
85.80	186.0
91.00	154.0
96.00	748.0
97.10	171.0
102.80	240.0
114.90	176.0
118.90	193.0
133.00	582.0
135.00	216.0
146.90	411.0
162.80	235.0
177.00	296.0
191.00	856.0
191.90	209.0
192.90	503.0
207.00	7036.0
208.00	1507.0
209.00	987.0
221.20	187.0
281.00	983.0
282.10	302.0
282.90	273.0
355.10	168.0

Instrument :
BNA_N
ClientSampleId :
DFTPP264

m/z	Abundance
35.90	511.0
37.00	1035.0
38.00	2714.0
39.00	15524.0
40.00	1590.0
41.00	666.0
43.00	179.0
43.90	1128.0
44.90	584.0
46.90	190.0
49.00	1619.0
50.00	53224.0
51.00	201472.0
52.00	10710.0
53.00	401.0
54.90	1245.0
56.00	5812.0
57.00	13468.0
57.90	694.0
59.00	350.0
61.00	2916.0
62.00	3214.0
63.00	8074.0
64.00	1035.0
65.00	4194.0
66.00	378.0
66.80	290.0
68.00	3750.0
69.00	211776.0
70.00	1358.0
73.00	2705.0
74.00	23304.0
75.00	34888.0
76.00	12216.0
77.00	253952.0
78.00	18608.0
79.00	16308.0
80.00	12827.0
81.00	17528.0
82.00	4466.0
83.00	3965.0
83.90	610.0
85.00	3212.0
86.00	5268.0
87.00	2575.0
87.90	786.0
89.00	534.0
91.00	3818.0
91.90	4361.0
93.00	28616.0
94.00	1915.0
94.90	505.0
95.90	1564.0
97.10	553.0
98.00	23640.0
99.00	19184.0
99.90	1862.0
101.00	9778.0
101.90	574.0
102.90	3128.0
103.90	6042.0
105.00	5608.0
106.00	1914.0
107.00	73992.0
108.00	10834.0
109.00	1825.0
110.00	129752.0
111.00	21296.0
112.00	2609.0
112.90	929.0
114.10	260.0
115.00	485.0
115.90	3765.0
117.00	70456.0
118.00	5117.0
118.90	821.0
120.00	1037.0
120.90	333.0
121.90	4938.0
123.00	7549.0
123.90	3510.0
124.90	2958.0
126.00	1108.0
127.00	248320.0
128.00	20048.0
129.00	97192.0
130.00	8290.0
131.00	1598.0
131.90	967.0
132.90	714.0
134.00	2690.0
135.00	8094.0

Instrument :
BNA_N
ClientSampleId :
DFTPP264

136.00	3321.0
137.00	5588.0
137.90	922.0
138.90	606.0
140.00	1094.0
141.00	12075.0
142.00	4227.0
143.00	2767.0
143.90	913.0
144.90	732.0
146.00	2266.0
147.00	7169.0
148.00	15577.0
149.00	3889.0
149.90	944.0
151.00	1579.0
151.70	956.0
152.90	3935.0
154.00	2770.0
155.00	7195.0
156.00	10296.0
157.00	2127.0
157.90	2164.0
159.00	2055.0
160.00	3850.0
161.00	5597.0
161.90	1765.0
162.90	690.0
164.00	795.0
165.00	4763.0
166.00	3572.0
167.00	28336.0
168.00	19608.0
169.00	2893.0
169.90	850.0
170.80	974.0
172.00	2199.0
173.00	2712.0
174.00	5042.0
175.00	9708.0
176.00	2997.0
177.00	5364.0
178.00	1438.0
178.90	19104.0
180.00	15166.0
181.00	5853.0
181.90	872.0
183.00	538.0
183.90	1468.0
185.00	8275.0
186.00	67256.0
187.00	19736.0
188.00	2596.0
189.00	4173.0
189.90	816.0
190.90	2608.0
192.00	5924.0
193.00	6958.0
194.00	1476.0
194.90	641.0
196.00	13554.0
196.70	2548.0
197.90	476288.0
198.90	34584.0
200.00	2860.0
201.40	2343.0
203.00	3546.0
204.00	17920.0
205.00	28856.0
206.00	126592.0
207.00	24656.0
208.00	5700.0
209.00	2410.0
210.00	1952.0
211.00	5060.0
212.90	328.0
214.00	232.0
214.90	1396.0
216.00	2689.0
217.00	31320.0
218.00	4668.0
218.90	524.0
221.00	18880.0
221.80	4580.0
223.00	7037.0
224.00	66320.0
225.00	17760.0
226.00	2540.0
226.90	30784.0
228.00	4042.0
228.90	5727.0
230.00	571.0
231.00	2588.0
232.00	374.0
233.00	603.0
233.90	1658.0

Instrument :
BNA_N
ClientSampleId :
DFTPP264

234.90	1724.0
235.90	1406.0
237.00	3780.0
238.00	583.0
239.00	1321.0
239.90	964.0
241.00	1457.0
242.00	3381.0
243.00	3784.0
244.00	53280.0
245.00	7283.0
246.00	10153.0
246.90	2000.0
247.90	433.0
248.90	2143.0
250.10	335.0
251.00	410.0
252.10	372.0
252.90	1004.0
254.00	1747.0
255.00	263168.0
256.00	41128.0
257.00	4140.0
257.90	13823.0
259.00	2500.0
259.90	469.0
261.00	542.0
262.90	257.0
263.80	426.0
265.00	5990.0
266.00	955.0
267.10	213.0
269.90	264.0
270.90	798.0
272.00	651.0
273.00	6982.0
274.00	21624.0
275.00	109304.0
276.00	16608.0
276.90	8819.0
278.00	1940.0
278.90	552.0
280.90	877.0
282.00	330.0
283.00	1266.0
284.00	793.0
285.00	1496.0
285.90	242.0
289.00	355.0
290.00	260.0
290.90	217.0
291.90	423.0
293.00	2830.0
293.90	831.0
294.90	522.0
296.00	31120.0
297.00	4307.0
297.80	373.0
301.00	377.0
302.00	550.0
303.00	3519.0
304.00	927.0
304.90	244.0
308.00	425.0
308.90	393.0
309.90	674.0
312.00	162.0
312.80	274.0
314.00	1638.0
315.00	3501.0
316.00	1704.0
316.90	279.0
320.90	1101.0
321.90	511.0
323.00	8516.0
324.00	2019.0
324.80	207.0
326.00	226.0
326.90	2084.0
327.90	941.0
328.90	221.0
331.80	775.0
333.00	1062.0
334.00	5866.0
334.90	1720.0
336.00	242.0
338.90	153.0
340.00	154.0
340.80	1183.0
341.80	311.0
345.90	2083.0
347.00	457.0
350.90	196.0
352.00	2853.0
353.00	1879.0
354.00	2999.0

Instrument :
BNA_N
ClientSampleId :
DFTPP264

355.00	762.0
358.90	160.0
364.90	11865.0
365.90	1937.0
369.80	274.0
370.90	803.0
371.90	5511.0
373.00	1374.0
373.90	261.0
376.80	191.0
382.90	1240.0
384.10	331.0
384.80	211.0
389.90	485.0
391.00	578.0
392.00	305.0
400.90	260.0
402.00	1778.0
403.00	2908.0
403.90	1342.0
404.80	386.0
414.80	186.0
421.00	2229.0
422.00	2453.0
423.00	16688.0
424.00	7681.0
424.90	1109.0
440.90	46792.0
442.00	305536.0
442.90	57592.0
444.00	6785.0
444.90	356.0

Instrument :
BNA_N
ClientSampleId :
DFTPP264

m/z	Abundance
36.00	335.0
37.00	1301.0
38.00	3297.0
39.00	19944.0
40.00	1619.0
41.00	875.0
43.00	219.0
43.90	1115.0
44.90	672.0
47.00	177.0
48.00	215.0
49.00	2343.0
50.00	69552.0
51.00	263616.0
52.00	13424.0
53.00	522.0
55.00	1570.0
55.90	7532.0
57.00	17424.0
57.90	805.0
59.00	230.0
60.00	316.0
61.00	3734.0
62.00	4397.0
63.00	11617.0
64.00	1627.0
65.00	5762.0
65.90	352.0
67.00	408.0
68.00	5177.0
69.00	290880.0
70.00	1571.0
71.00	168.0
73.00	3199.0
74.00	31720.0
75.00	48688.0
76.00	16912.0
77.00	342400.0
78.00	23896.0
79.00	23080.0
80.00	16928.0
81.00	24384.0
82.00	5968.0
83.00	5068.0
83.90	785.0
85.00	4062.0
86.00	6932.0
87.00	3460.0
88.00	1295.0
88.90	851.0
91.00	5265.0
92.00	6471.0
93.00	40856.0
94.00	2499.0
94.90	632.0
96.00	1950.0
96.80	706.0
97.90	32048.0
99.00	26672.0
99.90	2092.0
101.00	13740.0
102.00	727.0
103.00	4608.0
103.90	8994.0
105.00	8655.0
106.00	2483.0
107.00	105784.0
108.00	17112.0
109.00	3104.0
110.00	185856.0
111.00	28792.0
112.00	3545.0
113.00	1299.0
114.10	252.0
114.90	458.0
116.00	5256.0
117.00	97400.0
118.00	7376.0
119.00	1077.0
120.00	1480.0
120.80	522.0
122.00	7455.0
123.00	10915.0
124.00	4750.0
124.90	4219.0
127.00	362048.0
128.00	27984.0
129.00	141184.0
130.00	12457.0
130.90	2569.0
132.00	1159.0
132.90	1162.0

Instrument :
BNA_N
ClientSampleId :
DFTPP264

134.00	3675.0
135.00	11372.0
135.90	5051.0
137.00	6598.0
137.90	1367.0
138.90	962.0
140.00	1751.0
140.90	17704.0
142.00	6145.0
142.90	4406.0
143.90	907.0
145.00	929.0
146.00	3396.0
147.00	9718.0
148.00	24080.0
149.00	5086.0
149.90	1232.0
151.10	2221.0
153.00	5972.0
154.00	4564.0
155.00	10198.0
156.00	15260.0
157.00	2755.0
158.00	3524.0
159.00	2386.0
160.00	5654.0
161.00	9057.0
161.90	2373.0
162.90	960.0
164.00	968.0
165.00	6809.0
166.00	5598.0
167.00	41768.0
168.00	24504.0
169.00	3454.0
170.00	1293.0
171.00	1411.0
171.90	3622.0
173.00	4402.0
174.00	8103.0
175.00	14754.0
176.00	4315.0
177.00	7743.0
177.90	2408.0
178.90	28296.0
180.00	20048.0
181.00	9567.0
181.90	1547.0
182.80	915.0
183.90	2299.0
185.00	12715.0
186.00	103944.0
187.00	31320.0
188.00	3763.0
188.90	6781.0
190.00	976.0
190.90	3669.0
192.00	9885.0
193.00	10848.0
194.00	2299.0
195.00	1129.0
196.00	21528.0
197.00	4083.0
197.90	750976.0
199.00	52000.0
200.00	4187.0
201.50	3265.0
203.00	5369.0
204.00	27224.0
205.00	47928.0
206.00	196800.0
207.00	33456.0
208.00	8327.0
209.00	3304.0
210.10	3539.0
210.90	7879.0
213.00	520.0
214.00	252.0
215.00	2038.0
216.00	4178.0
216.90	52320.0
217.90	6453.0
219.00	563.0
221.00	30312.0
221.80	7454.0
223.00	11273.0
224.00	112112.0
225.00	28448.0
226.00	3776.0
226.90	45360.0
228.00	7222.0
228.90	9400.0
229.90	1214.0
231.00	4219.0
232.00	757.0
233.00	939.0

Instrument :
BNA_N
ClientSampleId :
DFTPP264

234.00	2624.0
235.00	3056.0
235.90	2345.0
237.00	4364.0
238.00	775.0
239.00	1876.0
240.10	1722.0
240.90	2924.0
242.00	5577.0
243.00	6215.0
244.00	85344.0
245.00	11565.0
246.00	18032.0
247.00	3408.0
248.00	859.0
248.90	3337.0
249.90	691.0
250.90	831.0
252.00	834.0
253.00	1882.0
255.00	432384.0
256.00	63832.0
257.00	5067.0
258.00	25880.0
258.90	4465.0
260.00	816.0
260.90	665.0
262.00	199.0
262.90	318.0
263.90	669.0
265.00	10560.0
266.00	1494.0
266.90	441.0
268.10	166.0
269.90	429.0
270.90	1225.0
272.00	1120.0
273.00	13374.0
274.00	35568.0
275.00	200640.0
276.00	25616.0
276.90	17648.0
278.00	2888.0
279.00	760.0
281.00	893.0
282.00	581.0
283.00	2177.0
284.00	1447.0
285.00	2811.0
286.00	449.0
287.90	271.0
289.00	747.0
289.90	480.0
291.00	490.0
292.00	812.0
292.90	3727.0
293.90	1287.0
296.00	53824.0
297.00	7895.0
297.90	715.0
301.00	797.0
302.00	974.0
303.00	6066.0
304.00	1667.0
305.00	307.0
308.00	717.0
308.90	543.0
309.90	827.0
311.00	188.0
311.80	276.0
312.90	498.0
314.00	2462.0
315.00	6137.0
316.00	3604.0
317.00	644.0
319.90	322.0
321.00	1522.0
321.90	778.0
323.00	16528.0
324.00	3301.0
325.00	455.0
326.00	436.0
326.90	3843.0
327.90	1642.0
329.00	446.0
331.90	1381.0
333.10	1807.0
334.00	12272.0
335.00	3037.0
335.90	426.0
338.90	422.0
340.20	239.0
341.00	1862.0
342.00	435.0
345.90	3903.0
346.90	764.0

Instrument :
BNA_N
ClientSampleId :
DFTPP264

351.00	397.0
352.00	5271.0
353.00	3833.0
354.00	5372.0
355.00	1403.0
359.00	460.0
364.90	26544.0
365.90	3940.0
366.90	360.0
370.00	714.0
370.90	1480.0
372.00	9365.0
373.00	2408.0
373.90	338.0
376.90	219.0
382.90	2149.0
383.90	659.0
384.90	176.0
389.90	1226.0
391.00	1035.0
391.90	639.0
400.90	460.0
402.00	3677.0
402.90	5046.0
403.90	1782.0
404.90	611.0
410.00	269.0
415.00	238.0
420.90	4964.0
421.90	4661.0
423.00	36904.0
424.00	11342.0
425.00	1826.0
425.90	221.0
440.00	546.0
441.00	97800.0
442.00	701760.0
443.00	123448.0
444.00	12318.0
444.90	823.0

Instrument :
BNA_N
ClientSampleId :
DFTPP264

m/z	Abundance
36.00	266.0
37.00	1025.0
38.00	2915.0
39.00	16456.0
40.00	1392.0
41.00	633.0
42.90	170.0
43.90	1052.0
45.00	551.0
46.90	226.0
47.90	222.0
49.00	1822.0
50.00	56928.0
51.00	214912.0
52.00	11525.0
53.00	360.0
55.00	1401.0
55.90	6531.0
57.00	14888.0
57.90	752.0
59.00	229.0
59.90	245.0
61.00	2845.0
62.00	3711.0
63.00	8715.0
64.00	1130.0
65.00	4598.0
65.90	356.0
66.90	416.0
68.00	4154.0
69.00	242240.0
70.00	1441.0
70.90	174.0
73.00	3212.0
74.00	25784.0
75.00	39992.0
76.00	14075.0
77.00	290688.0
78.00	20456.0
79.00	19168.0
80.00	13995.0
81.00	19888.0
82.00	4195.0
83.00	4677.0
83.90	698.0
84.90	4402.0
86.00	6416.0
87.00	2835.0
87.90	983.0
89.00	538.0
91.00	4806.0
92.00	5529.0
93.00	34504.0
93.90	2161.0
94.90	725.0
96.00	2285.0
97.10	696.0
98.00	28128.0
99.00	21640.0
100.00	1762.0
101.00	11391.0
101.90	686.0
102.90	3600.0
103.90	7530.0
104.90	6872.0
106.00	2158.0
107.00	90064.0
108.00	13848.0
109.00	2439.0
110.00	162944.0
111.00	24376.0
112.00	3008.0
112.90	910.0
114.10	251.0
114.90	555.0
116.00	5033.0
116.90	84448.0
118.00	5989.0
119.00	1000.0
120.00	1231.0
121.00	470.0
122.00	6129.0
123.00	10016.0
124.00	4711.0
125.00	4373.0
126.00	1453.0
127.00	312128.0
128.00	24320.0
129.00	120448.0
130.00	9869.0
130.90	1771.0
132.00	1027.0

Instrument :
BNA_N
ClientSampleId :
DFTPP264

132.90	1062.0
134.00	3928.0
135.00	10508.0
136.00	3972.0
137.00	6195.0
137.80	1084.0
138.90	573.0
140.00	1418.0
141.00	16140.0
142.00	5234.0
143.00	3478.0
143.90	805.0
145.00	970.0
145.90	2997.0
146.90	8827.0
148.00	21168.0
149.00	4164.0
150.00	1317.0
151.10	1863.0
151.80	1293.0
152.90	5179.0
154.00	4197.0
155.00	8954.0
156.00	14634.0
157.00	2512.0
158.00	2661.0
158.90	2200.0
160.00	4940.0
161.00	7573.0
162.00	2195.0
162.90	760.0
163.90	796.0
165.00	6025.0
166.00	5333.0
167.00	37120.0
168.00	21344.0
169.00	3072.0
170.00	1148.0
170.90	1752.0
171.90	3287.0
173.00	3640.0
174.00	7315.0
175.00	13642.0
176.00	3027.0
176.90	7065.0
177.90	2180.0
179.00	25320.0
180.00	18648.0
181.00	8592.0
182.00	1319.0
183.00	806.0
184.00	2035.0
185.00	11643.0
186.00	97624.0
187.00	27560.0
188.00	2901.0
189.00	5564.0
189.90	1051.0
191.00	3649.0
192.00	8939.0
193.00	10233.0
194.00	2090.0
195.00	1215.0
196.00	20112.0
197.00	3572.0
197.90	695872.0
198.90	45328.0
200.00	3516.0
201.40	2944.0
202.90	5231.0
204.00	25680.0
205.00	42696.0
206.00	182336.0
207.00	30368.0
208.00	7374.0
209.00	3219.0
210.10	3018.0
211.00	7742.0
211.70	1367.0
212.90	583.0
213.90	156.0
215.00	2419.0
216.00	4188.0
217.00	47648.0
218.00	6201.0
218.90	568.0
221.00	28984.0
221.80	7771.0
223.00	10870.0
224.00	104912.0
225.00	25816.0
226.10	3287.0
227.00	42664.0
228.00	6566.0
229.00	8556.0
230.00	1055.0

Instrument :
BNA_N
ClientSampleId :
DFTPP264

231.00	3715.0
231.90	589.0
232.90	715.0
234.00	2365.0
234.90	3325.0
235.90	2459.0
237.00	4226.0
237.90	639.0
238.90	1620.0
239.90	1293.0
241.00	2631.0
242.00	5618.0
243.00	6280.0
244.00	86168.0
245.00	10817.0
246.00	16688.0
247.00	3371.0
247.90	806.0
249.00	3061.0
249.90	600.0
250.90	798.0
252.00	622.0
252.90	1729.0
255.00	425920.0
256.00	64576.0
256.90	5031.0
258.00	23752.0
258.90	4088.0
260.00	826.0
260.90	720.0
262.80	260.0
263.90	565.0
265.00	10137.0
265.90	1458.0
267.00	275.0
268.90	178.0
270.10	479.0
270.90	942.0
272.00	1393.0
273.00	13724.0
274.00	34488.0
275.00	183680.0
276.00	24920.0
277.00	16512.0
277.90	2511.0
278.90	588.0
281.00	1198.0
281.90	459.0
283.00	2103.0
283.90	1281.0
285.00	3228.0
286.10	431.0
287.90	178.0
289.00	927.0
289.90	581.0
290.90	432.0
292.00	790.0
293.00	4193.0
294.00	1113.0
296.00	58584.0
297.00	8019.0
298.00	616.0
300.90	814.0
302.00	1141.0
303.00	6396.0
304.10	1760.0
305.00	249.0
308.00	854.0
308.90	532.0
309.90	584.0
310.90	201.0
311.90	207.0
312.90	561.0
314.00	2733.0
314.90	6124.0
316.00	3898.0
316.90	615.0
319.90	186.0
321.00	1915.0
322.00	993.0
323.00	18032.0
324.00	3419.0
324.90	248.0
325.90	459.0
326.90	3653.0
328.00	1630.0
329.00	211.0
332.00	1201.0
333.00	1705.0
333.90	11944.0
335.00	2783.0
336.00	372.0
338.90	206.0
340.00	219.0
341.00	2415.0
342.00	762.0

Instrument :
BNA_N
ClientSampleId :
DFTPP264

342.80	196.0
346.00	5162.0
346.90	936.0
349.90	251.0
351.10	307.0
352.00	6014.0
353.00	3898.0
354.00	6572.0
355.00	1290.0
356.00	174.0
358.90	412.0
364.00	221.0
364.90	24176.0
365.90	4141.0
366.80	251.0
369.90	562.0
370.90	1656.0
372.00	10074.0
373.00	2581.0
374.00	298.0
376.90	362.0
382.90	2739.0
383.90	976.0
384.90	261.0
389.90	1371.0
391.00	919.0
392.00	861.0
401.00	619.0
401.90	4302.0
403.00	5941.0
403.90	2284.0
404.90	366.0
409.90	241.0
414.90	500.0
419.90	171.0
420.90	6043.0
422.00	5613.0
423.00	39048.0
424.00	10664.0
425.00	1452.0
441.00	127904.0
442.00	783104.0
443.00	183232.0
444.00	14411.0
445.00	840.0

Instrument :
BNA_N
ClientSampleId :
DFTPP264